


245.1

LB90344
INSTRUMENT ID: CV2

Linear

 μ Abs.:

31631



A= 0.0000e+000

B= 3.1548e-004

C= 3.3968e-002

Rho= 0.9999869

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	-85	0.000	-85				
0.2	0.200	0.219	0.019	586	0.0 %	586				
2.5	2.500	2.486	-0.014	7774	0.0 %	7774				
5.0	5.000	4.964	-0.036	15628	0.0 %	15628				
7.5	7.500	7.510	0.010	23699	0.0 %	23699				
10.0	10.000	10.013	0.013	31631	0.0 %	31631				

LB90344 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Sep 2017 10:01:35

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	SO	-85	-	0.0000	245.1	PPB	27 Sep 2017 10:41:17	S	Std
0.2 - 1	SO.2	586	-	0.2000	245.1	PPB	27 Sep 2017 10:43:26	S	Std
2.5 - 1	SO.5	7774	-	2.5000	245.1	PPB	27 Sep 2017 10:45:34	S	Std
5.0 - 1	SO	15628	-	5.0000	245.1	PPB	27 Sep 2017 10:47:42	S	Std
7.5 - 1	SO.5	23699	-	7.5000	245.1	PPB	27 Sep 2017 10:49:50	S	Std
10.0 - 1	SO	31631	-	10.0000	245.1	PPB	27 Sep 2017 10:51:59	S	Std
ICV76 - 1	ICV76	12191	3.8799	-245.1		PPB	27 Sep 2017 11:00:43	U	SMPL
ICB76 - 1	ICB76	-187	-0.0250	-245.1		PPB	27 Sep 2017 11:02:50	U	SMPL
CCV32 - 1	CCV32	15735	4.9980	-245.1		PPB	27 Sep 2017 11:04:57	U	SMPL
CCB32 - 1	CCB32	-204	-0.0304	-245.1		PPB	27 Sep 2017 11:07:04	U	SMPL
CRA - 1	CRA	558	0.2100	-245.1		PPB	27 Sep 2017 11:09:11	U	SMPL
HighStd - 1	HighStd	31693	10.0324	-245.1		PPB	27 Sep 2017 11:11:17	U	SMPL
ChkStd - 1	ChkStd	21876	6.9353	-245.1		PPB	27 Sep 2017 11:13:24	U	SMPL
PB102631BL - 1	PBW	-289	-0.0572	-245.1		PPB	27 Sep 2017 11:15:33	U	SMPL
PB102631BS - 1	LCSW	12358	3.9326	-245.1		PPB	27 Sep 2017 11:17:42	U	SMPL
I5275-07 - 1	LOD-WATER-07-QT4-2017	10	0.0371	-245.1		PPB	27 Sep 2017 11:19:49	U	SMPL
I5275-08 - 1	LOQ-WATER-08-QT4-2017	1186	0.4081	-245.1		PPB	27 Sep 2017 11:21:57	U	SMPL
I5426-01 - 1	092517-TIDO-F-D-M	-180	-0.0228	-245.1		PPB	27 Sep 2017 11:24:06	U	SMPL
I5426-02 - 1	092517-TIDO-F-U-M	-179	-0.0225	-245.1		PPB	27 Sep 2017 11:26:14	U	SMPL
I5427-01 - 1	092217-SIDI-F-D-B	-174	-0.0209	-245.1		PPB	27 Sep 2017 11:28:21	U	SMPL
CCV33 - 1	CCV33	15903	5.0510	-245.1		PPB	27 Sep 2017 11:30:28	U	SMPL
CCB33 - 1	CCB33	-199	-0.0288	-245.1		PPB	27 Sep 2017 11:32:35	U	SMPL
I5427-02 - 1	092217-SIDI-F-D-M	-174	-0.0209	-245.1		PPB	27 Sep 2017 11:34:42	U	SMPL
I5427-03 - 1	092217-SIDI-F-D-S	-158	-0.0159	-245.1		PPB	27 Sep 2017 11:36:49	U	SMPL
I5427-04 - 1	092217-SIDI-F-U-B	-140	-0.0102	-245.1		PPB	27 Sep 2017 11:38:56	U	SMPL
I5427-05 - 1	092217-SIDI-F-U-M	-135	-0.0086	-245.1		PPB	27 Sep 2017 11:41:03	U	SMPL
I5427-06 - 1	092217-SIDI-F-U-S	-149	-0.0130	-245.1		PPB	27 Sep 2017 11:43:10	U	SMPL
I5427-06DUP - 1	092217-SIDI-F-U-SDUP	-151	-0.0137	-245.1		PPB	27 Sep 2017 11:45:18	U	SMPL
I5427-06MS - 1	092217-SIDI-F-U-SMS	14972	4.7573	-245.1		PPB	27 Sep 2017 11:47:26	U	SMPL
I5427-06MSD - 1	092217-SIDI-F-U-SMSD	15639	4.9677	-245.1		PPB	27 Sep 2017 11:49:34	U	SMPL
I5427-06LX5 - 1		-212	-0.0329	-245.1		PPB	27 Sep 2017 11:51:43	U	SMPL
I5427-06A - 1		15312	4.8646	-245.1		PPB	27 Sep 2017 11:53:50	U	SMPL
CCV34 - 1	CCV34	15506	4.9258	-245.1		PPB	27 Sep 2017 11:55:57	U	SMPL
CCB34 - 1	CCB34	-197	-0.0282	-245.1		PPB	27 Sep 2017 11:58:04	U	SMPL