

Analytical Summary ReportAnalysis Method: SM5220 DANALYST: IndianaParameter: CODSUPERVISOR REVIEW BY: apatelRun Number: LB90345

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP58116
COD calibration std. 100 ppm	WP58113
COD calibration std. 150 ppm	WP58114
COD calibration std. 50 ppm	WP58115
COD calibration std. 0 ppm	WP58111
COD Digestion Vials Low Level 0-150Mg/L	W2217
COD ICV-LCS std, 50ppm	WP58302
COD CCV std, 50ppm	WP58303

Temp In (C): <u>150</u>	Date In: <u>09/27/2017</u>	Time In: <u>09:25</u>
Temp Out (C): <u>150</u>	Date Out: <u>09/27/2017</u>	Time Out: <u>11:25</u>

Intercept: 0.4899Slope: 1.0276Regression: 0.9997

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	09/21/2017	12:44
2	CAL2	10	1	Water	12.000	09/21/2017	12:45
3	CAL3	50	1	Water	50.000	09/21/2017	12:45
4	CAL4	100	1	Water	105.000	09/21/2017	12:46
5	CAL5	150	1	Water	154.000	09/21/2017	12:46



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB90345

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	48.180	09/21/2017	12:47
2	ICB				1	Water	0.000	-0.477	09/21/2017	12:47
3	CCV1	50			1	Water	52.000	50.127	09/27/2017	13:00
4	CCB1				1	Water	0.000	-0.477	09/27/2017	13:00
5	LB90345BL				1	Water	0.000	-0.477	09/27/2017	13:01
6	LB90345BS	50			1	Water	51.000	49.153	09/27/2017	13:01
7	I5429-01				1	Water	6.000	5.362	09/27/2017	13:02
8	I5429-01DUP				1	Water	7.000	6.335	09/27/2017	13:02
9	I5429-01MS	50			1	Water	57.000	54.992	09/27/2017	13:03
10	I5429-01MSD	50			1	Water	59.000	56.939	09/27/2017	13:03
11	CCV2	50			1	Water	52.000	50.127	09/27/2017	13:04
12	CCB2				1	Water	0.000	-0.477	09/27/2017	13:04

2890345

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD092717

WorkList ID : 103306

Date : 9/27/2017 8:52:02 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2017	Water	I5429-01	COD	Conc H2SO4 tr	PSEG01	H52	SV-26493-LM	09/25/2017	SM5220 D

Date/Time 09/27/17 9:00
Received by: IB
Relinquished by:

Date/Time 09/27/17 10:00
Received by: IB
Relinquished by: IB