


7470A

LB90348
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36437



A= 0.0000e+000

B= 2.7789e-004

C= 2.9025e-002

Rho= 0.9993858

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	-48	0.000	-48				
0.2	0.200	0.238	0.038	753	0.0 %	753				
2.5	2.500	2.624	0.124	9338	0.0 %	9338				
5.0	5.000	4.788	-0.212	17125	0.0 %	17125				
7.5	7.500	7.380	-0.120	26453	0.0 %	26453				
10.0	10.000	10.154	0.154	36437	0.0 %	36437				

LB90348 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Sep 2017 12:07:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-48	-	0.0000	7470A	PPB	27 Sep 2017 12:17:50	S	Std
0.2 - 1	50.2	753	-	0.2000	7470A	PPB	27 Sep 2017 12:19:59	S	Std
2.5 - 1	52.5	9338	-	2.5000	7470A	PPB	27 Sep 2017 12:22:07	S	Std
5.0 - 1	55	17125	-	5.0000	7470A	PPB	27 Sep 2017 12:24:16	S	Std
7.5 - 1	57.5	26453	-	7.5000	7470A	PPB	27 Sep 2017 12:26:23	S	Std
10.0 - 1	510	36437	-	10.0000	7470A	PPB	27 Sep 2017 12:30:56	S	Std
ICV77 - 1	ICV77	13590	3.8055	-	7470A	PPB	27 Sep 2017 12:33:36	U	SMPL
ICB77 - 1	ICB77	-210	-0.0293	-	7470A	PPB	27 Sep 2017 12:35:43	U	SMPL
CCV35 - 1	CCV35	16788	4.6942	-	7470A	PPB	27 Sep 2017 12:37:49	U	SMPL
CCB35 - 1	CCB35	-198	-0.0260	-	7470A	PPB	27 Sep 2017 12:39:56	U	SMPL
CRA - 1	CRA	698	0.2230	-	7470A	PPB	27 Sep 2017 12:42:03	U	SMPL
HighStd - 1	HighStd	35775	9.9704	-	7470A	PPB	27 Sep 2017 12:44:10	U	SMPL
ChkStd - 1	ChkStd	23662	6.6044	-	7470A	PPB	27 Sep 2017 12:46:17	U	SMPL
PB102632BL - 1	PBW	-262	-0.0438	-	7470A	PPB	27 Sep 2017 12:48:26	U	SMPL
PB102632BS - 1	LCSW	13873	3.8841	-	7470A	PPB	27 Sep 2017 12:50:35	U	SMPL
PB102632TB - 1	PB102595TB	-184	-0.0221	-	7470A	PPB	27 Sep 2017 12:52:42	U	SMPL
I5415-01 - 1	OILY-RAGS-DEBRIS	-65	0.0110	-	7470A	PPB	27 Sep 2017 12:54:50	U	SMPL
I5421-02 - 1	JC-DRUM	97	0.0560	-	7470A	PPB	27 Sep 2017 12:56:59	U	SMPL
I5423-03 - 1	SA-06-092517-A	-73	0.0087	-	7470A	PPB	27 Sep 2017 12:59:08	U	SMPL
I5423-06 - 1	SA-06-092517-B	-86	0.0051	-	7470A	PPB	27 Sep 2017 13:01:14	U	SMPL
CCV36 - 1	CCV36	17061	4.7700	-	7470A	PPB	27 Sep 2017 13:03:21	U	SMPL
CCB36 - 1	CCB36	-194	-0.0249	-	7470A	PPB	27 Sep 2017 13:05:27	U	SMPL
I5425-01 - 1	ETGI-155	-13	0.0254	-	7470A	PPB	27 Sep 2017 13:07:35	U	SMPL
I5428-02 - 1	A-B-C-DRUM-COMP	-88	0.0046	-	7470A	PPB	27 Sep 2017 13:09:41	U	SMPL
I5429-02 - 1	SV-26493-LM	-59	0.0126	-	7470A	PPB	27 Sep 2017 13:11:48	U	SMPL
I5429-02DUP - 1	SV-26493-LMDUP	-27	0.0215	-	7470A	PPB	27 Sep 2017 13:13:55	U	SMPL
I5429-02MS - 1	SV-26493-LMMS	12473	3.4951	-	7470A	PPB	27 Sep 2017 13:16:02	U	SMPL
I5429-02MSD - 1	SV-26493-LMMSD	13253	3.7118	-	7470A	PPB	27 Sep 2017 13:18:10	U	SMPL
PB102633BL - 1	PBW	-170	-0.0182	-	7470A	PPB	27 Sep 2017 13:20:18	U	SMPL
PB102633BS - 1	LCSW	12590	3.5276	-	7470A	PPB	27 Sep 2017 13:22:26	U	SMPL
I5275-05 - 1	LOD-WATER-05-QT4-2017	39	0.0399	-	7470A	PPB	27 Sep 2017 13:24:35	U	SMPL
I5275-06 - 1	LOQ-WATER-06-QT4-2017	1326	0.3975	-	7470A	PPB	27 Sep 2017 13:26:42	U	SMPL
CCV37 - 1	CCV37	17160	4.7975	-	7470A	PPB	27 Sep 2017 13:28:48	U	SMPL
CCB37 - 1	CCB37	-194	-0.0249	-	7470A	PPB	27 Sep 2017 13:30:55	U	SMPL
I5429-01 - 1	SV-26493-LM	-180	-0.0210	-	7470A	PPB	27 Sep 2017 13:33:02	U	SMPL
I5429-01DUP - 1	SV-26493-LMDUP	-177	-0.0202	-	7470A	PPB	27 Sep 2017 13:35:09	U	SMPL
I5429-01MS - 1	SV-26493-LMMS	13765	3.8541	-	7470A	PPB	27 Sep 2017 13:37:16	U	SMPL
I5429-01MSD - 1	SV-26493-LMMSD	13352	3.7394	-	7470A	PPB	27 Sep 2017 13:39:23	U	SMPL
I5429-02LX5 - 1		-183	-0.0218	-	7470A	PPB	27 Sep 2017 13:41:31	U	SMPL
I5429-02A - 1		13864	3.8816	-	7470A	PPB	27 Sep 2017 13:43:38	U	SMPL
I5429-01LX5 - 1		-211	-0.0296	-	7470A	PPB	27 Sep 2017 13:45:46	U	SMPL
I5429-01A - 1		14774	4.1345	-	7470A	PPB	27 Sep 2017 13:47:54	U	SMPL
CCV38 - 1	CCV38	17341	4.8478	-	7470A	PPB	27 Sep 2017 13:50:03	U	SMPL
CCB38 - 1	CCB38	-218	-0.0316	-	7470A	PPB	27 Sep 2017 13:52:10	U	SMPL