


7471A

LB90356
INSTRUMENT ID: C41

Linear

 μ Abs.:

42728



A= 0.0000e+000

B= 2.3290e-004

C= -1.4539e-003

Rho= 0.9998606

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	25	0.000	25				
0.2	0.200	0.199	-0.001	862	0.0 %	862				
2.5	2.500	2.417	-0.083	10383	0.0 %	10383				
5.0	5.000	5.114	0.114	21966	0.0 %	21966				
7.5	7.500	7.515	0.015	32273	0.0 %	32273	0			
10.0	10.000	9.950	-0.050	42728	0.0 %	42728				

LB90356 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 27 Sep 2017 15:16:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	7471A	PPB	27 Sep 2017 15:19:16	S	Std
0.2 - 1	50.2	862	-	0.2000	7471A	PPB	27 Sep 2017 15:21:34	S	Std
2.5 - 1	52.5	10383	-	2.5000	7471A	PPB	27 Sep 2017 15:23:52	S	Std
5.0 - 1	55	21966	-	5.0000	7471A	PPB	27 Sep 2017 15:26:09	S	Std
7.5 - 1	57.5	32273	-	7.5000	7471A	PPB	27 Sep 2017 15:28:26	S	Std
10.0 - 1	59.5	42728	-	10.0000	7471A	PPB	27 Sep 2017 15:33:16	S	Std
ICV78 - 1	ICV78	16818	3.9155	-7471A	PPB	27 Sep 2017 15:36:00	U	SMPL	
ICB78 - 1	ICB78	-200	-0.0480	-7471A	PPB	27 Sep 2017 15:38:16	U	SMPL	
CCV39 - 1	CCV39	22019	5.1268	-7471A	PPB	27 Sep 2017 15:40:31	U	SMPL	
CCB39 - 1	CCB39	-201	-0.0483	-7471A	PPB	27 Sep 2017 15:42:47	U	SMPL	
CRA - 1	CRA	834	0.1928	-7471A	PPB	27 Sep 2017 15:45:03	U	SMPL	
HighStd - 1	HighStd	43157	10.0499	-7471A	PPB	27 Sep 2017 15:47:19	U	SMPL	
ChkStd - 1	ChkStd	29923	6.9677	-7471A	PPB	27 Sep 2017 15:49:35	U	SMPL	
PB102640BL - 1	PBS	-318	-0.0755	-7471A	PPB	27 Sep 2017 15:51:53	U	SMPL	
PB102640BS - 1	LCSS	14735	3.4304	-7471A	PPB	27 Sep 2017 15:54:11	U	SMPL	
I5433-01 - 1	FK-BPM-01-G7-G8-G9-G10	878	0.2030	-7471A	PPB	27 Sep 2017 15:56:27	U	SMPL	
I5433-01DUP - 1	FK-BPM-01-G7-G8-G9-G10DUP	982	0.2273	-7471A	PPB	27 Sep 2017 15:58:45	U	SMPL	
I5433-01MS - 1	FK-BPM-01-G7-G8-G9-G10MS	17149	3.9926	-7471A	PPB	27 Sep 2017 16:01:02	U	SMPL	
I5433-01MSD - 1	FK-BPM-01-G7-G8-G9-G10MSD	17047	3.9688	-7471A	PPB	27 Sep 2017 16:03:20	U	SMPL	
I5434-12 - 1	UT5400533	2407	0.5591	-7471A	PPB	27 Sep 2017 16:05:36	U	SMPL	
CCV40 - 1	CCV40	21884	5.0954	-7471A	PPB	27 Sep 2017 16:07:52	U	SMPL	
CCB40 - 1	CCB40	-220	-0.0527	-7471A	PPB	27 Sep 2017 16:10:08	U	SMPL	
I5434-13 - 1	UT5400534	2726	0.6334	-7471A	PPB	27 Sep 2017 16:12:25	U	SMPL	
I5434-14 - 1	UT5400535	2410	0.5598	-7471A	PPB	27 Sep 2017 16:14:41	U	SMPL	
I5434-15 - 1	UT5400536	2384	0.5538	-7471A	PPB	27 Sep 2017 16:16:57	U	SMPL	
I5436-01 - 1	OILY-GRAVEL	-1	-0.0017	-7471A	PPB	27 Sep 2017 16:19:13	U	SMPL	
I5440-01 - 1	W-1	71569	16.6672	-7471A	PPB	27 Sep 2017 16:21:30	U	SMPL	
I5440-02 - 1	W-2	187	0.0421	-7471A	PPB	27 Sep 2017 16:23:47	U	SMPL	
I5440-03 - 1	W-3	7605	1.7698	-7471A	PPB	27 Sep 2017 16:26:04	U	SMPL	
I5444-01 - 1	BU-02-092617	2521	0.5857	-7471A	PPB	27 Sep 2017 16:28:25	U	SMPL	
I5444-04 - 1	BU-03-092617	474	0.1089	-7471A	PPB	27 Sep 2017 16:30:43	U	SMPL	
I5445-01 - 1	AU-05-092617-A	1366	0.3167	-7471A	PPB	27 Sep 2017 16:32:59	U	SMPL	
CCV41 - 1	CCV41	22107	5.1473	-7471A	PPB	27 Sep 2017 16:35:16	U	SMPL	
CCB41 - 1	CCB41	-209	-0.0501	-7471A	PPB	27 Sep 2017 16:37:32	U	SMPL	
I5445-04 - 1	AU-05-092617-B	1368	0.3172	-7471A	PPB	27 Sep 2017 16:39:49	U	SMPL	
I5445-07 - 1	AU-05-092617-C	1716	0.3982	-7471A	PPB	27 Sep 2017 16:42:05	U	SMPL	
I5446-01 - 1	CC-1	428	0.0982	-7471A	PPB	27 Sep 2017 16:44:21	U	SMPL	
I5447-01 - 1	SASP2P-ENV-116-7(10-12)	5173	1.2034	-7471A	PPB	27 Sep 2017 16:46:38	U	SMPL	
I5447-02 - 1	SASP2P-ENV-116-7(25-27)	250	0.0568	-7471A	PPB	27 Sep 2017 16:48:55	U	SMPL	
I5447-03 - 1	SASP2P-ENV-115-5(0-5)	6585	1.5322	-7471A	PPB	27 Sep 2017 16:51:11	U	SMPL	
I5447-04 - 1	SASP2P-ENV-115-5(5-10)	247	0.0561	-7471A	PPB	27 Sep 2017 16:53:29	U	SMPL	
I5448-01 - 1	SASP2P-TP-106C-(0-5)	2387	0.5545	-7471A	PPB	27 Sep 2017 16:55:46	U	SMPL	
I5433-01LX5 - 1		132	0.0293	-7471A	PPB	27 Sep 2017 16:58:04	U	SMPL	
I5433-01A - 1		18036	4.1992	-7471A	PPB	27 Sep 2017 17:00:21	U	SMPL	
CCV42 - 1	CCV42	22269	5.1851	-7471A	PPB	27 Sep 2017 17:02:37	U	SMPL	
CCB42 - 1	CCB42	-246	-0.0587	-7471A	PPB	27 Sep 2017 17:04:53	U	SMPL	
I5440-01DLX5 - 1		14153	3.2948	-7471A	PPB	27 Sep 2017 17:07:11	U	SMPL	
CCV43 - 1	CCV43	22068	5.1382	-7471A	PPB	27 Sep 2017 17:09:27	U	SMPL	
CCB43 - 1	CCB43	-229	-0.0548	-7471A	PPB	27 Sep 2017 17:11:43	U	SMPL	