



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90366

Temp.: 19.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/27/2017	17:10
2	CAL2	1	Water			7.00	09/27/2017	17:11
3	CAL3	1	Water			10.02	09/27/2017	17:13
4	ICV	1	Water			7.01	09/27/2017	17:15
5	CCV1	1	Water			2.01	09/27/2017	17:16
6	I5455-01	1	Water			5.80	09/27/2017	17:19
7	I5455-01DUP	1	Water			5.81	09/27/2017	17:20
8	I5471-01	1	Water			5.22	09/27/2017	17:21
9	CCV2	1	Water			12.01	09/27/2017	17:23

1BQ0366

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5455 WorkList ID : 103342 Date : 9/27/2017 5:03:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5455-01-A	pH	Cool 4 deg C	STER15	H62	GILROY-OMEGA-CELL	09/15/2017	SM 4500-PH B
09/01/2017	Water	I5471-01-B	pH	Cool 4 deg C	AREC01	--Select--	PT-01	09/27/2017	SM 4500-PH B

Date/Time 69/27/17 5:03 PM
Received by: 71 CP
Relinquished by: _____

Date/Time 09/27/17 6:00 PM
Received by: CP
Relinquished by: JD