



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB90374
Analysis Date: 09/28/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 09/28/2017
Extraction IN Time: 08:15
Extraction OUT Time: 09:00

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB90374BL	LB90374BL	WATER	2	1000	100	2.3697	2.3697	3.02	2.3697	2.3697	0	0
2	LB90374BS	LB90374BS	WATER	2	1000	100	2.9144	2.9144	3.01	2.9332	2.9332	0.0188	18.8
3	I5429-01	SV-26493-LM	WATER	2	1000	100	3.0853	3.0853	3.03	3.086	3.086	0.0007	0.7
4	I5483-01	STORM-DRAIN-1	WATER	2	1000	100	3.095	3.095	3.02	3.0953	3.0953	0.0003	0.3
5	I5483-01DUP	STORM-DRAIN-1DUP	WATER	2	1000	100	3.1503	3.1503	3.01	3.1507	3.1507	0.0004	0.4
6	I5483-02	STORM-DRAIN-2	WATER	2	1000	100	2.958	2.958	3.04	2.9582	2.9582	0.0002	0.2
7	I5483-02MS	STORM-DRAIN-2	WATER	2	1000	100	2.8611	2.8611	3.03	2.8812	2.8812	0.0201	20.1



QC Batch# LB90374 Test: TPH Analysis Date: 09/28/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2272
pH Paper 0-14	W2248
Sodium Sulfate	EP1786
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD		
MS/MSD	10.00 ML	WP58005

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 10:00
1.0000 gram Balance: 1.0005 In Time: 09:30
Bal Check Time: 08:35

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 11:30
1.0000 gram Balance: 1.0003 Out Time: 11:00
Bal Check Time: 12:10

2390374

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I5429

WorkList ID : 103350

Date : 9/28/2017 7:59:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2017	Water	I5429-01	K TPH	Conc H2SO4 1c	PSEG01	H52	SV-26493-LM	09/25/2017	1664A
	Water	I5483-01	E, F TPH	Conc H2SO4 1c	KWAR01	I12	STORM-DRAIN-1	09/27/2017	1664A
	Water	I5483-02	E, F TPH	Conc H2SO4 1c	KWAR01	I12	STORM-DRAIN-2	09/27/2017	1664A

Date/Time 09/28/17 8:00 AM
Received by: JP
Relinquished by: JP

Date/Time 09/28/17 10:00 AM
Received by: JP
Relinquished by: JP