

Daily Analysis Runlog For Sequence/QC Batch ID # LB90377

Review By		Review On				
SubDirectory		LB90377		Test : Anions		

STD. NAME	STD REF.#
ICAL Standard	WP57723,WP57724,WP57725,WP57726,WP57727,WP57728,WP57729
ICV Standard	WP57730
CCV Standard	WP58275
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	WP58276
Chk Standard	WP57731,WP57732

Sr#	SampleId	ClientID	QcType	Date	Comment	Status
1	STD1	STD1	CAL1	09/06/17 12:53		OK
2	STD2	STD2	CAL2	09/06/17 13:24		OK
3	STD3	STD3	CAL3	09/06/17 13:55		OK
4	STD4	STD4	CAL4	09/06/17 14:26		OK
5	STD5	STD5	CAL5	09/06/17 14:57		OK
6	STD6	STD6	CAL6	09/06/17 15:28		OK
7	STD7	STD7	CAL7	09/06/17 15:59		OK
8	ICV1	ICV1	ICV	09/06/17 17:32		OK
9	ICB1	ICB1	ICB	09/06/17 18:03		OK
10	CCV1	CCV1	CCV	09/27/17 10:54		OK
11	CCB1	CCB1	CCB	09/27/17 11:24		OK
12	LB90377BLW	LB90377BLW	MB	09/27/17 11:54		OK
13	LB90377BSW	LB90377BSW	LCS	09/27/17 12:24		OK
14	I5455-01	GILROY-OMEGA-CE	SAM	09/27/17 14:19		OK
15	I5455-01DUP	GILROY-OMEGA-CE	DUP	09/27/17 14:49		OK
16	I5455-01MS	GILROY-OMEGA-CE	MS	09/27/17 15:19		OK
17	CCV2	CCV2	CCV	09/27/17 16:33		OK
18	CCB2	CCB2	CCB	09/27/17 17:03		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/15/2017 12:00:00 AM
Project:	LB90377	Date Received:	9/27/2017 12:00:00 AM
Client Sample ID:	GILROY-OMEGA-CELL	SDG No.:	LB90377
Lab Sample ID:	I5455-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	Bromide	0	U	1	0.066		0.5	mg/L		09/27/2017	300.0
	Chloride	0.12	J	1	0.075		0.15	mg/L		09/27/2017	300.0
	Fluoride	0	U	1	0.043		0.1	mg/L		09/27/2017	300.0
	Nitrate	0.14		1	0.027		0.13	mg/L		09/27/2017	300.0
	Nitrite	0	U	1	0.022		0.15	mg/L		09/27/2017	300.0
	Phosphorus, Ortho	0	U	1	0.01		0.25	mg/L		09/27/2017	300.0
	Sulfate	0	U	1	0.13		0.75	mg/L		09/27/2017	300.0



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/15/2017 12:00:00 AM
Project:	LB90377	Date Received:	9/27/2017 12:00:00 AM
Client Sample ID:	GILROY-OMEGA-CELL	SDG No.:	LB90377
Lab Sample ID:	I5455-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB90377

SDG No.: LB90377

Contract: _____ Lab Code: CHEM

Case No.: LB90377

SAS No.: LB90377

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Bromide	9.8	10	98	0	90 - 110		09/06/2017	17:32	LB90377
	Chloride	2.9	3	96.7	0	90 - 110		09/06/2017	17:32	LB90377
	Fluoride	2	2	100	0	90 - 110		09/06/2017	17:32	LB90377
	Phosphorus, Ortho	5.2	5	104	0	90 - 110		09/06/2017	17:32	LB90377
	Nitrite	2.9	3	96.7	0	90 - 110		09/06/2017	17:32	LB90377
	Nitrate	2.4	2.5	96	0	90 - 110		09/06/2017	17:32	LB90377
	Sulfate	14.6	15	97.3	0	90 - 110		09/06/2017	17:32	LB90377



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB90377

SDG No.: LB90377

Contract: _____ Lab Code: CHEM

Case No.: LB90377

SAS No.: LB90377

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Bromide	9.5	10	95	0	90 - 110		09/27/2017	10:54	LB90377
	Chloride	2.9	3	96.7	0	90 - 110		09/27/2017	10:54	LB90377
	Fluoride	2.1	2	105	0	90 - 110		09/27/2017	10:54	LB90377
	Phosphorus, Ortho	4.9	5	98	0	90 - 110		09/27/2017	10:54	LB90377
	Nitrite	2.8	3	93.3	0	90 - 110		09/27/2017	10:54	LB90377
	Nitrate	2.3	2.5	92	0	90 - 110		09/27/2017	10:54	LB90377
	Sulfate	14	15	93.3	0	90 - 110		09/27/2017	10:54	LB90377
CCV2	Bromide	9.5	10	95	0	90 - 110		09/27/2017	16:33	LB90377
	Chloride	2.9	3	96.7	0	90 - 110		09/27/2017	16:33	LB90377
	Fluoride	2.1	2	105	0	90 - 110		09/27/2017	16:33	LB90377
	Phosphorus, Ortho	5.3	5	106	0	90 - 110		09/27/2017	16:33	LB90377
	Nitrite	2.8	3	93.3	0	90 - 110		09/27/2017	16:33	LB90377
	Nitrate	2.4	2.5	96	0	90 - 110		09/27/2017	16:33	LB90377
	Sulfate	14.1	15	94	0	90 - 110		09/27/2017	16:33	LB90377

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Bromide	0.066	+/-0.5	U		0.5		09/06/2017	18:03	LB90377
	Chloride	0.075	+/-0.15	U		0.15		09/06/2017	18:03	LB90377
	Fluoride	0.043	+/-0.1	U		0.1		09/06/2017	18:03	LB90377
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		09/06/2017	18:03	LB90377
	Nitrite	0.022	+/-0.15	U		0.15		09/06/2017	18:03	LB90377
	Nitrate	0.027	+/-0.13	U		0.13		09/06/2017	18:03	LB90377
	Sulfate	0.13	+/-0.75	U		0.75		09/06/2017	18:03	LB90377

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Bromide	0.066	+/-0.5	U		0.5		09/27/2017	11:24	LB90377
	Chloride	0.075	+/-0.15	U		0.15		09/27/2017	11:24	LB90377
	Fluoride	0.043	+/-0.1	U		0.1		09/27/2017	11:24	LB90377
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		09/27/2017	11:24	LB90377
	Nitrite	0.022	+/-0.15	U		0.15		09/27/2017	11:24	LB90377
	Nitrate	0.027	+/-0.13	U		0.13		09/27/2017	11:24	LB90377
	Sulfate	0.13	+/-0.75	U		0.75		09/27/2017	11:24	LB90377
CCB2	Bromide	0.066	+/-0.5	U		0.5		09/27/2017	17:03	LB90377
	Chloride	0.075	+/-0.15	U		0.15		09/27/2017	17:03	LB90377
	Fluoride	0.043	+/-0.1	U		0.1		09/27/2017	17:03	LB90377
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		09/27/2017	17:03	LB90377
	Nitrite	0.022	+/-0.15	U		0.15		09/27/2017	17:03	LB90377
	Nitrate	0.027	+/-0.13	U		0.13		09/27/2017	17:03	LB90377
	Sulfate	0.13	+/-0.75	U		0.75		09/27/2017	17:03	LB90377



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB90377

SDG No.: LB90377

Contract: Lab Code: CHEM

Case No.: LB90377

SAS No.: LB90377

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB90377

SDG No.: LB90377

Instrument: IC-2

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
LB90377BLW		WATER		Batch Number:				Prep Date:		
	Bromide	0.066	<0.5	U		0.5		09/27/2017	11:54	LB90377
	Chloride	0.075	<0.15	U		0.15		09/27/2017	11:54	LB90377
	Fluoride	0.043	<0.1	U		0.1		09/27/2017	11:54	LB90377
	Phosphorus, Ortho	0.01	<0.25	U		0.25		09/27/2017	11:54	LB90377
	Nitrite	0.022	<0.15	U		0.15		09/27/2017	11:54	LB90377
	Nitrate	0.027	<0.13	U		0.13		09/27/2017	11:54	LB90377
	Sulfate	0.13	<0.75	U		0.75		09/27/2017	11:54	LB90377



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB90377 level: low sdg no.: LB90377
contract: _____ lab code: CHEM case no.: LB90377 sas no.: LB90377
matrix: WATER sample id: I5455-01 client id: GILROY-OMEGA-CELLMS
Percent Solids for Sample: 0 Spiked ID: I5455-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Bromide	mg/L	80 - 120	9.9		0.066	U	10	99		
Chloride	mg/L	80 - 120	3.2		0.12	J	3	103		
Fluoride	mg/L	80 - 120	2.1		0.043	U	2	105		
Phosphorus, Ortho	mg/L	80 - 120	5.3		0.01	U	5	106		
Nitrite	mg/L	80 - 120	2.9		0.022	U	3	96.7		
Nitrate	mg/L	80 - 120	2.5		0.14		2.5	94		
Sulfate	mg/L	80 - 120	14.8		0.13	U	15	98.7		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90377 **Level:** LOW **SDG No.:** LB90377
Contract: **Lab Code:** CHEM **Case No.:** LB90377 **SAS No.:** LB90377
Matrix: WATER **Sample ID:** I5455-01 **Client ID:** GILROY-OMEGA-CELLDUP
Percent Solids for Sample: 0 **Duplicate ID** I5455-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Bromide	mg/L	20	0.066	U	0	U			
Chloride	mg/L	20	0.12	J	0.12	J	0		
Fluoride	mg/L	20	0.043	U	0	U			
Phosphorus, Ortho	mg/L	20	0.01	U	0	U			
Nitrite	mg/L	20	0.022	U	0	U			
Nitrate	mg/L	20	0.14		0.14		0		
Sulfate	mg/L	20	0.13	U	0	U			



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB90377

SDG No.: LB90377

Contract:

Lab Code: CHEM

Case No.: LB90377

SAS No.: LB90377

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB90377BSW							
Bromide	mg/L	10	9.5		95	90 - 110	
Chloride	mg/L	3	2.9		96.7	90 - 110	
Fluoride	mg/L	2	2.1		105	90 - 110	
Phosphorus, Ortho	mg/L	5	5.3		106	90 - 110	
Nitrite	mg/L	3	2.8		93.3	90 - 110	
Nitrate	mg/L	2.5	2.4		96	90 - 110	
Sulfate	mg/L	15	14.1		94	90 - 110	
