



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB90379

Reagent/Standard	Lot/Log #
conductivity standard	WP54988

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB90379BL		1	22.1	0.90	0.95	09/28/2017	09:01
2	LB90379BS	1413	1	21.7	1237.00	1320.21	09/28/2017	09:06
3	I5471-01		1	22.2	382.00	403.58	09/28/2017	09:10
4	I5471-01DUP		1	22.2	385.00	406.75	09/28/2017	09:15

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

0090379

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CON092817

WorkList ID : 103376

Date : 9/28/2017 8:26:13 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/01/2017	Water	I5471-01	Conductance	Cool 4 deg C	AREC01	--Select--	PT-01	09/27/2017	SM2510 B

Date/Time 9/28/17 8:48
Received by: AD
Relinquished by: AD

Date/Time 9/28/17 9:30
Received by: AD
Relinquished by: AD