

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: DHVANI

Date: 09/28/2017

DATE & TIME IN: 09/28/2017 12:00

OVEN TEMP: 103 °C

Run Number: LB90382

DATE & TIME OUT: 09/28/2017 16:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB90382BL	LB90382BL	1.396	1.396	100	1.396	1.396	1.396	0	0
2	LB90382BS	LB90382BS	1.3906	1.3906	100	1.4453	1.4453	1.4453	0.0547	547
3	I5483-01	STORM-DRAIN-1	1.389	1.389	1000	1.3903	1.3903	1.3903	0.0013	1.3
4	I5483-02	STORM-DRAIN-2	1.4034	1.4034	500	1.4043	1.4043	1.4043	0.0009	1.8
5	I5483-02DUP	STORM-DRAIN-2DUP	1.4017	1.4017	500	1.402	1.402	1.402	0.0003	0.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

UB 90 382

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss092817-2

WorkList ID : 103393

Date : 9/28/2017 11:23:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5483-01	—C TSS	Cool 4 deg C	KWAR01	I12	STORM-DRAIN-1	09/27/2017	SM2540 D
	Water	I5483-02	—C TSS	Cool 4 deg C	KWAR01	I12	STORM-DRAIN-2	09/27/2017	SM2540 D

Date/Time 09/28/17 11:45

Received by: DP

Relinquished by: CP

Date/Time 09/28/17 12:10

Received by: CP

Relinquished by: DP