


7470A

LB90383
INSTRUMENT ID: CV1

Linear

μ Abs.:

34237



A= 0.0000e+000

B= 2.9131e-004

C= 2.3722e-002

Rho= 0.9998986

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	-38	0.000	-38				
0.2	0.200	0.189	-0.011	568	0.0 %	568				
2.5	2.500	2.447	-0.053	8317	0.0 %	8317				
5.0	5.000	5.104	0.104	17441	0.0 %	17441				
7.5	7.500	7.450	-0.050	25492	0.0 %	25492				
10.0	10.000	9.997	-0.003	34237	0.0 %	34237				

LB90383 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Sep 2017 11:43:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-38	-	0.0000	7470A	PPB	28 Sep 2017 11:47:21	S	Std
0.2 - 1	50.2	568	-	0.2000	7470A	PPB	28 Sep 2017 11:49:30	S	Std
2.5 - 1	52.5	8317	-	2.5000	7470A	PPB	28 Sep 2017 11:51:38	S	Std
5.0 - 1	55	17441	-	5.0000	7470A	PPB	28 Sep 2017 11:53:46	S	Std
7.5 - 1	57.5	25492	-	7.5000	7470A	PPB	28 Sep 2017 11:55:54	S	Std
10.0 - 1	510	34237	-	10.0000	7470A	PPB	28 Sep 2017 11:58:02	S	Std
ICV79 - 1	ICV79	12948	3.7956	-	7470A	PPB	28 Sep 2017 12:00:44	U	SMPL
ICB79 - 1	ICB79	-159	-0.0226	-	7470A	PPB	28 Sep 2017 12:02:50	U	SMPL
CCV44 - 1	CCV44	17477	5.1150	-	7470A	PPB	28 Sep 2017 12:04:57	U	SMPL
CCB44 - 1	CCB44	-186	-0.0305	-	7470A	PPB	28 Sep 2017 12:07:04	U	SMPL
CRA - 1	CRA	547	0.1831	-	7470A	PPB	28 Sep 2017 12:09:10	U	SMPL
HighStd - 1	HighStd	34302	10.0163	-	7470A	PPB	28 Sep 2017 12:11:17	U	SMPL
ChkStd - 1	ChkStd	23284	6.8066	-	7470A	PPB	28 Sep 2017 12:13:24	U	SMPL
PB102668BL - 1	PBW	-244	-0.0474	-	7470A	PPB	28 Sep 2017 12:15:33	U	SMPL
PB102668BS - 1	LCSW	13168	3.8597	-	7470A	PPB	28 Sep 2017 12:17:41	U	SMPL
PB102668TB - 1	PB102629TB	-152	-0.0206	-	7470A	PPB	28 Sep 2017 12:19:48	U	SMPL
I5436-02 - 1	OILY-GRAVEL	-42	0.0115	-	7470A	PPB	28 Sep 2017 12:21:57	U	SMPL
I5444-03 - 1	BU-02-092617	1	0.0240	-	7470A	PPB	28 Sep 2017 12:24:05	U	SMPL
I5444-06 - 1	BU-03-092617	325	0.1184	-	7470A	PPB	28 Sep 2017 12:26:13	U	SMPL
I5445-03 - 1	AU-05-092617-A	-35	0.0135	-	7470A	PPB	28 Sep 2017 12:28:20	U	SMPL
CCV45 - 1	CCV45	17475	5.1144	-	7470A	PPB	28 Sep 2017 12:30:27	U	SMPL
CCB45 - 1	CCB45	-177	-0.0278	-	7470A	PPB	28 Sep 2017 12:32:34	U	SMPL
I5445-06 - 1	AU-05-092617-B	-58	0.0068	-	7470A	PPB	28 Sep 2017 12:34:41	U	SMPL
I5445-09 - 1	AU-05-092617-C	-38	0.0127	-	7470A	PPB	28 Sep 2017 12:36:48	U	SMPL
I5445-09DUP - 1	AU-05-092617-CDUP	-32	0.0144	-	7470A	PPB	28 Sep 2017 12:38:55	U	SMPL
I5445-09MS - 1	AU-05-092617-CMS	13184	3.8644	-	7470A	PPB	28 Sep 2017 12:41:02	U	SMPL
I5445-09MSD - 1	AU-05-092617-CMSD	13055	3.8268	-	7470A	PPB	28 Sep 2017 12:43:10	U	SMPL
I5445-09LX5 - 1		-116	-0.0101	-	7470A	PPB	28 Sep 2017 12:45:17	U	SMPL
I5445-09A - 1		14405	4.2201	-	7470A	PPB	28 Sep 2017 12:47:26	U	SMPL
CCV46 - 1	CCV46	17811	5.2123	-	7470A	PPB	28 Sep 2017 12:49:34	U	SMPL
CCB46 - 1	CCB46	-193	-0.0325	-	7470A	PPB	28 Sep 2017 12:51:43	U	SMPL