



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90384

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/27/2017	17:15
2	CAL2	1	Water			7.00	09/27/2017	17:16
3	CAL3	1	Water			10.03	09/27/2017	17:19
4	ICV	1	Water			7.00	09/27/2017	17:20
5	CCV1	1	Water			2.01	09/27/2017	17:21
6	I5472-02	1	Solid	20.02	20	7.69	09/27/2017	17:22
7	I5473-03	1	Solid	20.03	20	7.46	09/27/2017	17:23
8	I5475-03	1	Solid	20.04	20	6.91	09/27/2017	17:25
9	I5478-01	1	Solid	20.02	20	7.63	09/27/2017	17:26
10	I5481-03	1	Solid	20.01	20	9.11	09/27/2017	17:29
11	I5487-03	1	Solid	20.02	20	8.61	09/27/2017	17:31
12	I5487-06	1	Solid	20.03	20	8.50	09/27/2017	17:33
13	I5487-09	1	Solid	20.04	20	8.92	09/27/2017	17:35
14	I5487-12	1	Solid	20.01	20	9.16	09/27/2017	17:36
15	I5491-03	1	Solid	20.01	20	9.03	09/27/2017	17:37
16	CCV2	1	Water			12.01	09/27/2017	17:39
17	I5491-03DUP	1	Solid	20.03	20	9.04	09/27/2017	17:40
18	CCV3	1	Water			12.02	09/27/2017	17:41

1390384

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5440 WorkList ID : 103345 Date : 9/27/2017 5:00:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/03/2017	Solid	I5472-02	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-1-2-3-4-COMP	09/27/2017	9045C
	Solid	I5473-03	Corrosivity	Cool 4 deg C	PSEG05	H52	TR-04-092717	09/27/2017	9045C
10/03/2017	Solid	I5475-03	Corrosivity	Cool 4 deg C	PSEG01	H52	WS-20408-COMP	09/27/2017	9045C
	Solid	I5478-01	Corrosivity	Cool 4 deg C	SEAC01	I12	1612	09/27/2017	9045C
	Solid	I5481-03	Corrosivity	Cool 4 deg C	PSEG05	H52	HR-06-092717	09/27/2017	9045C
	Solid	I5487-03	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-02-092717-A	09/27/2017	9045C
09/28/2017	Solid	I5487-06	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-02-092717-B	09/27/2017	9045C
09/28/2017	Solid	I5487-09	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-02-092717-C	09/27/2017	9045C
09/28/2017	Solid	I5487-12	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-02-092717-D	09/27/2017	9045C
	Solid	I5491-03	Corrosivity	Cool 4 deg C	PSEG05	H52	OR-02-092717	09/27/2017	9045C

Date/Time 09/27/17 5:00P
Received by:
Relinquished by:

Date/Time 09/27/17 7:30P
Received by:
Relinquished by: