


7471A

LB90390
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40675



A= 0.0000e+000

B= 2.4660e-004

C= 2.4092e-002

Rho= 0.9999221

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	42	0.000	42				
0.2	0.200	0.184	-0.016	649	0.0 %	649				
2.5	2.500	2.538	0.038	10195	0.0 %	10195				
5.0	5.000	4.926	-0.074	19879	0.0 %	19879				
7.5	7.500	7.462	-0.038	30163	0.0 %	30163	0			
10.0	10.000	10.055	0.055	40675	0.0 %	40675				

LB90390 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 28 Sep 2017 14:00:43					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	42	-	0.0000	7471A	PPB	28 Sep 2017 14:09:55	S	Std
0.2 - 1	50.2	649	-	0.2000	7471A	PPB	28 Sep 2017 14:12:12	S	Std
2.5 - 1	52.5	10195	-	2.5000	7471A	PPB	28 Sep 2017 14:14:30	S	Std
5.0 - 1	55	19879	-	5.0000	7471A	PPB	28 Sep 2017 14:16:48	S	Std
7.5 - 1	57.5	30163	-	7.5000	7471A	PPB	28 Sep 2017 14:19:05	S	Std
10.0 - 1	510	40675	-	10.0000	7471A	PPB	28 Sep 2017 14:24:06	S	Std
ICV80 - 1	ICV80	17072	4.2341	-	7471A	PPB	28 Sep 2017 14:29:46	U	SMPL
ICB80 - 1	ICB80	-100	-0.0006	-	7471A	PPB	28 Sep 2017 14:32:02	U	SMPL
CCV47 - 1	CCV47	19886	4.9280	-	7471A	PPB	28 Sep 2017 14:34:18	U	SMPL
CCB47 - 1	CCB47	-184	-0.0213	-	7471A	PPB	28 Sep 2017 14:36:34	U	SMPL
CRA - 1	CRA	589	0.1693	-	7471A	PPB	28 Sep 2017 14:38:50	U	SMPL
HighStd - 1	HighStd	41175	10.1779	-	7471A	PPB	28 Sep 2017 14:41:06	U	SMPL
ChkStd - 1	ChkStd	27236	6.7405	-	7471A	PPB	28 Sep 2017 14:43:22	U	SMPL
PB102673BL - 1	PBS	-275	-0.0437	-	7471A	PPB	28 Sep 2017 14:45:40	U	SMPL
PB102673BS - 1	LCSS	15414	3.8252	-	7471A	PPB	28 Sep 2017 14:47:58	U	SMPL
I5275-01 - 1	LOD-SOIL-01-QT4-2017	81	0.0441	-	7471A	PPB	28 Sep 2017 14:50:15	U	SMPL
I5275-02 - 1	LOQ-SOIL-02-QT4-2017	1655	0.4322	-	7471A	PPB	28 Sep 2017 14:52:33	U	SMPL
I5472-01 - 1	TP-1-2-3-4-COMP	1016	0.2746	-	7471A	PPB	28 Sep 2017 14:54:51	U	SMPL
I5473-01 - 1	TR-04-092717	1611	0.4214	-	7471A	PPB	28 Sep 2017 14:57:08	U	SMPL
I5475-01 - 1	WS-20408-COMP	1208	0.3220	-	7471A	PPB	28 Sep 2017 14:59:25	U	SMPL
CCV48 - 1	CCV48	19856	4.9206	-	7471A	PPB	28 Sep 2017 15:01:40	U	SMPL
CCB48 - 1	CCB48	-163	-0.0161	-	7471A	PPB	28 Sep 2017 15:03:56	U	SMPL
I5481-01 - 1	HR-06-092717	4274	1.0781	-	7471A	PPB	28 Sep 2017 15:06:13	U	SMPL
I5485-01 - 1	WALDWICK-SWITCH	1427	0.3760	-	7471A	PPB	28 Sep 2017 15:08:29	U	SMPL
I5487-01 - 1	CL-02-092717-A	1292	0.3427	-	7471A	PPB	28 Sep 2017 15:10:45	U	SMPL
I5487-04 - 1	CL-02-092717-B	16457	4.0824	-	7471A	PPB	28 Sep 2017 15:13:01	U	SMPL
I5488-01 - 1	PAMT-COMP-C	4809	1.2100	-	7471A	PPB	28 Sep 2017 15:15:18	U	SMPL
I5488-01DUP - 1	PAMT-COMP-CDUP	5138	1.2911	-	7471A	PPB	28 Sep 2017 15:17:35	U	SMPL
I5488-02 - 1	I5488-01MS	20056	4.9699	-	7471A	PPB	28 Sep 2017 15:19:52	U	SMPL
I5488-03 - 1	I5488-01MSD	19976	4.9502	-	7471A	PPB	28 Sep 2017 15:22:10	U	SMPL
I5488-04 - 1	PAMT-COMP-D	7860	1.9624	-	7471A	PPB	28 Sep 2017 15:24:29	U	SMPL
I5491-01 - 1	OR-02-092717	1000	0.2707	-	7471A	PPB	28 Sep 2017 15:26:45	U	SMPL
CCV49 - 1	CCV49	19812	4.9098	-	7471A	PPB	28 Sep 2017 15:29:01	U	SMPL
CCB49 - 1	CCB49	-179	-0.0200	-	7471A	PPB	28 Sep 2017 15:31:17	U	SMPL
I5488-01LX5 - 1		942	0.2564	-	7471A	PPB	28 Sep 2017 15:33:35	U	SMPL
I5488-01A - 1		21696	5.3744	-	7471A	PPB	28 Sep 2017 15:35:52	U	SMPL
CCV50 - 1	CCV50	19920	4.9364	-	7471A	PPB	28 Sep 2017 15:38:08	U	SMPL
CCB50 - 1	CCB50	-184	-0.0213	-	7471A	PPB	28 Sep 2017 15:40:25	U	SMPL