


245.1

LB 90406
INSTRUMENT ID: CV1

Linear

 μ Abs.:

32208



A= 0.0000e+000

B= 3.1147e-004

C= 1.5708e-002

Rho= 0.9999621

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	-57	0.000	-57				
0.2	0.200	0.220	0.020	656	0.0 %	656				
2.5	2.500	2.506	0.006	7994	0.0 %	7994				
5.0	5.000	4.986	-0.014	15959	0.0 %	15959				
7.5	7.500	7.442	-0.058	23844	0.0 %	23844				
10.0	10.000	10.048	0.048	32208	0.0 %	32208				

LB90406 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Sep 2017 17:22:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-57	-	0.0000	245.1	PPB	28 Sep 2017 17:26:26	S	Std
0.2 - 1	SO.2	656	-	0.2000	245.1	PPB	28 Sep 2017 17:28:34	S	Std
2.5 - 1	32.5	7994	-	2.5000	245.1	PPB	28 Sep 2017 17:30:43	S	Std
5.0 - 1	55	15959	-	5.0000	245.1	PPB	28 Sep 2017 17:32:51	S	Std
7.5 - 1	57.5	23844	-	7.5000	245.1	PPB	28 Sep 2017 17:34:58	S	Std
10.0 - 1	510	32208	-	10.0000	245.1	PPB	28 Sep 2017 17:37:07	S	Std
ICV81 - 1	ICV81	12809	4.0053	-245.1		PPB	28 Sep 2017 17:48:01	U	SMPL
ICB81 - 1	ICB81	-184	-0.0416	-245.1		PPB	28 Sep 2017 17:50:08	U	SMPL
CCV51 - 1	CCV51	15926	4.9762	-245.1		PPB	28 Sep 2017 17:52:14	U	SMPL
CCB51 - 1	CCB51	-220	-0.0528	-245.1		PPB	28 Sep 2017 17:54:21	U	SMPL
CRA - 1	CRA	641	0.2154	-245.1		PPB	28 Sep 2017 17:56:28	U	SMPL
HighStd - 1	HighStd	31778	9.9136	-245.1		PPB	28 Sep 2017 17:58:34	U	SMPL
ChkStd - 1	ChkStd	21860	6.8244	-245.1		PPB	28 Sep 2017 18:00:41	U	SMPL
PB102722BL - 1	PBW	-256	-0.0640	-245.1		PPB	28 Sep 2017 18:02:50	U	SMPL
PB102722BS - 1	LCSW	13133	4.1062	-245.1		PPB	28 Sep 2017 18:04:59	U	SMPL
I5463-01 - 1	092717-TIDO-E-D-M	-132	-0.0254	-245.1		PPB	28 Sep 2017 18:11:43	U	SMPL
I5463-02 - 1	092717-TIDO-E-U-M	-145	-0.0295	-245.1		PPB	28 Sep 2017 18:13:51	U	SMPL
I5464-01 - 1	092617-TIDO-E-D-M	-160	-0.0341	-245.1		PPB	28 Sep 2017 18:15:59	U	SMPL
I5464-02 - 1	092617-TIDO-E-U-M	-171	-0.0376	-245.1		PPB	28 Sep 2017 18:18:06	U	SMPL
CCV52 - 1	CCV52	15967	4.9890	-245.1		PPB	28 Sep 2017 18:20:12	U	SMPL
CCB52 - 1	CCB52	-218	-0.0522	-245.1		PPB	28 Sep 2017 18:22:19	U	SMPL
I5464-02DUP - 1	092617-TIDO-E-U-MDUP	-142	-0.0285	-245.1		PPB	28 Sep 2017 18:24:26	U	SMPL
I5464-02MS - 1	092617-TIDO-E-U-MMS	13456	4.2069	-245.1		PPB	28 Sep 2017 18:26:33	U	SMPL
I5464-02MSD - 1	092617-TIDO-E-U-MMSD	13548	4.2355	-245.1		PPB	28 Sep 2017 18:28:39	U	SMPL
I5268-08 - 1	PT-HG-WP	66915	20.8577	-245.1		PPB	28 Sep 2017 18:32:53	U	SMPL
I5464-02LX5 - 1		-183	-0.0413	-245.1		PPB	28 Sep 2017 18:41:58	U	SMPL
I5464-02A - 1		13064	4.0848	-245.1		PPB	28 Sep 2017 18:44:06	U	SMPL
CCV53 - 1	CCV53	15804	4.9382	-245.1		PPB	28 Sep 2017 18:46:13	U	SMPL
CCB53 - 1	CCB53	-220	-0.0528	-245.1		PPB	28 Sep 2017 18:48:20	U	SMPL
I5268-08DLX5 - 1	PT-HG-WP	12980	4.0586	-245.1		PPB	28 Sep 2017 18:50:28	U	SMPL
CCV54 - 1	CCV54	15973	4.9908	-245.1		PPB	28 Sep 2017 18:58:56	U	SMPL
CCB54 - 1	CCB54	-210	-0.0497	-245.1		PPB	28 Sep 2017 19:01:03	U	SMPL