



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90411

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/28/2017	16:00
2	CAL2	1	Water			7.01	09/28/2017	16:01
3	CAL3	1	Water			10.02	09/28/2017	16:02
4	ICV	1	Water			7.00	09/28/2017	16:03
5	CCV1	1	Water			2.02	09/28/2017	16:05
6	I5509-03	1	Solid	20.02	20	6.65	09/28/2017	16:06
7	I5511-03	1	Solid	20.03	20	8.51	09/28/2017	16:09
8	I5514-03	1	Solid	20.04	20	8.43	09/28/2017	16:10
9	I5515-03	1	Solid	20.02	20	8.94	09/28/2017	16:11
10	I5515-03DUP	1	Solid	20.03	20	8.95	09/28/2017	16:12
11	CCV2	1	Water			12.02	09/28/2017	16:15

JB90411

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5515

WorkList ID : 103412

Date : 9/28/2017 2:28:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5509-03	Corrosivity	Cool 4 deg C	PSEG01	H52	LEONIA-GENERATOR-COMF	09/28/2017	9045C
	Solid	I5511-03	Corrosivity	Cool 4 deg C	PSEG05	H52	EO-01-092817	09/28/2017	9045C
	Solid	I5514-03	Corrosivity	Cool 4 deg C	PSEG05	H52	SU-03-092817	09/28/2017	9045C
	Solid	I5515-03	Corrosivity	Cool 4 deg C	PSEG05	H52	HD-02-092817	09/28/2017	9045C

Date/Time 09/28/17 2:28 PM
 Received by: JP
 Relinquished by: CB

Date/Time 09/28/17 5:35 PM
 Received by: CP
 Relinquished by: JP