


7470A

LB 90413
INSTRUMENT ID: CXL

Linear

 μ Abs.:

35009



A= 0.0000e+000

B= 2.8800e-004

C= 1.9860e-002

Rho= 0.9998165

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	-26	0.000	-26				
0.2	0.200	0.238	0.038	759	0.0 %	759				
2.5	2.500	2.531	0.031	8718	0.0 %	8718				
5.0	5.000	4.904	-0.096	16958	0.0 %	16958				
7.5	7.500	7.412	-0.088	25669	0.0 %	25669				
10.0	10.000	10.102	0.102	35009	0.0 %	35009				

LB90413 INSTRUMENT ID ; CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Sep 2017 20:07:58

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-26	-	0.0000	7470A	PPB	28 Sep 2017 20:09:07	S	Std
0.2 - 1	50.2	759	-	0.2000	7470A	PPB	28 Sep 2017 20:11:15	S	Std
2.5 - 1	52.5	8718	-	2.5000	7470A	PPB	28 Sep 2017 20:13:24	S	Std
5.0 - 1	55	16958	-	5.0000	7470A	PPB	28 Sep 2017 20:15:32	S	Std
7.5 - 1	57.5	25669	-	7.5000	7470A	PPB	28 Sep 2017 20:17:40	S	Std
10.0 - 1	510	35009	-	10.0000	7470A	PPB	28 Sep 2017 20:19:47	S	Std
ICV82 - 1	ICV82	13370	3.8704	-	7470A	PPB	28 Sep 2017 20:22:33	U	SMPL
ICB82 - 1	ICB82	-154	-0.0245	-	7470A	PPB	28 Sep 2017 20:24:39	U	SMPL
CCV55 - 1	CCV55	17196	4.9723	-	7470A	PPB	28 Sep 2017 20:26:45	U	SMPL
CCB55 - 1	CCB55	-201	-0.0380	-	7470A	PPB	28 Sep 2017 20:28:52	U	SMPL
CRA - 1	CRA	725	0.2287	-	7470A	PPB	28 Sep 2017 20:30:59	U	SMPL
HighStd - 1	HighStd	35270	10.1775	-	7470A	PPB	28 Sep 2017 20:33:05	U	SMPL
ChkStd - 1	ChkStd	23648	6.8304	-	7470A	PPB	28 Sep 2017 20:35:12	U	SMPL
PB102695BL - 1	PBW	-276	-0.0596	-	7470A	PPB	28 Sep 2017 20:37:21	U	SMPL
PB102695BS - 1	LCSW	13047	3.7774	-	7470A	PPB	28 Sep 2017 20:39:29	U	SMPL
I5269-08 - 1	PT-HG-WP	68391	19.7163	-	7470A	PPB	28 Sep 2017 20:48:06	U	SMPL
CCV56 - 1	CCV56	16289	4.7111	-	7470A	PPB	28 Sep 2017 20:50:15	U	SMPL
CCB56 - 1	CCB56	-264	-0.0562	-	7470A	PPB	28 Sep 2017 20:52:24	U	SMPL
I5269-08DLX5 - 1	PT-HG-WP	13867	4.0135	-	7470A	PPB	28 Sep 2017 20:54:32	U	SMPL
CCV57 - 1	CCV57	16929	4.8954	-	7470A	PPB	28 Sep 2017 21:02:59	U	SMPL
CCB57 - 1	CCB57	-194	-0.0360	-	7470A	PPB	28 Sep 2017 21:05:06	U	SMPL