


7470A

LB90427
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35825



A= 0.0000e+000

B= 2.8461e-004

C= 9.1516e-004

Rho= 0.9990101

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	24	0.000	24				
0.2	0.200	0.216	0.016	754	0.0 %	754				
2.5	2.500	2.574	0.074	9042	0.0 %	9042				
5.0	5.000	5.044	0.044	17721	0.0 %	17721				
7.5	7.500	7.161	-0.339	25157	0.0 %	25157				
10.0	10.000	10.197	0.197	35825	0.0 %	35825				

LB90427 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 29 Sep 2017 10:50:42					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7470A	PPB	29 Sep 2017 10:54:25	S	Std
0.2 - 1	50.2	754	-	0.2000	7470A	PPB	29 Sep 2017 10:56:33	S	Std
2.5 - 1	52.5	9042	-	2.5000	7470A	PPB	29 Sep 2017 10:58:42	S	Std
5.0 - 1	55	17721	-	5.0000	7470A	PPB	29 Sep 2017 11:00:50	S	Std
7.5 - 1	57.5	25157	-	7.5000	7470A	PPB	29 Sep 2017 11:02:58	S	Std
10.0 - 1	510	35825	-	10.0000	7470A	PPB	29 Sep 2017 11:07:32	S	Std
ICV83 - 1	ICV83	13073	3.7216	-	7470A	PPB	29 Sep 2017 11:10:17	U	SMPL
ICB83 - 1	ICB83	-165	-0.0460	-	7470A	PPB	29 Sep 2017 11:12:23	U	SMPL
CCV58 - 1	CCV58	16586	4.7215	-	7470A	PPB	29 Sep 2017 11:14:30	U	SMPL
CCB58 - 1	CCB58	-164	-0.0458	-	7470A	PPB	29 Sep 2017 11:16:37	U	SMPL
CRA - 1	CRA	745	0.2129	-	7470A	PPB	29 Sep 2017 11:18:43	U	SMPL
HighStd - 1	HighStd	35743	10.1737	-	7470A	PPB	29 Sep 2017 11:23:36	U	SMPL
ChkStd - 1	ChkStd	23574	6.7103	-	7470A	PPB	29 Sep 2017 11:25:43	U	SMPL
PB102714BL - 1	PBW	-104	-0.0287	-	7470A	PPB	29 Sep 2017 11:29:05	U	SMPL
PB102714BS - 1	LCSW	13189	3.7546	-	7470A	PPB	29 Sep 2017 11:31:12	U	SMPL
PB102714TB - 1	PB102669TB	-156	-0.0435	-	7470A	PPB	29 Sep 2017 11:33:20	U	SMPL
I5450-13 - 1	20170660-A-C-1	-28	-0.0071	-	7470A	PPB	29 Sep 2017 11:35:28	U	SMPL
I5450-14 - 1	20170661-A-C-2	-41	-0.0108	-	7470A	PPB	29 Sep 2017 11:37:36	U	SMPL
I5450-15 - 1	20170662-A-C-3	14	0.0049	-	7470A	PPB	29 Sep 2017 11:39:45	U	SMPL
I5450-16 - 1	20170663-A-C-4	277	0.0798	-	7470A	PPB	29 Sep 2017 11:41:52	U	SMPL
CCV59 - 1	CCV59	17457	4.9693	-	7470A	PPB	29 Sep 2017 11:43:59	U	SMPL
CCB59 - 1	CCB59	-172	-0.0480	-	7470A	PPB	29 Sep 2017 11:46:05	U	SMPL
I5450-17 - 1	20170664-A-C-5	-61	-0.0164	-	7470A	PPB	29 Sep 2017 11:48:13	U	SMPL
I5450-18 - 1	20170665-A-C-6	-26	-0.0065	-	7470A	PPB	29 Sep 2017 11:50:20	U	SMPL
I5450-19 - 1	20170666-A-C-7	-31	-0.0079	-	7470A	PPB	29 Sep 2017 11:52:26	U	SMPL
I5450-20 - 1	20170667-A-C-8	-9	-0.0016	-	7470A	PPB	29 Sep 2017 11:54:33	U	SMPL
I5450-21 - 1	20170668-A-C-9	2	0.0015	-	7470A	PPB	29 Sep 2017 11:56:41	U	SMPL
I5450-22 - 1	20170669-A-C-10	-9	-0.0016	-	7470A	PPB	29 Sep 2017 11:58:48	U	SMPL
I5450-23 - 1	20170670-A-C-11	-29	-0.0073	-	7470A	PPB	29 Sep 2017 12:00:56	U	SMPL
I5450-24 - 1	20170671-A-C-12	-1	0.0006	-	7470A	PPB	29 Sep 2017 12:03:05	U	SMPL
I5472-02 - 1	TP-1-2-3-4-COMP	-33	-0.0085	-	7470A	PPB	29 Sep 2017 12:05:14	U	SMPL
I5473-03 - 1	TR-04-092717	-37	-0.0096	-	7470A	PPB	29 Sep 2017 12:07:21	U	SMPL
CCV60 - 1	CCV60	17347	4.9380	-	7470A	PPB	29 Sep 2017 12:09:27	U	SMPL
CCB60 - 1	CCB60	-135	-0.0375	-	7470A	PPB	29 Sep 2017 12:11:35	U	SMPL
I5475-03 - 1	WS-20408-COMP	-35	-0.0090	-	7470A	PPB	29 Sep 2017 12:13:42	U	SMPL
I5475-03DUP - 1	WS-20408-COMPDUP	-31	-0.0079	-	7470A	PPB	29 Sep 2017 12:15:50	U	SMPL
I5475-03MS - 1	WS-20408-COMPMS	12612	3.5904	-	7470A	PPB	29 Sep 2017 12:17:57	U	SMPL
I5475-03MSD - 1	WS-20408-COMPMSD	12487	3.5548	-	7470A	PPB	29 Sep 2017 12:20:04	U	SMPL
PB102715BL - 1	PBW	-153	-0.0426	-	7470A	PPB	29 Sep 2017 12:22:11	U	SMPL
PB102715BS - 1	LCSW	12722	3.6217	-	7470A	PPB	29 Sep 2017 12:24:19	U	SMPL
PB102715TB - 1	PB102686TB	-159	-0.0443	-	7470A	PPB	29 Sep 2017 12:26:27	U	SMPL
I5478-01 - 1	1612	-69	-0.0187	-	7470A	PPB	29 Sep 2017 12:28:36	U	SMPL
I5481-03 - 1	HR-06-092717	-72	-0.0196	-	7470A	PPB	29 Sep 2017 12:30:44	U	SMPL
I5485-02 - 1	WALDWICK-SWITCH	-50	-0.0133	-	7470A	PPB	29 Sep 2017 12:32:51	U	SMPL
CCV61 - 1	CCV61	17084	4.8632	-	7470A	PPB	29 Sep 2017 12:34:59	U	SMPL
CCB61 - 1	CCB61	-162	-0.0452	-	7470A	PPB	29 Sep 2017 12:37:06	U	SMPL
I5487-03 - 1	CL-02-092717-A	-45	-0.0119	-	7470A	PPB	29 Sep 2017 12:39:14	U	SMPL
I5487-06 - 1	CL-02-092717-B	-33	-0.0085	-	7470A	PPB	29 Sep 2017 12:41:21	U	SMPL
I5487-09 - 1	CL-02-092717-C	-47	-0.0125	-	7470A	PPB	29 Sep 2017 12:43:28	U	SMPL
I5487-12 - 1	CL-02-092717-D	-28	-0.0071	-	7470A	PPB	29 Sep 2017 12:45:35	U	SMPL
I5491-03 - 1	OR-02-092717	-55	-0.0147	-	7470A	PPB	29 Sep 2017 12:47:42	U	SMPL
I5491-03DUP - 1	OR-02-092717DUP	-54	-0.0145	-	7470A	PPB	29 Sep 2017 12:49:50	U	SMPL
I5491-03MS - 1	OR-02-092717MS	12925	3.6795	-	7470A	PPB	29 Sep 2017 12:51:58	U	SMPL
I5491-03MSD - 1	OR-02-092717MSD	12884	3.6678	-	7470A	PPB	29 Sep 2017 12:54:07	U	SMPL
CCV62 - 1	CCV62	17176	4.8894	-	7470A	PPB	29 Sep 2017 12:56:15	U	SMPL
CCB62 - 1	CCB62	-183	-0.0512	-	7470A	PPB	29 Sep 2017 12:58:23	U	SMPL
I5475-03LX5 - 1		-70	-0.0190	-	7470A	PPB	29 Sep 2017 13:00:30	U	SMPL
I5475-03A - 1		14557	4.1440	-	7470A	PPB	29 Sep 2017 13:02:37	U	SMPL
I5491-03LX5 - 1		-146	-0.0406	-	7470A	PPB	29 Sep 2017 13:04:44	U	SMPL
I5491-03A - 1		13875	3.9499	-	7470A	PPB	29 Sep 2017 13:06:51	U	SMPL
CCV63 - 1	CCV63	16649	4.7394	-	7470A	PPB	29 Sep 2017 13:08:58	U	SMPL
CCB63 - 1	CCB63	-158	-0.0441	-	7470A	PPB	29 Sep 2017 13:11:06	U	SMPL