

DATE & TIME IN: 09/29/2017 17:30

OVEN TEMP: 103 °C

Run Number: LB90430

DATE & TIME OUT: 09/30/2017 10:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result %
1	LB90430BL	LB90430BL	75.3293	75.3293	753293	75.3293	75.3293	75.3293	0	0
2	I5442-01	20170660-A-C-1	95.4884	95.4884	100	95.5875	95.5875	95.5875	0.0991	991
3	I5442-01DUP	20170660-A-C-1DUP	84.9763	84.9763	100	85.071	85.071	85.071	0.0947	947
4	I5442-02	20170661-A-C-2	89.2207	89.2207	100	89.3239	89.3239	89.3239	0.1032	1032
5	I5442-03	20170662-A-C-3	85.12	85.12	100	85.235	85.235	85.235	0.115	1150
6	I5442-04	20170663-A-C-4	77.575	77.575	100	77.693	77.693	77.693	0.118	1180
7	I5442-05	20170664-A-C-5	95.1596	95.1596	100	95.2656	95.2656	95.2656	0.106	1060
8	I5442-06	20170665-A-C-6	104.689	104.689	100	104.784	104.784	104.784	0.095	950
9	I5442-07	20170666-A-C-7	74.8002	74.8002	100	74.9055	74.9055	74.9055	0.1053	1053
10	I5442-08	20170667-A-C-8	83.5514	83.5514	100	83.6625	83.6625	83.6625	0.1111	1111
11	I5442-09	20170668-A-C-9	75.551	75.551	100	75.6572	75.6572	75.6572	0.1062	1062
12	I5442-10	20170669-A-C-10	97.4966	97.4966	100	97.5943	97.5943	97.5943	0.0977	977
13	I5442-11	20170670-A-C-11	68.8366	68.8366	100	68.937	68.937	68.937	0.1004	1004
14	I5442-12	20170671-A-C-12	63.8223	63.8223	100	63.9252	63.9252	63.9252	0.1029	1029

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

LB90430

WorkList Name : TS092917

WorkList ID : 103460

Date : 9/29/2017 2:05:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5442-01 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170660-A-C-1	09/20/2017	SM2540 B
	Solid	I5442-02 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170661-A-C-2	09/20/2017	SM2540 B
	Solid	I5442-03 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170662-A-C-3	09/20/2017	SM2540 B
	Solid	I5442-04 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170663-A-C-4	09/20/2017	SM2540 B
	Solid	I5442-05 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170664-A-C-5	09/25/2017	SM2540 B
	Solid	I5442-06 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170665-A-C-6	09/25/2017	SM2540 B
	Solid	I5442-07 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170666-A-C-7	09/25/2017	SM2540 B
	Solid	I5442-08 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170667-A-C-8	09/25/2017	SM2540 B
	Solid	I5442-09 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170668-A-C-9	09/25/2017	SM2540 B
	Solid	I5442-10 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170669-A-C-10	09/25/2017	SM2540 B
	Solid	I5442-11 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170670-A-C-11	09/25/2017	SM2540 B
	Solid	I5442-12 ✓	ASTM TS	Cool 4 deg C	TULL02	H42	20170671-A-C-12	09/25/2017	SM2540 B

Date/Time 09/29/17 17:00

Received by: DP

Relinquished by: CP

Date/Time 09/29/17 17:35

Received by: CP

Relinquished by: DP