


ISM02.4

LB90441
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42274



A= 0.0000e+000

B= 2.3466e-004

C= 4.8346e-003

Rho= 0.9997996

Accept=Accepted

slope
 $y = 1.2162e-02$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	-0.046	-0.046	-217	0.000	-217					-
0.2	0.200	0.221	0.021	920	0.0 %	920					11
2.5	2.500	2.545	0.045	10825	0.0 %	10825					2
5.0	5.000	4.926	-0.074	20970	0.0 %	20970					-1
7.5	7.500	7.630	0.130	32496	0.0 %	32496					2
10.0	10.000	9.925	-0.075	42274	0.0 %	42274					-1

103
9/29/18

LB90441
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-217	-		0 ISM02.4	PPB	9/29/2017 17:03	Std
	0.2 S0.2	920	-		0.2 ISM02.4	PPB	9/29/2017 17:05	Std
	2.5 S2.5	10825	-		2.5 ISM02.4	PPB	9/29/2017 17:08	Std
	5 S5	20970	-		5 ISM02.4	PPB	9/29/2017 17:10	Std
	7.5 S7.5	32496	-		7.5 ISM02.4	PPB	9/29/2017 17:12	Std
	10 S10	42274	-		10 ISM02.4	PPB	9/29/2017 17:15	Std
ICV	ICV	16592	3.8982 -		ISM02.4	PPB	9/29/2017 17:17	SMPL
ICB	ICB	-167	-0.0344 -		ISM02.4	PPB	9/29/2017 17:20	SMPL
CCV067	CCV067	20928	4.9157 -		ISM02.4	PPB	9/29/2017 17:22	SMPL
CCB067	CCB067	-162	-0.0332 -		ISM02.4	PPB	9/29/2017 17:24	SMPL
PB102732BL	PBS004	129	0.0351 -		ISM02.4	PPB	9/29/2017 17:26	SMPL
I5449-01	B-41(0-1)-092617	61234	14.3737 -		ISM02.4	PPB	9/29/2017 17:29	SMPL
I5449-02	B-40(0-1)-092617	32393	7.606 -		ISM02.4	PPB	9/29/2017 17:31	SMPL
I5449-03	B-60(FILL)-092617	185926	43.6333 -		ISM02.4	PPB	9/29/2017 17:33	SMPL
I5449-04	B-60(0-1)-092617	8791	2.0677 -		ISM02.4	PPB	9/29/2017 17:36	SMPL
I5449-05	B-60(5-7)-092617	242984	57.0223 -		ISM02.4	PPB	9/29/2017 17:38	SMPL
PB102733BL	PBS004	-35	-0.0034 -		ISM02.4	PPB	9/29/2017 17:43	SMPL
I5490-01	B-43(0-1)-092617	94214	22.1126 -		ISM02.4	PPB	9/29/2017 17:45	SMPL
I5490-02	B-43(5-7)-092617	31599	7.4197 -		ISM02.4	PPB	9/29/2017 17:48	SMPL
I5490-03	B-70(0-1)-092717	39573	9.2908 -		ISM02.4	PPB	9/29/2017 17:50	SMPL
I5490-04	B-70(5-7)-092717	17311	4.067 -		ISM02.4	PPB	9/29/2017 17:52	SMPL
I5490-05	B-69(0-1)-092717	4852	1.1434 -		ISM02.4	PPB	9/29/2017 17:55	SMPL
I5490-06	B-69(1-3)-092717	12406	2.916 -		ISM02.4	PPB	9/29/2017 17:57	SMPL
I5490-07	B-41(5-7)-092717	4678	1.1026 -		ISM02.4	PPB	9/29/2017 17:59	SMPL
I5490-08	B-40(5-7)-092717	20156	4.7345 -		ISM02.4	PPB	9/29/2017 18:01	SMPL
I5490-09	B-58(1-2)-092717	5346	1.2593 -		ISM02.4	PPB	9/29/2017 18:04	SMPL
I5490-10	B-58(6-8)-092717	4091	0.9648 -		ISM02.4	PPB	9/29/2017 18:06	SMPL
PB102756BL	PBS004	117	0.0323 -		ISM02.4	PPB	9/29/2017 18:08	SMPL
I5533-01	B-56(1-2)-092717	16201	3.8065 -		ISM02.4	PPB	9/29/2017 18:11	SMPL
I5533-02	B-56(2-4)-092717	26904	6.318 -		ISM02.4	PPB	9/29/2017 18:13	SMPL
CCV068	CCV068	21220	4.9842 -		ISM02.4	PPB	9/29/2017 18:15	SMPL
CCB068	CCB068	-187	-0.039 -		ISM02.4	PPB	9/29/2017 18:17	SMPL
I5533-03	B-77(0-1)-092817	25703	6.0362 -		ISM02.4	PPB	9/29/2017 18:20	SMPL
I5533-15	B-77(0-1)-092817D	25176	5.9125 -		ISM02.4	PPB	9/29/2017 18:27	SMPL
I5533-16	B-77(0-1)-092817S	35597	8.3579 -		ISM02.4	PPB	9/29/2017 18:30	SMPL
I5533-06	B-77(1-3)-092817	15266	3.5871 -		ISM02.4	PPB	9/29/2017 18:32	SMPL
I5533-07	DUP-1	26553	6.2356 -		ISM02.4	PPB	9/29/2017 18:34	SMPL
I5533-08	B-54(7-8)-092817	12992	3.0535 -		ISM02.4	PPB	9/29/2017 18:36	SMPL
I5533-09	B-54(1-2)-092817	11203	2.6337 -		ISM02.4	PPB	9/29/2017 18:39	SMPL
I5533-10	B-53(0-1)-092817	15955	3.7488 -		ISM02.4	PPB	9/29/2017 18:41	SMPL
CCV069	CCV069	21104	4.957 -		ISM02.4	PPB	9/29/2017 18:43	SMPL
CCB069	CCB069	158	0.0419 -		ISM02.4	PPB	9/29/2017 18:48	SMPL
I5449-01DLX5	B-41(0-1)-092617	11929	2.804 -		ISM02.4	PPB	9/29/2017 18:51	SMPL
I5449-03DLX10	B-60(FILL)-092617	18909	4.4419 -		ISM02.4	PPB	9/29/2017 18:53	SMPL
I5449-05DLX10	B-60(5-7)-092617	25310	5.944 -		ISM02.4	PPB	9/29/2017 18:55	SMPL
I5490-01DLX5	B-43(0-1)-092617	18058	4.2422 -		ISM02.4	PPB	9/29/2017 18:58	SMPL

LB90441

INSTRUMENT ID : CV1

CCV070	CCV070	21336	5.0114 -	ISM02.4	PPB	9/29/2017 19:03	SMPL
CCB070	CCB070	-157	-0.032 -	ISM02.4	PPB	9/29/2017 19:05	SMPL