



Analytical Summary Report

Analysis Method: 1498

Reviewed By: JIGNESH

Parameter: Redox Potential

Supervisor Review By: apatel

Run Number: LB90445

Reagent/Standard	Lot/Log #
ZOBELLS SOLUTION ORP STD 500ML	W2255

Seq	LabID	ClientID	True value	DL	Initial Wt (g)	Final Vol (mL)	Temp (°C)	Result (mV)	Anal Date	Anal Time
1	LB90445BL	LB90445BL		1			19.4	0.00	09/29/2017	16:40
2	LB90445BS	LB90445BS	228	1			19.1	219.70	09/29/2017	16:42
3	I5543-01	MW-1		1			19.6	206.10	09/29/2017	16:43
4	I5543-01DUP	MW-1DUP		1			19.5	207.20	09/29/2017	16:44
5	I5543-02	MW-3		1			19.2	38.20	09/29/2017	16:45
6	I5543-03	MW-4		1			19.5	24.60	09/29/2017	16:47
7	I5543-04	MW-2		1			19.4	7.60	09/29/2017	16:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : rp I5543

WorkList ID : 103464

Date : 9/29/2017 4:07:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5543-01-H	Redox Potential	Cool 4 deg C	LOCK01	I13	MW-1	09/27/2017	1498
	Water	I5543-02-H	Redox Potential	Cool 4 deg C	LOCK01	I13	MW-3	09/27/2017	1498
	Water	I5543-03-H	Redox Potential	Cool 4 deg C	LOCK01	I13	MW-4	09/27/2017	1498
	Water	I5543-04-H	Redox Potential	Cool 4 deg C	LOCK01	I13	MW-2	09/27/2017	1498

Date/Time 09/29/17 2:06 PM
 Received by: JD
 Relinquished by: JD

Date/Time 09/29/17 6:00 PM
 Received by: JD
 Relinquished by: JD