



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90470

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/29/2017	17:25
2	CAL2	1	Water			7.01	09/29/2017	17:26
3	CAL3	1	Water			10.03	09/29/2017	17:29
4	ICV	1	Water			7.01	09/29/2017	17:30
5	CCV1	1	Water			2.01	09/29/2017	17:32
6	I5546-03	1	Solid	20.02	20	5.25	09/29/2017	17:33
7	I5547-03	1	Solid	20.04	20	8.09	09/29/2017	17:35
8	I5551-03	1	Solid	20.02	20	9.10	09/29/2017	17:36
9	I5551-06	1	Solid	20.05	20	9.13	09/29/2017	17:37
10	I5551-09	1	Solid	20.03	20	8.97	09/29/2017	17:38
11	I5551-09DUP	1	Solid	20.04	20	8.97	09/29/2017	17:39
12	CCV2	1	Water			12.01	09/29/2017	17:40

1390470

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5563 WorkList ID : 103471 Date : 9/29/2017 5:46:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2017	Solid	I5546-03	Corrosivity	Cool 4 deg C	PSEG01	H52	PAINT-CHIPS-DRUMS-COMP	09/29/2017	9045C
	Solid	I5547-03	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-02-092917	09/29/2017	9045C
	Solid	I5551-03	Corrosivity	Cool 4 deg C	PSEG05	H53	JC-03-092917-A	09/29/2017	9045C
	Solid	I5551-06	Corrosivity	Cool 4 deg C	PSEG05	H53	JC-03-092917-B	09/29/2017	9045C
	Solid	I5551-09	Corrosivity	Cool 4 deg C	PSEG05	H53	JC-03-092917-C	09/29/2017	9045C

Date/Time 09/29/17 5:00PM
Received by: TL
Relinquished by: CG

Date/Time 09/29/17 8:10PM
Received by: JP
Relinquished by: JP