

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: DHVANI

Date: 10/02/2017

DATE & TIME IN: 10/02/2017 11:01

OVEN TEMP: 103 °C

Run Number: LB90484

DATE & TIME OUT: 10/02/2017 15:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB90484BL	LB90484BL	1.4065	1.4065	100	1.4065	1.4065	1.4065	0	0
2	LB90484BS	LB90484BS	1.4146	1.4146	100	1.4693	1.4693	1.4693	0.0547	547
3	I5543-01	MW-1	1.394	1.394	500	1.3956	1.3956	1.3956	0.0016	3.2
4	I5543-02	MW-3	1.4043	1.4043	200	1.409	1.409	1.409	0.0047	23.5
5	I5543-02DUP	MW-3DUP	1.3984	1.3984	200	1.4033	1.4033	1.4033	0.0049	24.5
6	I5543-03	MW-4	1.4457	1.4457	500	1.4566	1.4566	1.4566	0.0109	21.8
7	I5543-04	MW-2	1.4038	1.4038	300	1.4163	1.4163	1.4163	0.0125	41.7
8	I5571-01	PAMT-COMP-C	1.4019	1.4019	30	1.4339	1.4339	1.4339	0.032	1066.7
9	I5571-02	I5571-01MS	1.4264	1.4264	30	1.4537	1.4537	1.4537	0.0273	910
10	I5571-03	I5571-01MSD	1.4335	1.4335	30	1.4715	1.4715	1.4715	0.038	1266.7
11	I5571-04	PAMT-COMP-D	1.4396	1.4396	30	1.4711	1.4711	1.4711	0.0315	1050
12	I5571-05	PAMT-COMP-C-D	1.4253	1.4253	960	1.4282	1.4282	1.4282	0.0029	3
13	I5571-06	I5571-05MS	1.4036	1.4036	700	1.4061	1.4061	1.4061	0.0025	3.6
14	I5571-07	I5571-05MSD	1.4254	1.4254	700	1.4282	1.4282	1.4282	0.0028	4
15	I5571-08	PAMT-COMP-D	1.4262	1.4262	700	1.4341	1.4341	1.4341	0.0079	11.3

WORKLIST(Hardcopy Internal Chain)

LB90484

WorkList Name : tss021017

WorkList ID : 103558

Date : 10/2/2017 9:43:47 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5543-01-H	TSS	Cool 4 deg C	LOCK01	I13	MW-1	09/27/2017	SM2540 D
	Water	I5543-02-H	TSS	Cool 4 deg C	LOCK01	I13	MW-3	09/27/2017	SM2540 D
	Water	I5543-03-H	TSS	Cool 4 deg C	LOCK01	I13	MW-4	09/27/2017	SM2540 D
	Water	I5543-04-H	TSS	Cool 4 deg C	LOCK01	I13	MW-2	09/27/2017	SM2540 D
	Water	I5571-01-H	TSS	Cool 4 deg C	ADVA06	H42	PAMT-COMP-C	09/29/2017	SM2540 D
	Water	I5571-02-H	TSS	Cool 4 deg C	ADVA06	H42	I5571-01MS	09/29/2017	SM2540 D
	Water	I5571-03-H	TSS	Cool 4 deg C	ADVA06	H42	I5571-01MSD	09/29/2017	SM2540 D
	Water	I5571-04-H	TSS	Cool 4 deg C	ADVA06	H42	PAMT-COMP-D	09/29/2017	SM2540 D
	Water	I5571-05-H	TSS	Cool 4 deg C	ADVA06	H42	PAMT-COMP-C-D	09/29/2017	SM2540 D
	Water	I5571-06-H	TSS	Cool 4 deg C	ADVA06	H42	I5571-05MS	09/29/2017	SM2540 D
	Water	I5571-07-H	TSS	Cool 4 deg C	ADVA06	H42	I5571-05MSD	09/29/2017	SM2540 D
	Water	I5571-08-H	TSS	Cool 4 deg C	ADVA06	H42	PAMT-COMP-D	09/29/2017	SM2540 D

Date/Time 10/02/17 10:05

Received by: DP

Relinquished by: CP

Date/Time 10/02/17 10:30

Received by: CP

Relinquished by: DP