

7470A


LB 90485

INSTRUMENT ID: CY1

Linear

 μ Abs.:

33912



A= 0.0000e+000

B= 2.9688e-004

C= 2.4529e-002

Rho= 0.9997702

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	8	0.000	8				
0.2	0.200	0.172	-0.028	496	0.0 %	496				
2.5	2.500	2.598	0.098	8667	0.0 %	8667				
5.0	5.000	4.914	-0.086	16469	0.0 %	16469				
7.5	7.500	7.397	-0.103	24834	0.0 %	24834				
10.0	10.000	10.092	0.092	33912	0.0 %	33912				

LB90485 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Oct 2017 11:12:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	8	-	0.0000	7470A	PPB	02 Oct 2017 11:16:38	S	Std
0.2 - 1	80.2	496	-	0.2000	7470A	PPB	02 Oct 2017 11:18:47	S	Std
2.5 - 1	82.5	8667	-	2.5000	7470A	PPB	02 Oct 2017 11:20:56	S	Std
5.0 - 1	85	16469	-	5.0000	7470A	PPB	02 Oct 2017 11:23:04	S	Std
7.5 - 1	87.5	24834	-	7.5000	7470A	PPB	02 Oct 2017 11:25:11	S	Std
10.0 - 1	810	33912	-	10.0000	7470A	PPB	02 Oct 2017 11:29:49	S	Std
ICV84 - 1	ICV84	13066	3.9036	-	7470A	PPB	02 Oct 2017 11:32:19	U	SMPL
ICB84 - 1	ICB84	-148	-0.0194	-	7470A	PPB	02 Oct 2017 11:34:26	U	SMPL
CCV71 - 1	CCV71	16386	4.8893	-	7470A	PPB	02 Oct 2017 11:36:33	U	SMPL
CCB71 - 1	CCB71	-175	-0.0274	-	7470A	PPB	02 Oct 2017 11:38:39	U	SMPL
CRA - 1	CRA	462	0.1617	-	7470A	PPB	02 Oct 2017 11:40:47	U	SMPL
HighStd - 1	HighStd	33593	9.9977	-	7470A	PPB	02 Oct 2017 11:42:53	U	SMPL
ChkStd - 1	ChkStd	23365	6.9612	-	7470A	PPB	02 Oct 2017 11:45:00	U	SMPL
PB102716BL - 1	PBW	-263	-0.0536	-	7470A	PPB	02 Oct 2017 11:47:10	U	SMPL
PB102716BS - 1	LCSW	13277	3.9663	-	7470A	PPB	02 Oct 2017 11:49:19	U	SMPL
I5486-01 - 1	PAMT-SSW	-215	-0.0393	-	7470A	PPB	02 Oct 2017 11:51:26	U	SMPL
I5496-01 - 1	PAMT-COMP-A	556	0.1896	-	7470A	PPB	02 Oct 2017 11:53:34	U	SMPL
I5496-02 - 1	PAMT-COMP-DUP	178	0.0774	-	7470A	PPB	02 Oct 2017 11:55:43	U	SMPL
I5496-03 - 1	PAMT-COMP-B	1297	0.4096	-	7470A	PPB	02 Oct 2017 11:57:52	U	SMPL
I5496-04 - 1	PAMT-COMP-E	5617	1.6921	-	7470A	PPB	02 Oct 2017 11:59:58	U	SMPL
CCV72 - 1	CCV72	16601	4.9531	-	7470A	PPB	02 Oct 2017 12:02:04	U	SMPL
CCB72 - 1	CCB72	-131	-0.0144	-	7470A	PPB	02 Oct 2017 12:04:11	U	SMPL
I5496-05 - 1	PAMT-COMP-A-D	-137	-0.0161	-	7470A	PPB	02 Oct 2017 12:06:18	U	SMPL
I5496-06 - 1	PAMT-COMP-DUP-D	-133	-0.0150	-	7470A	PPB	02 Oct 2017 12:08:25	U	SMPL
I5496-07 - 1	PAMT-COMP-B-D	-99	-0.0049	-	7470A	PPB	02 Oct 2017 12:10:32	U	SMPL
I5496-08 - 1	PAMT-COMP-E-D	-83	-0.0001	-	7470A	PPB	02 Oct 2017 12:12:39	U	SMPL
I5503-01 - 1	S32-0622	-147	-0.0191	-	7470A	PPB	02 Oct 2017 12:14:47	U	SMPL
I5503-01DUP - 1	S32-0622DUP	-151	-0.0203	-	7470A	PPB	02 Oct 2017 12:16:54	U	SMPL
I5503-01MS - 1	S32-0622MS	13107	3.9158	-	7470A	PPB	02 Oct 2017 12:19:03	U	SMPL
I5503-01MSD - 1	S32-0622MSD	13565	4.0518	-	7470A	PPB	02 Oct 2017 12:21:11	U	SMPL
I5503-02 - 1	S32-0595	-226	-0.0426	-	7470A	PPB	02 Oct 2017 12:23:20	U	SMPL
I5503-03 - 1	S32-0597	-111	-0.0084	-	7470A	PPB	02 Oct 2017 12:25:27	U	SMPL
CCV73 - 1	CCV73	16432	4.9029	-	7470A	PPB	02 Oct 2017 12:27:34	U	SMPL
CCB73 - 1	CCB73	-147	-0.0191	-	7470A	PPB	02 Oct 2017 12:29:41	U	SMPL
I5503-04 - 1	S32-0596	-89	-0.0019	-	7470A	PPB	02 Oct 2017 12:31:48	U	SMPL
I5503-05 - 1	S32-0598	-71	0.0035	-	7470A	PPB	02 Oct 2017 12:33:55	U	SMPL
I5503-06 - 1	S32-0599	-77	0.0017	-	7470A	PPB	02 Oct 2017 12:36:02	U	SMPL
PB102755BL - 1	PBW	-31	0.0153	-	7470A	PPB	02 Oct 2017 12:38:09	U	SMPL
PB102755BS - 1	LCSW	11604	3.4696	-	7470A	PPB	02 Oct 2017 12:40:17	U	SMPL
PB102755TB - 1	PB102710TB	-111	-0.0084	-	7470A	PPB	02 Oct 2017 12:42:25	U	SMPL
I5509-03 - 1	LEONIA-GENERATOR-COMP	-30	0.0156	-	7470A	PPB	02 Oct 2017 12:44:33	U	SMPL
I5511-03 - 1	EO-01-092817	-24	0.0174	-	7470A	PPB	02 Oct 2017 12:46:41	U	SMPL
I5514-03 - 1	SU-03-092817	6	0.0263	-	7470A	PPB	02 Oct 2017 12:48:49	U	SMPL
I5514-03DUP - 1	SU-03-092817DUP	36	0.0352	-	7470A	PPB	02 Oct 2017 12:50:56	U	SMPL
CCV74 - 1	CCV74	16634	4.9629	-	7470A	PPB	02 Oct 2017 12:53:04	U	SMPL
CCB74 - 1	CCB74	-164	-0.0242	-	7470A	PPB	02 Oct 2017 12:55:11	U	SMPL
I5514-03MS - 1	SU-03-092817MS	12106	3.6186	-	7470A	PPB	02 Oct 2017 12:57:19	U	SMPL
I5514-03MSD - 1	SU-03-092817MSD	11758	3.5153	-	7470A	PPB	02 Oct 2017 12:59:26	U	SMPL
I5515-03 - 1	HD-02-092817	-125	-0.0126	-	7470A	PPB	02 Oct 2017 13:01:33	U	SMPL
I5503-01LX5 - 1		-68	0.0043	-	7470A	PPB	02 Oct 2017 13:03:40	U	SMPL
I5503-01A - 1		14071	4.2020	-	7470A	PPB	02 Oct 2017 13:05:48	U	SMPL
I5514-03LX5 - 1		-140	-0.0170	-	7470A	PPB	02 Oct 2017 13:07:55	U	SMPL
I5514-03A - 1		14399	4.2994	-	7470A	PPB	02 Oct 2017 13:10:03	U	SMPL
CCV75 - 1	CCV75	16378	4.8869	-	7470A	PPB	02 Oct 2017 13:12:12	U	SMPL
CCB75 - 1	CCB75	-120	-0.0111	-	7470A	PPB	02 Oct 2017 13:14:20	U	SMPL