

LB90487


7471A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39324



A= 0.0000e+000

B= 2.5292e-004

C= 3.3574e-004

Rho= 0.9998696

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	-5	0.000	-5				
0.2	0.200	0.208	0.008	822	0.0 %	822				
2.5	2.500	2.416	-0.084	9550	0.0 %	9550				
5.0	5.000	5.102	0.102	20171	0.0 %	20171				
7.5	7.500	7.529	0.029	29767	0.0 %	29767	0			
10.0	10.000	9.946	-0.054	39324	0.0 %	39324				

LB90487 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 02 Oct 2017 13:45:13

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-5	-	0.0007471A	PPB		02 Oct 2017 13:45:37	S	Std
0.2 - 1	50.2	822	-	0.20007471A	PPB		02 Oct 2017 13:47:55	S	Std
2.5 - 1	52.5	9550	-	2.50007471A	PPB		02 Oct 2017 13:50:13	S	Std
5.0 - 1	55	20171	-	5.00007471A	PPB		02 Oct 2017 13:52:30	S	Std
7.5 - 1	57.5	29767	-	7.50007471A	PPB		02 Oct 2017 13:54:47	S	Std
10.0 - 1	50	39324	-	10.00007471A	PPB		02 Oct 2017 13:57:04	S	Std
ICV85 - 1	ICV85	15515	3.9244	-7471A	PPB		02 Oct 2017 13:59:45	U	SMPL
ICB85 - 1	ICB85	-197	-0.0495	-7471A	PPB		02 Oct 2017 14:02:00	U	SMPL
CCV76 - 1	CCV76	19794	5.0066	-7471A	PPB		02 Oct 2017 14:04:16	U	SMPL
CCB76 - 1	CCB76	-213	-0.0535	-7471A	PPB		02 Oct 2017 14:06:32	U	SMPL
CRA - 1	CRA	731	0.1852	-7471A	PPB		02 Oct 2017 14:08:48	U	SMPL
HighStd - 1	HighStd	39344	9.9511	-7471A	PPB		02 Oct 2017 14:11:04	U	SMPL
ChkStd - 1	ChkStd	26611	6.7307	-7471A	PPB		02 Oct 2017 14:13:20	U	SMPL
PB102767BL - 1	PBS	-258	-0.0649	-7471A	PPB		02 Oct 2017 14:15:39	U	SMPL
PB102767BS - 1	LCSS	14205	3.5930	-7471A	PPB		02 Oct 2017 14:17:56	U	SMPL
I5487-07 - 1	CL-02-092717-C	1313	0.3324	-7471A	PPB		02 Oct 2017 14:20:13	U	SMPL
I5487-10 - 1	CL-02-092717-D	1794	0.4541	-7471A	PPB		02 Oct 2017 14:22:30	U	SMPL
I5509-01 - 1	LEONIA-GENERATOR-COMP	3311	0.8377	-7471A	PPB		02 Oct 2017 14:24:48	U	SMPL
I5511-01 - 1	EO-01-092817	3313	0.8383	-7471A	PPB		02 Oct 2017 14:27:06	U	SMPL
I5514-01 - 1	SU-03-092817	8504	2.1512	-7471A	PPB		02 Oct 2017 14:29:22	U	SMPL
CCV77 - 1	CCV77	19682	4.9783	-7471A	PPB		02 Oct 2017 14:31:37	U	SMPL
CCB77 - 1	CCB77	-209	-0.0525	-7471A	PPB		02 Oct 2017 14:33:54	U	SMPL
I5515-01 - 1	HD-02-092817	926	0.2345	-7471A	PPB		02 Oct 2017 14:36:10	U	SMPL
I5527-01 - 1	STONE-DUST	739	0.1872	-7471A	PPB		02 Oct 2017 14:38:25	U	SMPL
I5531-01 - 1	G-62-0-2.5	1341	0.3395	-7471A	PPB		02 Oct 2017 14:40:42	U	SMPL
I5531-03 - 1	G-62-2.5-5	7029	1.7781	-7471A	PPB		02 Oct 2017 14:42:58	U	SMPL
I5531-05 - 1	G-63-0-2.5	1441	0.3648	-7471A	PPB		02 Oct 2017 14:45:15	U	SMPL
I5531-07 - 1	G-63-2.5-5	12349	3.1236	-7471A	PPB		02 Oct 2017 14:47:32	U	SMPL
I5531-09 - 1	G-62A-0-2.5	1589	0.4022	-7471A	PPB		02 Oct 2017 14:49:50	U	SMPL
I5531-11 - 1	G-62A-2.5-5	3947	0.9986	-7471A	PPB		02 Oct 2017 14:52:07	U	SMPL
I5531-13 - 1	G-63A-0-3	3930	0.9943	-7471A	PPB		02 Oct 2017 14:54:25	U	SMPL
I5531-15 - 1	G-63A-3-6	8958	2.2660	-7471A	PPB		02 Oct 2017 14:56:41	U	SMPL
CCV78 - 1	CCV78	19821	5.0134	-7471A	PPB		02 Oct 2017 14:58:58	U	SMPL
CCB78 - 1	CCB78	-193	-0.0485	-7471A	PPB		02 Oct 2017 15:01:14	U	SMPL
I5531-17 - 1	G-64-0-1.5	10826	2.7384	-7471A	PPB		02 Oct 2017 15:03:31	U	SMPL
I5531-19 - 1	G-64-1.5-3	6249	1.5808	-7471A	PPB		02 Oct 2017 15:05:48	U	SMPL
I5534-01 - 1	SASP2P-ENV-114-1(0-5)	58335	14.7543	-7471A	PPB		02 Oct 2017 15:08:04	U	SMPL
I5534-02 - 1	SASP2P-ENV-114-1(5-10)	12315	3.1150	-7471A	PPB		02 Oct 2017 15:10:20	U	SMPL
I5534-02DUP - 1	SASP2P-ENV-114-1(5-10)DUP	10402	2.6312	-7471A	PPB		02 Oct 2017 15:12:39	U	SMPL
I5534-02MS - 1	SASP2P-ENV-114-1(5-10)MS	26456	6.6915	-7471A	PPB		02 Oct 2017 15:17:32	U	SMPL
I5534-02MSD - 1	SASP2P-ENV-114-1(5-10)MSD	26429	6.6847	-7471A	PPB		02 Oct 2017 15:19:49	U	SMPL
I5534-03 - 1	SASP2P-ENV-114-1(10-16)	108	0.0277	-7471A	PPB		02 Oct 2017 15:22:08	U	SMPL
PB102768BL - 1	PBS	-58	-0.0143	-7471A	PPB		02 Oct 2017 15:24:27	U	SMPL
PB102768BS - 1	LCSS	14851	3.7564	-7471A	PPB		02 Oct 2017 15:26:43	U	SMPL
CCV79 - 1	CCV79	19935	5.0423	-7471A	PPB		02 Oct 2017 15:29:00	U	SMPL
CCB79 - 1	CCB79	-231	-0.0581	-7471A	PPB		02 Oct 2017 15:31:16	U	SMPL
I5532-01 - 1	G-64A-0-1.5	7081	1.7912	-7471A	PPB		02 Oct 2017 15:33:34	U	SMPL
I5532-03 - 1	G-64A-1.5-3	8101	2.0492	-7471A	PPB		02 Oct 2017 15:35:51	U	SMPL
I5532-05 - 1	G-65-0-1	9534	2.4117	-7471A	PPB		02 Oct 2017 15:38:07	U	SMPL
I5532-07 - 1	G-65-1-2	10718	2.7111	-7471A	PPB		02 Oct 2017 15:40:23	U	SMPL
I5532-09 - 1	G-66-0-1	6576	1.6635	-7471A	PPB		02 Oct 2017 15:42:40	U	SMPL
I5532-09DUP - 1	G-66-0-1DUP	7261	1.8368	-7471A	PPB		02 Oct 2017 15:44:57	U	SMPL
I5532-09MS - 1	G-66-0-1MS	20901	5.2866	-7471A	PPB		02 Oct 2017 15:47:14	U	SMPL
I5532-09MSD - 1	G-66-0-1MSD	22338	5.6500	-7471A	PPB		02 Oct 2017 15:49:32	U	SMPL
I5532-11 - 1	G-66-1-2	2599	0.6577	-7471A	PPB		02 Oct 2017 15:51:50	U	SMPL
I5546-01 - 1	PAINT-CHIPS-DRUMS-COMP	771	0.1953	-7471A	PPB		02 Oct 2017 15:54:06	U	SMPL
CCV80 - 1	CCV80	20502	5.1857	-7471A	PPB		02 Oct 2017 15:56:22	U	SMPL
CCB80 - 1	CCB80	-214	-0.0538	-7471A	PPB		02 Oct 2017 15:58:39	U	SMPL
I5547-01 - 1	CL-02-092917	14624	3.6990	-7471A	PPB		02 Oct 2017 16:00:55	U	SMPL
I5551-01 - 1	JC-03-092917-A	8090	2.0464	-7471A	PPB		02 Oct 2017 16:03:11	U	SMPL
I5551-04 - 1	JC-03-092917-B	19962	5.0491	-7471A	PPB		02 Oct 2017 16:05:28	U	SMPL
I5551-07 - 1	JC-03-092917-C	7696	1.9468	-7471A	PPB		02 Oct 2017 16:07:44	U	SMPL
I5554-01 - 1	BRICK-FOR-WASTE	30108	7.6152	-7471A	PPB		02 Oct 2017 16:10:00	U	SMPL
I5563-01 - 1	NYSEG-PH4-SB-ESMI-1	15321	3.8753	-7471A	PPB		02 Oct 2017 16:12:17	U	SMPL
I5563-02 - 1	NYSEG-PH4-SB-ESMI-2	36314	9.1848	-7471A	PPB		02 Oct 2017 16:14:36	U	SMPL
I5563-03 - 1	NYSEG-PH4-SB-ESMI-3	7051	1.7837	-7471A	PPB		02 Oct 2017 16:16:53	U	SMPL
I5563-04 - 1	NYSEG-PH4-SB-ESMI-4	132347	33.4733	-7471A	PPB		02 Oct 2017 16:19:13	U	SMPL
I5563-05 - 1	NYSEG-PH4-SB-ESMI-5	76791	19.4222	-7471A	PPB		02 Oct 2017 16:21:29	U	SMPL
CCV81 - 1	CCV81	19798	5.0076	-7471A	PPB		02 Oct 2017 16:23:53	U	SMPL
CCB81 - 1	CCB81	-266	-0.0669	-7471A	PPB		02 Oct 2017 16:26:12	U	SMPL
I5563-06 - 1	NYSEG-PH4-SB-ESMI-6	32185	8.1405	-7471A	PPB		02 Oct 2017 16:28:31	U	SMPL
I5563-07 - 1	NYSEG-PH4-SB-ESMI-7	73581	18.6103	-7471A	PPB		02 Oct 2017 16:30:48	U	SMPL
I5563-08 - 1	NYSEG-PH4-SB-ESMI-8	56298	14.2391	-7471A	PPB		02 Oct 2017 16:33:06	U	SMPL
I5534-02LX5 - 1		1803	0.4563	-7471A	PPB		02 Oct 2017 16:35:25	U	SMPL
I5534-02A - 1		28211	7.1354	-7471A	PPB		02 Oct 2017 16:37:44	U	SMPL
I5532-09LX5 - 1		962	0.2436	-7471A	PPB		02 Oct 2017 16:40:01	U	SMPL
I5532-09A - 1		23281	5.8885	-7471A	PPB		02 Oct 2017 16:42:19	U	SMPL
I5534-01DLX5 - 1		10510	2.6585	-7471A	PPB		02 Oct 2017 16:44:40	U	SMPL
CCV82 - 1	CCV82	20490	5.1826	-7471A	PPB		02 Oct 2017 16:46:59	U	SMPL
CCB82 - 1	CCB82	-211	-0.0530	-7471A	PPB		02 Oct 2017 16:49:16	U	SMPL
I5563-04DLX10 - 1		13006	3.2898	-7471A	PPB		02 Oct 2017 16:51:33	U	SMPL
I5563-05DLX5 - 1		14902	3.7693	-7471A	PPB		02 Oct 2017 16:53:50	U	SMPL
I5563-07DLX5 - 1		13980	3.5361	-7471A	PPB		02 Oct 2017 16:56:07	U	SMPL
I5563-08DLX5 - 1		10838	2.7415	-7471A	PPB		02 Oct 2017 16:58:24	U	SMPL
CCV83 - 1	CCV83	19987	5.0554	-7471A	PPB		02 Oct 2017 17:00:41	U	SMPL
CCB83 - 1	CCB83	-200	-0.0502	-7471A	PPB		02 Oct 2017 17:02:58	U	SMPL