



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By :

Run Number: LB90493

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/02/2017	14:30
2	CAL2	1	Water			7.01	10/02/2017	14:31
3	CAL3	1	Water			10.03	10/02/2017	14:32
4	ICV	1	Water			6.99	10/02/2017	14:35
5	CCV1	1	Water			2.01	10/02/2017	14:36
6	I5328-03	1	Solid	20.02	20	5.57	10/02/2017	14:37
7	I5328-03DUP	1	Solid	20.04	20	5.58	10/02/2017	14:38
8	CCV2	1	Water			12.01	10/02/2017	14:40

1390493

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5328

WorkList ID : 103578

Date : 10/2/2017 2:25:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/20/2017	Solid	I5328-03	pH	Cool 4 deg C	JERS02	JMP-TOPSOIL	JMP-TOPSOIL	09/20/2017	9045C

Date/Time

Received by:

Relinquished by:

10/02/17 2:25 PM
JD
CD

Date/Time

Received by:

Relinquished by:

10/02/17 3:35 PM
CD
JD