



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90499

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/02/2017	17:00
2	CAL2	1	Water			7.01	10/02/2017	17:01
3	CAL3	1	Water			10.03	10/02/2017	17:03
4	ICV	1	Water			7.00	10/02/2017	17:05
5	CCV1	1	Water			2.01	10/02/2017	17:06
6	I5598-01	1	Solid	20.02	20	5.46	10/02/2017	17:08
7	I5598-01DUP	1	Solid	20.04	20	5.47	10/02/2017	17:09
8	I5601-01	1	Solid	20.01	20	5.64	10/02/2017	17:11
9	I5605-01	1	Solid	20.04	20	5.86	10/02/2017	17:13
10	CCV2	1	Water			12.02	10/02/2017	17:15

2390499

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5601

WorkList ID : 103582

Date : 10/2/2017 4:50:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5598-01 - 1A	pH	Cool 4 deg C	TULL02	I12	FK-PS-01-017	10/02/2017	9045C
	Solid	I5601-01 - 1A	pH	Cool 4 deg C	TULL02	I12	FK-PS-01-016	10/02/2017	9045C
	Solid	I5605-01 - 1A	pH	Cool 4 deg C	TULL02	I11	FK-PS-01-015	10/02/2017	9045C

Date/Time 10/02/17 4:50 PM
Received by: ZP
Relinquished by: CP

Date/Time 10/02/17 6:00 PM
Received by: CP
Relinquished by: ZP