



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB90509

ANALYST: Priscilla
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: 0.0001 Slope: 0.7771 Regression: 0.999974

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Priscilla

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB90509

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		1.66	0.000	0.394	0.394	0.507	10/02/2017	16:05
2	ICB		1	100	100		1.85	0.000	0.000	0.000	0.000	10/02/2017	16:06
3	CCV1	0.5	1	100	100		2.26	0.000	0.395	0.395	0.508	10/02/2017	16:06
4	CCB1		1	100	100		2.22	0.000	0.000	0.000	0.000	10/02/2017	16:07
5	PB102813BL		1	2.55	100	7.30	1.67	0.000	0.000	0.000	0.000	10/02/2017	16:07
6	PB102813BS	20	1	2.53	100	7.41	1.85	0.001	0.415	0.414	0.533	10/02/2017	16:08
7	I5546-01		1	2.50	100	7.84	1.96	0.208	2.988	2.780	3.577	10/02/2017	16:08
8	I5547-01		1	2.50	100	7.51	2.22	0.018	0.018	0.000	0.000	10/02/2017	16:09
9	I5551-01		1	2.50	100	7.98	2.16	0.012	0.012	0.000	0.000	10/02/2017	16:09
10	I5551-01DUP		1	2.50	100	7.62	2.03	0.013	0.013	0.000	0.000	10/02/2017	16:10
11	I5551-01MSPre	40	2	2.53	100	7.33	1.95	0.008	0.130	0.122	0.157	10/02/2017	16:10
12	I5551-01MS2Ins	1284	40	2.53	100	7.11	2.06	0.000	0.642	0.642	0.826	10/02/2017	16:11
13	I5551-01MS3Post	40	2	2.54	100	7.52	1.98	0.013	0.335	0.322	0.414	10/02/2017	16:11
14	I5551-04		1	2.53	100	7.77	1.64	0.009	0.009	0.000	0.000	10/02/2017	16:12
15	CCV2	0.5	1	100	100		1.72	0.000	0.394	0.394	0.507	10/02/2017	16:12
16	CCB2		1	100	100		2.36	0.000	0.000	0.000	0.000	10/02/2017	16:13
17	I5551-07		1	2.54	100	7.38	2.45	0.009	0.009	0.000	0.000	10/02/2017	16:13
18	I5584-01		1	2.55	100	7.16	1.64	0.004	0.005	0.001	0.001	10/02/2017	16:14
19	I5585-01		1	2.55	100	7.91	2.07	0.036	0.037	0.001	0.001	10/02/2017	16:14
20	I5585-03		1	2.53	100	7.54	1.78	0.018	0.019	0.001	0.001	10/02/2017	16:15
21	I5585-05		1	2.55	100	7.25	2.09	0.044	0.045	0.001	0.001	10/02/2017	16:15
22	I5585-07		1	2.52	100	7.26	2.24	0.008	0.008	0.000	0.000	10/02/2017	16:16
23	I5585-09		1	2.53	100	7.83	2.36	0.032	0.033	0.001	0.001	10/02/2017	16:16
24	I5585-11		1	2.55	100	7.41	2.33	0.009	0.009	0.000	0.000	10/02/2017	16:17
25	I5585-13		1	2.55	100	7.34	1.95	0.016	0.016	0.000	0.000	10/02/2017	16:17
26	I5585-15		1	2.55	100	7.28	1.69	0.019	0.019	0.000	0.000	10/02/2017	16:18
27	CCV3	0.5	1	100	100		1.59	0.000	0.394	0.394	0.507	10/02/2017	16:18
28	CCB3		1	100	100		2.45	0.000	0.000	0.000	0.000	10/02/2017	16:19
29	I5585-17		1	2.54	100	7.83	1.95	0.013	0.014	0.001	0.001	10/02/2017	16:19
30	I5585-19		1	2.55	100	7.63	2.08	0.004	0.004	0.000	0.000	10/02/2017	16:20
31	I5546-01		10	2.50	100	7.84	1.96	0.007	0.333	0.326	0.419	10/02/2017	16:20
32	CCV4	0.5	1	100	100		2.15	0.000	0.394	0.394	0.507	10/02/2017	16:21
33	CCB4		1	100	100		1.67	0.000	0.000	0.000	0.000	10/02/2017	16:21