


7471A

LB90516
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4138E



A= 0.0000e+000

B= 2.4258e-004

C= 2.4397e-002

Rho= 0.9999264

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	30	0.000	30				
0.2	0.200	0.210	0.010	766	0.0 %	766				
2.5	2.500	2.507	0.007	10235	0.0 %	10235				
5.0	5.000	4.929	-0.071	20220	0.0 %	20220				
7.5	7.500	7.458	-0.042	30642	0.0 %	30642	0			
10.0	10.000	10.064	0.064	41386	0.0 %	41386				

LB90516 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 03 Oct 2017 11:54:46				
Sample ID	Extended ID	µ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	30	-	0.00007471A	PPB	03 Oct 2017 12:10:36	S	Std
0.2 - 1	50.2	766	-	0.20007471A	PPB	03 Oct 2017 12:12:54	S	Std
2.5 - 1	52.5	10235	-	2.50007471A	PPB	03 Oct 2017 12:15:12	S	Std
5.0 - 1	55	20220	-	5.00007471A	PPB	03 Oct 2017 12:17:29	S	Std
7.5 - 1	57.5	30642	-	7.50007471A	PPB	03 Oct 2017 12:19:47	S	Std
10.0 - 1	59	41386	-	10.00007471A	PPB	03 Oct 2017 12:24:42	S	Std
ICV86 - 1	ICV86	16277	3.9729	-7471A	PPB	03 Oct 2017 12:27:21	U	SMPL
ICB86 - 1	ICB86	-118	-0.0042	-7471A	PPB	03 Oct 2017 12:29:36	U	SMPL
CCV86 - 1	CCV86	19849	4.8394	-7471A	PPB	03 Oct 2017 12:31:53	U	SMPL
CCB86 - 1	CCB86	-89	0.0028	-7471A	PPB	03 Oct 2017 12:34:09	U	SMPL
CRA - 1	CRA	796	0.2175	-7471A	PPB	03 Oct 2017 12:36:24	U	SMPL
HighStd - 1	HighStd	41412	10.0702	-7471A	PPB	03 Oct 2017 12:38:40	U	SMPL
ChkStd - 1	ChkStd	28624	6.9681	-7471A	PPB	03 Oct 2017 12:40:56	U	SMPL
PB102846BL - 1	PBS	-241	-0.0341	-7471A	PPB	03 Oct 2017 12:43:13	U	SMPL
PB102846BS - 1	LCSS	14415	3.5212	-7471A	PPB	03 Oct 2017 12:45:32	U	SMPL
I5432-01 - 1	FK-BPM-01-E7-E8-F7-F8	953	0.2556	-7471A	PPB	03 Oct 2017 12:47:49	U	SMPL
I5584-01 - 1	TP-1-COMP	1089	0.2886	-7471A	PPB	03 Oct 2017 12:50:06	U	SMPL
I5587-01 - 1	G28A-0-2	1069	0.2837	-7471A	PPB	03 Oct 2017 12:52:23	U	SMPL
I5587-03 - 1	G28A-2-4	2495	0.6296	-7471A	PPB	03 Oct 2017 12:54:41	U	SMPL
I5587-05 - 1	G10-0-1.5	8453	2.0749	-7471A	PPB	03 Oct 2017 12:56:57	U	SMPL
CCV87 - 1	CCV87	19455	4.7438	-7471A	PPB	03 Oct 2017 12:59:13	U	SMPL
CCB87 - 1	CCB87	-143	-0.0103	-7471A	PPB	03 Oct 2017 13:01:29	U	SMPL
I5587-07 - 1	G10-1.5-3	8156	2.0029	-7471A	PPB	03 Oct 2017 13:03:46	U	SMPL
I5588-01 - 1	SASP2P-EPE-115-4B(10-12)	143	0.0591	-7471A	PPB	03 Oct 2017 13:06:01	U	SMPL
I5588-02 - 1	SASP2P-ENV-114-4(0-5)	10471	2.5645	-7471A	PPB	03 Oct 2017 13:08:18	U	SMPL
I5588-03 - 1	SASP2P-EPE-T8(0-5)	172	0.0661	-7471A	PPB	03 Oct 2017 13:10:34	U	SMPL
I5588-04 - 1	SASP2P-EPE-T8(5-10)	1015	0.2706	-7471A	PPB	03 Oct 2017 13:12:51	U	SMPL
I5589-01 - 1	SASP2P-TP-106-T(0-5)	2959	0.7422	-7471A	PPB	03 Oct 2017 13:15:08	U	SMPL
I5598-01 - 1	FK-PS-01-017	531	0.1532	-7471A	PPB	03 Oct 2017 13:17:25	U	SMPL
I5601-01 - 1	FK-PS-01-016	332	0.1049	-7471A	PPB	03 Oct 2017 13:19:43	U	SMPL
I5604-01 - 1	SP-A-100217	4576	1.1345	-7471A	PPB	03 Oct 2017 13:22:00	U	SMPL
I5604-01DUP - 1	SP-A-100217DUP	4920	1.2179	-7471A	PPB	03 Oct 2017 13:24:17	U	SMPL
CCV88 - 1	CCV88	19537	4.7637	-7471A	PPB	03 Oct 2017 13:26:33	U	SMPL
CCB88 - 1	CCB88	-129	-0.0069	-7471A	PPB	03 Oct 2017 13:28:49	U	SMPL
I5604-01MS - 1	SP-A-100217MS	17749	4.3300	-7471A	PPB	03 Oct 2017 13:31:06	U	SMPL
I5604-01MSD - 1	SP-A-100217MSD	18045	4.4018	-7471A	PPB	03 Oct 2017 13:33:23	U	SMPL
I5604-04 - 1	SP-B-100217	16560	4.0416	-7471A	PPB	03 Oct 2017 13:35:40	U	SMPL
I5605-01 - 1	FK-PS-01-015	237	0.0819	-7471A	PPB	03 Oct 2017 13:37:56	U	SMPL
I5610-01 - 1	1A-(1-2)-COMP-(0-5)	18783	4.5808	-7471A	PPB	03 Oct 2017 13:40:13	U	SMPL
I5610-02 - 1	1B-(1-2)-COMP-(0-5)	24081	5.8660	-7471A	PPB	03 Oct 2017 13:42:30	U	SMPL
I5610-03 - 1	2A-(1-2)-COMP-(0-5)	22570	5.4995	-7471A	PPB	03 Oct 2017 13:44:47	U	SMPL
I5610-04 - 1	2B-(1-2)-COMP-(0-5)	26841	6.5355	-7471A	PPB	03 Oct 2017 13:47:06	U	SMPL
PB102847BL - 1	PBS	-236	-0.0329	-7471A	PPB	03 Oct 2017 13:49:25	U	SMPL
PB102847BS - 1	LCSS	14609	3.5683	-7471A	PPB	03 Oct 2017 13:51:42	U	SMPL
CCV89 - 1	CCV89	19328	4.7130	-7471A	PPB	03 Oct 2017 13:53:59	U	SMPL
CCB89 - 1	CCB89	-145	-0.0108	-7471A	PPB	03 Oct 2017 13:56:15	U	SMPL
I5585-01 - 1	G67-0-1	5110	1.2640	-7471A	PPB	03 Oct 2017 13:58:32	U	SMPL
I5585-03 - 1	G67-1-2	6955	1.7116	-7471A	PPB	03 Oct 2017 14:00:48	U	SMPL
I5585-05 - 1	G68-0-1.5	6153	1.5170	-7471A	PPB	03 Oct 2017 14:03:04	U	SMPL
I5585-07 - 1	G68-1.5-3	1901	0.4855	-7471A	PPB	03 Oct 2017 14:05:21	U	SMPL
I5585-09 - 1	G69-0-1.5	7268	1.7875	-7471A	PPB	03 Oct 2017 14:07:37	U	SMPL
I5585-11 - 1	G69-1.5-3	278	0.0918	-7471A	PPB	03 Oct 2017 14:09:54	U	SMPL
I5585-13 - 1	G70-0-2.5	3552	0.8860	-7471A	PPB	03 Oct 2017 14:12:11	U	SMPL
I5585-15 - 1	G70-2.5-5	2219	0.5627	-7471A	PPB	03 Oct 2017 14:14:29	U	SMPL
I5585-17 - 1	G71-0-1.5	2636	0.6638	-7471A	PPB	03 Oct 2017 14:16:47	U	SMPL
I5585-19 - 1	G71-1.5-3	3289	0.8222	-7471A	PPB	03 Oct 2017 14:19:03	U	SMPL
CCV90 - 1	CCV90	20107	4.9020	-7471A	PPB	03 Oct 2017 14:21:19	U	SMPL
CCB90 - 1	CCB90	-146	-0.0110	-7471A	PPB	03 Oct 2017 14:23:36	U	SMPL
I5586-01 - 1	G72-0-1	3411	0.8518	-7471A	PPB	03 Oct 2017 14:25:52	U	SMPL
I5586-01DUP - 1	G72-0-1DUP	3322	0.8303	-7471A	PPB	03 Oct 2017 14:28:09	U	SMPL
I5586-01MS - 1	G72-0-1MS	17980	4.3860	-7471A	PPB	03 Oct 2017 14:30:25	U	SMPL
I5586-01MSD - 1	G72-0-1MSD	17732	4.3259	-7471A	PPB	03 Oct 2017 14:32:42	U	SMPL
I5586-03 - 1	G72-1-2	2605	0.6563	-7471A	PPB	03 Oct 2017 14:34:58	U	SMPL
I5586-05 - 1	G47A-0-2.5	3765	0.9377	-7471A	PPB	03 Oct 2017 14:37:16	U	SMPL
I5586-07 - 1	G47A-2.5-5	5389	1.3317	-7471A	PPB	03 Oct 2017 14:39:33	U	SMPL
I5586-09 - 1	G54A-0-3.5	7826	1.9228	-7471A	PPB	03 Oct 2017 14:41:51	U	SMPL
I5586-11 - 1	G54A-3.5-7	7349	1.8071	-7471A	PPB	03 Oct 2017 14:44:08	U	SMPL
I5586-13 - 1	G54B-0-3	4334	1.0757	-7471A	PPB	03 Oct 2017 14:46:25	U	SMPL
CCV91 - 1	CCV91	19980	4.8712	-7471A	PPB	03 Oct 2017 14:48:42	U	SMPL
CCB91 - 1	CCB91	-146	-0.0110	-7471A	PPB	03 Oct 2017 14:50:59	U	SMPL
I5586-15 - 1	G54B-3-6	31041	7.5544	-7471A	PPB	03 Oct 2017 14:53:15	U	SMPL
I5586-17 - 1	G56A-0-3	4018	0.9991	-7471A	PPB	03 Oct 2017 14:55:32	U	SMPL
I5586-19 - 1	G56A-3-6	7243	1.7814	-7471A	PPB	03 Oct 2017 14:57:50	U	SMPL
I5604-01LX5 - 1		824	0.2243	-7471A	PPB	03 Oct 2017 15:00:06	U	SMPL
I5604-01A - 1		21428	5.2224	-7471A	PPB	03 Oct 2017 15:02:23	U	SMPL
I5586-01LX5 - 1		440	0.1311	-7471A	PPB	03 Oct 2017 15:04:40	U	SMPL
I5586-01A - 1		19665	4.7948	-7471A	PPB	03 Oct 2017 15:06:58	U	SMPL
CCV92 - 1	CCV92	19889	4.8491	-7471A	PPB	03 Oct 2017 15:09:15	U	SMPL
CCB92 - 1	CCB92	-204	-0.0251	-7471A	PPB	03 Oct 2017 15:11:33	U	SMPL