

7470A


LB 90523

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35238



A= 0.0000e+000

B= 2.8371e-004

C= -1.6921e-002

Rho= 0.9999431

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	82	0.000	82				
0.2	0.200	0.196	-0.004	751	0.0 %	751				
2.5	2.500	2.522	0.022	8948	0.0 %	8948				
5.0	5.000	4.933	-0.067	17447	0.0 %	17447				
7.5	7.500	7.562	0.062	26715	0.0 %	26715				
10.0	10.000	9.980	-0.020	35238	0.0 %	35238				

LB90523 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Oct 2017 15:20:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	82	-	0.0000	7470A	PPB	03 Oct 2017 15:20:55	S	Std
0.2 - 1	50.2	751	-	0.2000	7470A	PPB	03 Oct 2017 15:23:03	S	Std
2.5 - 1	52.5	8948	-	2.5000	7470A	PPB	03 Oct 2017 15:25:12	S	Std
5.0 - 1	55	17447	-	5.0000	7470A	PPB	03 Oct 2017 15:27:20	S	Std
7.5 - 1	57.5	26715	-	7.5000	7470A	PPB	03 Oct 2017 15:29:28	S	Std
10.0 - 1	51.1	35238	-	10.0000	7470A	PPB	03 Oct 2017 15:31:35	S	Std
ICV87 - 1	ICV87	13317	3.7612	-	7470A	PPB	03 Oct 2017 15:39:06	U	SMPL
ICB87 - 1	ICB87	-79	-0.0393	-	7470A	PPB	03 Oct 2017 15:41:12	U	SMPL
CCV93 - 1	CCV93	17157	4.8507	-	7470A	PPB	03 Oct 2017 15:43:19	U	SMPL
CCB93 - 1	CCB93	-100	-0.0453	-	7470A	PPB	03 Oct 2017 15:45:25	U	SMPL
CRA - 1	CRA	685	0.1774	-	7470A	PPB	03 Oct 2017 15:47:32	U	SMPL
HighStd - 1	HighStd	35003	9.9138	-	7470A	PPB	03 Oct 2017 15:49:39	U	SMPL
ChkStd - 1	ChkStd	24536	6.9442	-	7470A	PPB	03 Oct 2017 15:51:45	U	SMPL
PB102878BL - 1	PBW	-244	-0.0861	-	7470A	PPB	03 Oct 2017 15:53:54	U	SMPL
PB102878BS - 1	LCSW	12576	3.5510	-	7470A	PPB	03 Oct 2017 15:56:03	U	SMPL
PB102878TB - 1	PB102748TB	-198	-0.0731	-	7470A	PPB	03 Oct 2017 15:58:10	U	SMPL
I5527-02 - 1	STONE-DUST	-99	-0.0450	-	7470A	PPB	03 Oct 2017 16:00:18	U	SMPL
I5527-02DUP - 1	STONE-DUSTDUP	-65	-0.0354	-	7470A	PPB	03 Oct 2017 16:02:27	U	SMPL
I5527-02MS - 1	STONE-DUSTMS	13044	3.6838	-	7470A	PPB	03 Oct 2017 16:04:35	U	SMPL
I5527-02MSD - 1	STONE-DUSTMSD	12601	3.5581	-	7470A	PPB	03 Oct 2017 16:06:42	U	SMPL
CCV94 - 1	CCV94	17706	5.0064	-	7470A	PPB	03 Oct 2017 16:08:49	U	SMPL
CCB94 - 1	CCB94	-123	-0.0518	-	7470A	PPB	03 Oct 2017 16:10:56	U	SMPL
I5527-02LX5 - 1		-21	-0.0229	-	7470A	PPB	03 Oct 2017 16:13:04	U	SMPL
I5527-02A - 1		13474	3.8058	-	7470A	PPB	03 Oct 2017 16:15:10	U	SMPL
CCV95 - 1	CCV95	17656	4.9923	-	7470A	PPB	03 Oct 2017 16:17:17	U	SMPL
CCB95 - 1	CCB95	-107	-0.0473	-	7470A	PPB	03 Oct 2017 16:19:25	U	SMPL