



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90535

Temp.: 21.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/03/2017	16:40
2	CAL2	1	Water			7.01	10/03/2017	16:41
3	CAL3	1	Water			10.04	10/03/2017	16:42
4	ICV	1	Water			7.00	10/03/2017	16:43
5	CCV1	1	Water			2.01	10/03/2017	16:44
6	I5623-04	1	Water			6.59	10/03/2017	16:45
7	I5631-04	1	Water			6.73	10/03/2017	16:46
8	I5631-04DUP	1	Water			6.74	10/03/2017	16:47
9	CCV2	1	Water			12.01	10/03/2017	16:49

2090535

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I5623

WorkList ID : 103646

Date : 10/3/2017 6:04:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5623-04	Corrosivity	Cool 4 deg C	ATCE02	J11	170911BSEU-04-LCC-WW	09/11/2017	9040C
	Water	I5631-04	Corrosivity	Cool 4 deg C	ATCE02		170917GCTT42-04-LCC-WW	09/17/2017	9040C

Date/Time

Received by:

Relinquished by:

10/03/17 6:04 PM

JH

JP

Date/Time

Received by:

Relinquished by:

10/03/17 7:30 PM

JH

JP