


7470A

LB 90548
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

34045



A= 0.0000e+000

B= 2.9315e-004

C= -2.2674e-002

Rho= 0.9999542

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	44	0.000	44				
0.2	0.200	0.186	-0.014	713	0.0 %	713				
2.5	2.500	2.527	0.027	8699	0.0 %	8699				
5.0	5.000	4.975	-0.025	17048	0.0 %	17048				
7.5	7.500	7.563	0.063	25878	0.0 %	25878				
10.0	10.000	9.958	-0.042	34045	0.0 %	34045				

LB90548 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Oct 2017 09:58:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	44	-	0.0000	7470A	PPB	04 Oct 2017 10:22:08	S	Std
0.2 - 1	50.2	713	-	0.2000	7470A	PPB	04 Oct 2017 10:24:16	S	Std
2.5 - 1	52.5	8699	-	2.5000	7470A	PPB	04 Oct 2017 10:26:25	S	Std
5.0 - 1	55	17048	-	5.0000	7470A	PPB	04 Oct 2017 10:28:33	S	Std
7.5 - 1	57.5	25878	-	7.5000	7470A	PPB	04 Oct 2017 10:30:40	S	Std
10.0 - 1	510	34045	-	10.0000	7470A	PPB	04 Oct 2017 10:32:48	S	Std
ICV89 - 1	ICV89	13297	3.8753	-	7470A	PPB	04 Oct 2017 10:35:31	U	SMPL
ICB89 - 1	ICB89	-20	-0.0285	-	7470A	PPB	04 Oct 2017 10:37:37	U	SMPL
CCV08 - 1	CCV08	16615	4.8480	-	7470A	PPB	04 Oct 2017 10:39:44	U	SMPL
CCB08 - 1	CCB08	58	-0.0057	-	7470A	PPB	04 Oct 2017 10:41:50	U	SMPL
CRA - 1	CRA	702	0.1831	-	7470A	PPB	04 Oct 2017 10:43:57	U	SMPL
HighStd - 1	HighStd	34528	10.0992	-	7470A	PPB	04 Oct 2017 10:46:03	U	SMPL
ChkStd - 1	ChkStd	24830	7.2562	-	7470A	PPB	04 Oct 2017 10:48:10	U	SMPL
PB102891BL - 1	PBW	-168	-0.0719	-	7470A	PPB	04 Oct 2017 10:50:19	U	SMPL
PB102891BS - 1	LCSW	12965	3.7780	-	7470A	PPB	04 Oct 2017 10:52:28	U	SMPL
PB102891TB - 1	PB102851TB	-82	-0.0467	-	7470A	PPB	04 Oct 2017 10:54:36	U	SMPL
I5584-04 - 1	TP-1-COMP	-19	-0.0282	-	7470A	PPB	04 Oct 2017 10:56:44	U	SMPL
I5587-09 - 1	G28A-0-2	-2	-0.0233	-	7470A	PPB	04 Oct 2017 10:58:52	U	SMPL
I5587-10 - 1	G28A-2-4	22	-0.0162	-	7470A	PPB	04 Oct 2017 11:01:01	U	SMPL
I5587-11 - 1	G10-0-1.5	16	-0.0180	-	7470A	PPB	04 Oct 2017 11:03:08	U	SMPL
CCV09 - 1	CCV09	16850	4.9169	-	7470A	PPB	04 Oct 2017 11:05:14	U	SMPL
CCB09 - 1	CCB09	47	-0.0089	-	7470A	PPB	04 Oct 2017 11:07:20	U	SMPL
I5587-12 - 1	G10-1.5-3	-4	-0.0238	-	7470A	PPB	04 Oct 2017 11:09:27	U	SMPL
I5604-03 - 1	SP-A-100217	-6	-0.0244	-	7470A	PPB	04 Oct 2017 11:11:34	U	SMPL
I5604-06 - 1	SP-B-100217	3	-0.0218	-	7470A	PPB	04 Oct 2017 11:13:41	U	SMPL
I5604-06DUP - 1	SP-B-100217DUP	17	-0.0177	-	7470A	PPB	04 Oct 2017 11:15:48	U	SMPL
I5604-06MS - 1	SP-B-100217MS	12303	3.5839	-	7470A	PPB	04 Oct 2017 11:17:55	U	SMPL
I5604-06MSD - 1	SP-B-100217MSD	12527	3.6496	-	7470A	PPB	04 Oct 2017 11:20:03	U	SMPL
I5604-06LX5 - 1		42	-0.0104	-	7470A	PPB	04 Oct 2017 11:22:11	U	SMPL
I5604-06A - 1		14021	4.0876	-	7470A	PPB	04 Oct 2017 11:24:20	U	SMPL
CCV10 - 1	CCV10	16975	4.9535	-	7470A	PPB	04 Oct 2017 11:26:28	U	SMPL
CCB10 - 1	CCB10	22	-0.0162	-	7470A	PPB	04 Oct 2017 11:28:34	U	SMPL