



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90555

Temp.: 19.2 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/04/2017	13:35
2	CAL2	1	Water			7.00	10/04/2017	13:36
3	CAL3	1	Water			10.04	10/04/2017	13:37
4	ICV	1	Water			7.00	10/04/2017	13:39
5	CCV1	1	Water			2.01	10/04/2017	13:40
6	I5634-01	1	Water			7.21	10/04/2017	13:41
7	I5634-01DUP	1	Water			7.22	10/04/2017	13:42
8	CCV2	1	Water			12.01	10/04/2017	13:43

LB90555

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5638

WorkList ID : 103693

Date : 10/4/2017 1:25:38 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5634-01-1	pH	Cool 4 deg C	MINO01	J22	OUTFALL	10/02/2017	9040C

Date/Time 10/04/17 1:25 PM
Received by: JH
Relinquished by:

Date/Time 10/04/17 2:00 PM
Received by: CP
Relinquished by: JH