


ISM02.4

LB 90565

 μ Abs.:

43823



Linear

A= 0.0000e+000

B= 2.2947e-004 *slope*C= -2.0180e-002 *x-intercept*

Rho= 0.9999335

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.023	-0.023	-12	0.000	-12				
0.2	0.200	0.194	-0.006	932	0.0 %	932				
2.5	2.500	2.565	0.065	11264	0.0 %	11264				
5.0	5.000	4.998	-0.002	21870	0.0 %	21870				
7.5	7.500	7.430	-0.070	32469	0.0 %	32469				
10.0	10.000	10.036	0.036	43823	0.0 %	43823				

*0.101**-3**3**0**-1**0**10/4/12*

LB90565
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-12	-		0 ISM02.4	PPB	10/4/2017 15:53	Std
	0.2 S0.2	932	-		0.2 ISM02.4	PPB	10/4/2017 15:56	Std
	2.5 S2.5	11264	-		2.5 ISM02.4	PPB	10/4/2017 15:58	Std
	5 S5	21870	-		5 ISM02.4	PPB	10/4/2017 16:00	Std
	7.5 S7.5	32469	-		7.5 ISM02.4	PPB	10/4/2017 16:02	Std
	10 S10	43823	-		10 ISM02.4	PPB	10/4/2017 16:05	Std
ICV	ICV	17177	3.9214	-	ISM02.4	PPB	10/4/2017 16:07	SMPL
ICB	ICB	-228	-0.0725	-	ISM02.4	PPB	10/4/2017 16:10	SMPL
CCV014	CCV014	22095	5.05	-	ISM02.4	PPB	10/4/2017 16:12	SMPL
CCB014	CCB014	-220	-0.0707	-	ISM02.4	PPB	10/4/2017 16:14	SMPL
PB102893BL	PBS004	50	-0.0087	-	ISM02.4	PPB	10/4/2017 16:16	SMPL
I5524-01	MC0AD2	2173	0.4785	-	ISM02.4	PPB	10/4/2017 16:19	SMPL
I5524-02	MC0AD3	2979	0.6634	-	ISM02.4	PPB	10/4/2017 16:21	SMPL
I5524-03	MC0AD4	2609	0.5785	-	ISM02.4	PPB	10/4/2017 16:23	SMPL
I5524-04	MC0AD5	6911	1.5657	-	ISM02.4	PPB	10/4/2017 16:26	SMPL
I5524-05	MC0AD5D	8003	1.8163	-	ISM02.4	PPB	10/4/2017 16:28	SMPL
I5524-06	MC0AD5S	19079	4.3579	-	ISM02.4	PPB	10/4/2017 16:30	SMPL
I5524-07	MC0AD6	7072	1.6026	-	ISM02.4	PPB	10/4/2017 16:32	SMPL
I5524-08	MC0AD7	8549	1.9416	-	ISM02.4	PPB	10/4/2017 16:35	SMPL
I5524-09	MC0AD8	6552	1.4833	-	ISM02.4	PPB	10/4/2017 16:37	SMPL
I5524-10	MC0AD9	7599	1.7236	-	ISM02.4	PPB	10/4/2017 16:39	SMPL
I5524-11	MC0AE0	2493	0.5519	-	ISM02.4	PPB	10/4/2017 16:41	SMPL
I5524-12	MC0AE1	2073	0.4555	-	ISM02.4	PPB	10/4/2017 16:44	SMPL
I5524-13	MC0AE2	2214	0.4879	-	ISM02.4	PPB	10/4/2017 16:46	SMPL
I5524-14	MC0AE3	7876	1.7871	-	ISM02.4	PPB	10/4/2017 16:48	SMPL
I5524-15	MC0AB2	1386	0.2979	-	ISM02.4	PPB	10/4/2017 16:51	SMPL
I5524-16	MC0AB3	1840	0.402	-	ISM02.4	PPB	10/4/2017 16:53	SMPL
I5524-17	MC0AB4	2071	0.4551	-	ISM02.4	PPB	10/4/2017 16:55	SMPL
I5524-18	MC0AB5	1480	0.3194	-	ISM02.4	PPB	10/4/2017 16:57	SMPL
I5524-19	MC0AB6	12767	2.9095	-	ISM02.4	PPB	10/4/2017 17:00	SMPL
CCV015	CCV015	22140	5.0603	-	ISM02.4	PPB	10/4/2017 17:02	SMPL
CCB015	CCB015	-233	-0.0736	-	ISM02.4	PPB	10/4/2017 17:04	SMPL
I5524-20	MC0AB7	2060	0.4525	-	ISM02.4	PPB	10/4/2017 17:07	SMPL
I5524-21	MC0AB8	1517	0.3279	-	ISM02.4	PPB	10/4/2017 17:09	SMPL
I5524-22	MC0AB9	1033	0.2169	-	ISM02.4	PPB	10/4/2017 17:11	SMPL
PB102894BL	PBS004	-10	-0.0225	-	ISM02.4	PPB	10/4/2017 17:13	SMPL
I5568-01	B-53(1-3)-092817	14355	3.2739	-	ISM02.4	PPB	10/4/2017 17:16	SMPL
I5568-02	B-51(1-2)-092817	11957	2.7236	-	ISM02.4	PPB	10/4/2017 17:18	SMPL
I5568-03	B-51(5-7)-092817	5389	1.2164	-	ISM02.4	PPB	10/4/2017 17:20	SMPL
I5568-04	B-75(0-1)-092917	2392	0.5287	-	ISM02.4	PPB	10/4/2017 17:22	SMPL
I5568-04D	B-75(0-1)-092917D	2309	0.5097	-	ISM02.4	PPB	10/4/2017 17:25	SMPL
I5568-04S	B-75(0-1)-092917S	13660	3.1144	-	ISM02.4	PPB	10/4/2017 17:27	SMPL
I5568-05	B-75(1-3)-092917	12057	2.7465	-	ISM02.4	PPB	10/4/2017 17:29	SMPL
I5568-06	B-29(0-1)-092917	52825	12.1016	-	ISM02.4	PPB	10/4/2017 17:32	SMPL
I5568-07	B-29(1-3)-092917	66758	15.2988	-	ISM02.4	PPB	10/4/2017 17:34	SMPL
PB102895BL	PBS004	-11	-0.0227	-	ISM02.4	PPB	10/4/2017 17:44	SMPL
I5570-01	MC0ACO	8617	1.9572	-	ISM02.4	PPB	10/4/2017 17:46	SMPL

LB90565
INSTRUMENT ID : CV1

I5570-02	MC0AC1	7318	1.6591 -	ISM02.4	PPB	10/4/2017 17:49	SMPL
I5570-03	MC0AC2	2395	0.5294 -	ISM02.4	PPB	10/4/2017 17:51	SMPL
I5570-04	MC0AC3	2299	0.5074 -	ISM02.4	PPB	10/4/2017 17:53	SMPL
I5570-05	MC0AC4	4289	0.964 -	ISM02.4	PPB	10/4/2017 17:56	SMPL
I5570-06	MC0AC5	8604	1.9542 -	ISM02.4	PPB	10/4/2017 17:58	SMPL
CCV016	CCV016	22226	5.08 -	ISM02.4	PPB	10/4/2017 18:00	SMPL
CCB016	CCB016	-219	-0.0704 -	ISM02.4	PPB	10/4/2017 18:02	SMPL
I5570-07	MC0AC6	3428	0.7664 -	ISM02.4	PPB	10/4/2017 18:05	SMPL
I5570-08	MC0AC7	6517	1.4753 -	ISM02.4	PPB	10/4/2017 18:07	SMPL
I5570-09	MC0AC8	3872	0.8683 -	ISM02.4	PPB	10/4/2017 18:09	SMPL
I5570-10	MC0AC8D	4393	0.9879 -	ISM02.4	PPB	10/4/2017 18:12	SMPL
I5570-11	MC0AC8S	15527	3.5428 -	ISM02.4	PPB	10/4/2017 18:14	SMPL
I5570-12	MC0AC9	7594	1.7224 -	ISM02.4	PPB	10/4/2017 18:16	SMPL
I5570-13	MC0AD0	3056	0.6811 -	ISM02.4	PPB	10/4/2017 18:18	SMPL
I5570-14	MC0AD1	8676	1.9707 -	ISM02.4	PPB	10/4/2017 18:21	SMPL
I5570-15	MC0AE7	1312	0.2809 -	ISM02.4	PPB	10/4/2017 18:23	SMPL
I5570-16	MC0AE8	4715	1.0618 -	ISM02.4	PPB	10/4/2017 18:25	SMPL
I5570-17	MC0AF0	17232	3.934 -	ISM02.4	PPB	10/4/2017 18:27	SMPL
I5570-18	MC0AF1	7495	1.6997 -	ISM02.4	PPB	10/4/2017 18:30	SMPL
I5570-19	MC0AF4	552	0.1065 -	ISM02.4	PPB	10/4/2017 18:32	SMPL
I5570-20	MC0AF5	1045	0.2196 -	ISM02.4	PPB	10/4/2017 18:34	SMPL
I5570-21	MC0AF6	1883	0.4119 -	ISM02.4	PPB	10/4/2017 18:37	SMPL
I5570-22	MC0AF7	5791	1.3087 -	ISM02.4	PPB	10/4/2017 18:39	SMPL
PB102896BL	PBS004	4	-0.0193 -	ISM02.4	PPB	10/4/2017 18:41	SMPL
I5590-01	MC0AF8	10494	2.3879 -	ISM02.4	PPB	10/4/2017 18:43	SMPL
I5590-02	MC0AF8D	10433	2.3739 -	ISM02.4	PPB	10/4/2017 18:49	SMPL
I5590-03	MC0AF8S	21396	4.8896 -	ISM02.4	PPB	10/4/2017 18:51	SMPL
CCV017	CCV017	22134	5.0589 -	ISM02.4	PPB	10/4/2017 18:53	SMPL
CCB017	CCB017	-229	-0.0727 -	ISM02.4	PPB	10/4/2017 18:56	SMPL
PB102897BL	PBS004	45	-0.0099 -	ISM02.4	PPB	10/4/2017 18:58	SMPL
I5611-01	B-44(0-1)_100217	13962	3.1837 -	ISM02.4	PPB	10/4/2017 19:00	SMPL
I5611-02	B-44(5-5.5)_100217	14508	3.309 -	ISM02.4	PPB	10/4/2017 19:03	SMPL
I5611-03	B-42(0.3-1.3)_100217	39286	8.9948 -	ISM02.4	PPB	10/4/2017 19:05	SMPL
I5611-03D	B-42(0.3-1.3)_100217D	41840	9.5808 -	ISM02.4	PPB	10/4/2017 19:07	SMPL
I5611-03S	B-42(0.3-1.3)_100217S	50286	11.5189 -	ISM02.4	PPB	10/4/2017 19:14	SMPL
I5568-06DLX2	B-2990(0-1)-092917	26228	5.9984 -	ISM02.4	PPB	10/4/2017 19:16	SMPL
I5568-07DLX2	B-2990(1-3)-092917	33233	7.6058 -	ISM02.4	PPB	10/4/2017 19:19	SMPL
I5611-03DLX2	B-44(0-1)_100217	19360	4.4224 -	ISM02.4	PPB	10/4/2017 19:21	SMPL
I5611-03DDLX2	B-44(0-1)_100217D	20857	4.7659 -	ISM02.4	PPB	10/4/2017 19:23	SMPL
I5611-03SDLX2	B-44(0-1)_100217S	25705	5.8783 -	ISM02.4	PPB	10/4/2017 19:26	SMPL
CCV018	CCV018	22175	5.0683 -	ISM02.4	PPB	10/4/2017 19:28	SMPL
CCB018	CCB018	-185	-0.0626 -	ISM02.4	PPB	10/4/2017 19:30	SMPL