



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90571

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/04/2017	15:30
2	CAL2	1	Water			7.01	10/04/2017	15:31
3	CAL3	1	Water			10.04	10/04/2017	15:33
4	ICV	1	Water			7.00	10/04/2017	15:35
5	CCV1	1	Water			2.01	10/04/2017	15:36
6	I5658-03	1	Solid	20.02	20	8.70	10/04/2017	15:37
7	I5678-03	1	Solid	20.04	20	8.54	10/04/2017	15:38
8	I5678-03DUP	1	Solid	20.02	20	8.55	10/04/2017	15:40
9	CCV2	1	Water			12.01	10/04/2017	15:41

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5658

WorkList ID : 103692

Date : 10/4/2017 1:23:43 PM

2B90571

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5658-03	Corrosivity	Cool 4 deg C	PSEG05	H52	SU-02-100417	10/04/2017	9045C
	Solid	I5678-03	Corrosivity	Cool 4 deg C	PSEG05	H52	OR-03-100417	10/04/2017	9045C

Date/Time 10/04/17 1:20 PM
 Received by: JP
 Relinquished by: CS

Date/Time 10/04/17 5:00 PM
 Received by: CS
 Relinquished by: JP