

245.1


LB 90586

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32406



A= 0.0000e+000

B= 3.1151e-004

C= -4.3657e-002

Rho= 0.9999198

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	186	0.000	186				
0.2	0.200	0.196	-0.004	769	0.0 %	769				
2.5	2.500	2.494	-0.006	8145	0.0 %	8145				
5.0	5.000	5.038	0.038	16314	0.0 %	16314				
7.5	7.500	7.407	-0.093	23916	0.0 %	23916				
10.0	10.000	10.051	0.051	32406	0.0 %	32406				

LB90586 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Oct 2017 11:02:09

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	186	-	0.0000	245.1	PPB	05 Oct 2017 11:13:22	S	Std
0.2 - 1	50.2	769	-	0.2000	245.1	PPB	05 Oct 2017 11:15:30	S	Std
2.5 - 1	52.5	8145	-	2.5000	245.1	PPB	05 Oct 2017 11:17:39	S	Std
5.0 - 1	55	16314	-	5.0000	245.1	PPB	05 Oct 2017 11:19:47	S	Std
7.5 - 1	57.5	23916	-	7.5000	245.1	PPB	05 Oct 2017 11:21:55	S	Std
10.0 - 1	510	32406	-	10.0000	245.1	PPB	05 Oct 2017 11:27:20	S	Std
ICV90 - 1	ICV90	12639	3.8936	-245.1		PPB	05 Oct 2017 11:29:48	U	SMPL
ICB90 - 1	ICB90	35	-0.0328	-245.1		PPB	05 Oct 2017 11:31:54	U	SMPL
CCV19 - 1	CCV19	16132	4.9817	-245.1		PPB	05 Oct 2017 11:34:01	U	SMPL
CCB19 - 1	CCB19	38	-0.0318	-245.1		PPB	05 Oct 2017 11:36:08	U	SMPL
CRA - 1	CRA	763	0.1940	-245.1		PPB	05 Oct 2017 11:38:16	U	SMPL
HighStd - 1	HighStd	31714	9.8357	-245.1		PPB	05 Oct 2017 11:40:22	U	SMPL
ChkStd - 1	ChkStd	21925	6.7863	-245.1		PPB	05 Oct 2017 11:42:29	U	SMPL
PB102919BL - 1	PBW	-147	-0.0894	-245.1		PPB	05 Oct 2017 11:44:38	U	SMPL
PB102919BS - 1	LCSW	12306	3.7898	-245.1		PPB	05 Oct 2017 11:46:46	U	SMPL
I5645-01 - 1	100317-TIDO-F-D-B	-147	-0.0894	-245.1		PPB	05 Oct 2017 11:48:54	U	SMPL
I5645-02 - 1	100317-TIDO-F-D-S	-103	-0.0757	-245.1		PPB	05 Oct 2017 11:51:02	U	SMPL
I5645-03 - 1	100317-TIDO-F-U-B	-64	-0.0636	-245.1		PPB	05 Oct 2017 11:53:10	U	SMPL
I5645-04 - 1	100317-TIDO-F-U-S	-74	-0.0667	-245.1		PPB	05 Oct 2017 11:55:19	U	SMPL
I5646-01 - 1	100217-TIDO-E-D-M	-51	-0.0595	-245.1		PPB	05 Oct 2017 11:57:27	U	SMPL
CCV20 - 1	CCV20	16414	5.0695	-245.1		PPB	05 Oct 2017 12:05:27	U	SMPL
CCB20 - 1	CCB20	53	-0.0271	-245.1		PPB	05 Oct 2017 12:07:33	U	SMPL
I5646-02 - 1	100217-TIDO-E-U-M	-90	-0.0717	-245.1		PPB	05 Oct 2017 12:09:40	U	SMPL
I5646-02DUP - 1	100217-TIDO-E-U-MDUP	-88	-0.0711	-245.1		PPB	05 Oct 2017 12:11:47	U	SMPL
I5646-02MS - 1	100217-TIDO-E-U-MMS	13660	4.2116	-245.1		PPB	05 Oct 2017 12:13:53	U	SMPL
I5646-02MSD - 1	100217-TIDO-E-U-MMSD	14203	4.3808	-245.1		PPB	05 Oct 2017 12:16:00	U	SMPL
I5646-02LX5 - 1		-2	-0.0443	-245.1		PPB	05 Oct 2017 12:18:08	U	SMPL
I5646-02A - 1		15716	4.8521	-245.1		PPB	05 Oct 2017 12:20:16	U	SMPL
CCV21 - 1	CCV21	16245	5.0169	-245.1		PPB	05 Oct 2017 12:22:24	U	SMPL
CCB21 - 1	CCB21	69	-0.0222	-245.1		PPB	05 Oct 2017 12:24:33	U	SMPL