


7470A

LB 90595
INSTRUMENT ID: CV4

Linear

 μ Abs.:

35915



A= 0.0000e+000

B= 2.8128e-004

C= 2.4576e-002

Rho= 0.9996645

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.038	0.038	46	0.000	46				
0.2	0.200	0.221	0.021	698	0.0 %	698				
2.5	2.500	2.473	-0.027	8703	0.0 %	8703				
5.0	5.000	5.030	0.030	17795	0.0 %	17795				
7.5	7.500	7.312	-0.188	25909	0.0 %	25909				
10.0	10.000	10.127	0.127	35915	0.0 %	35915				

LB90595 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Oct 2017 12:36:48

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7470A	PPB	05 Oct 2017 12:38:32	S	Std
0.2 - 1	50.2	698	-	0.2000	7470A	PPB	05 Oct 2017 12:40:40	S	Std
2.5 - 1	52.5	8703	-	2.5000	7470A	PPB	05 Oct 2017 12:42:49	S	Std
5.0 - 1	55	17795	-	5.0000	7470A	PPB	05 Oct 2017 12:44:57	S	Std
7.5 - 1	57.5	25909	-	7.5000	7470A	PPB	05 Oct 2017 12:47:05	S	Std
10.0 - 1	510	35915	-	10.0000	7470A	PPB	05 Oct 2017 12:49:13	S	Std
ICV91 - 1	ICV91	13389	3.7906	-	7470A	PPB	05 Oct 2017 12:51:50	U	SMPL
ICB91 - 1	ICB91	-100	-0.0036	-	7470A	PPB	05 Oct 2017 12:53:56	U	SMPL
CCV22 - 1	CCV22	18061	5.1048	-	7470A	PPB	05 Oct 2017 12:56:03	U	SMPL
CCB22 - 1	CCB22	-79	0.0024	-	7470A	PPB	05 Oct 2017 12:58:10	U	SMPL
CRA - 1	CRA	697	0.2206	-	7470A	PPB	05 Oct 2017 13:00:16	U	SMPL
HighStd - 1	HighStd	35674	10.0590	-	7470A	PPB	05 Oct 2017 13:02:23	U	SMPL
ChkStd - 1	ChkStd	24387	6.8842	-	7470A	PPB	05 Oct 2017 13:04:30	U	SMPL
PB102912BL - 1	PBW	-204	-0.0328	-	7470A	PPB	05 Oct 2017 13:06:39	U	SMPL
PB102912BS - 1	LCSW	12617	3.5735	-	7470A	PPB	05 Oct 2017 13:08:47	U	SMPL
PB102912TB - 1	PB102907TB	-71	0.0046	-	7470A	PPB	05 Oct 2017 13:10:55	U	SMPL
I5585-21 - 1	G67-0-1	2	0.0251	-	7470A	PPB	05 Oct 2017 13:13:03	U	SMPL
I5585-22 - 1	G67-1-2	7	0.0265	-	7470A	PPB	05 Oct 2017 13:15:11	U	SMPL
I5585-23 - 1	G68-0-1.5	30	0.0330	-	7470A	PPB	05 Oct 2017 13:17:19	U	SMPL
I5585-24 - 1	G68-1.5-3	52	0.0392	-	7470A	PPB	05 Oct 2017 13:19:26	U	SMPL
CCV23 - 1	CCV23	17354	4.9059	-	7470A	PPB	05 Oct 2017 13:21:33	U	SMPL
CCB23 - 1	CCB23	-101	-0.0038	-	7470A	PPB	05 Oct 2017 13:23:40	U	SMPL
I5585-25 - 1	G69-0-1.5	-1	0.0243	-	7470A	PPB	05 Oct 2017 13:25:48	U	SMPL
I5585-26 - 1	G69-1.5-3	18	0.0296	-	7470A	PPB	05 Oct 2017 13:27:54	U	SMPL
I5585-27 - 1	G70-0-2.5	44	0.0370	-	7470A	PPB	05 Oct 2017 13:30:01	U	SMPL
I5585-28 - 1	G70-2.5-5	17	0.0294	-	7470A	PPB	05 Oct 2017 13:32:08	U	SMPL
I5585-29 - 1	G71-0-1.5	39	0.0355	-	7470A	PPB	05 Oct 2017 13:34:16	U	SMPL
I5585-30 - 1	G71-1.5-3	51	0.0389	-	7470A	PPB	05 Oct 2017 13:36:24	U	SMPL
I5586-21 - 1	G72-0-1	62	0.0420	-	7470A	PPB	05 Oct 2017 13:38:32	U	SMPL
I5586-22 - 1	G72-1-2	45	0.0372	-	7470A	PPB	05 Oct 2017 13:40:40	U	SMPL
I5586-23 - 1	G47A-0-2.5	25	0.0316	-	7470A	PPB	05 Oct 2017 13:42:49	U	SMPL
I5586-24 - 1	G47A-2.5-5	25	0.0316	-	7470A	PPB	05 Oct 2017 13:44:56	U	SMPL
CCV24 - 1	CCV24	17472	4.9391	-	7470A	PPB	05 Oct 2017 13:47:03	U	SMPL
CCB24 - 1	CCB24	-111	-0.0066	-	7470A	PPB	05 Oct 2017 13:49:10	U	SMPL
I5586-25 - 1	G54A-0-3.5	8	0.0268	-	7470A	PPB	05 Oct 2017 13:51:17	U	SMPL
I5586-26 - 1	G54A-3.5-7	39	0.0355	-	7470A	PPB	05 Oct 2017 13:53:24	U	SMPL
I5586-27 - 1	G54B-0-3	38	0.0353	-	7470A	PPB	05 Oct 2017 13:55:30	U	SMPL
I5586-28 - 1	G54B-3-6	35	0.0344	-	7470A	PPB	05 Oct 2017 13:57:37	U	SMPL
I5586-29 - 1	G56A-0-3	23	0.0310	-	7470A	PPB	05 Oct 2017 13:59:45	U	SMPL
I5586-30 - 1	G56A-3-6	17	0.0294	-	7470A	PPB	05 Oct 2017 14:01:53	U	SMPL
I5586-30DUP - 1	G56A-3-6DUP	30	0.0330	-	7470A	PPB	05 Oct 2017 14:04:01	U	SMPL
I5586-30MS - 1	G56A-3-6MS	12838	3.6357	-	7470A	PPB	05 Oct 2017 14:06:09	U	SMPL
I5586-30MSD - 1	G56A-3-6MSD	12867	3.6438	-	7470A	PPB	05 Oct 2017 14:08:18	U	SMPL
PB102913BL - 1	PBW	-74	0.0038	-	7470A	PPB	05 Oct 2017 14:10:25	U	SMPL
CCV25 - 1	CCV25	17674	4.9959	-	7470A	PPB	05 Oct 2017 14:12:32	U	SMPL
CCB25 - 1	CCB25	-98	-0.0030	-	7470A	PPB	05 Oct 2017 14:14:39	U	SMPL
PB102913BS - 1	LCSW	13283	3.7608	-	7470A	PPB	05 Oct 2017 14:16:47	U	SMPL
PB102913TB - 1	PB102874TB	-97	-0.0027	-	7470A	PPB	05 Oct 2017 14:18:54	U	SMPL
I5610-25 - 1	1A-(1-2)-COMP-(0-5)	37	0.0350	-	7470A	PPB	05 Oct 2017 14:21:02	U	SMPL
I5610-26 - 1	1B-(1-2)-COMP-(0-5)	35	0.0344	-	7470A	PPB	05 Oct 2017 14:23:09	U	SMPL
I5610-27 - 1	2A-(1-2)-COMP-(0-5)	52	0.0392	-	7470A	PPB	05 Oct 2017 14:25:16	U	SMPL
I5610-28 - 1	2B-(1-2)-COMP-(0-5)	77	0.0462	-	7470A	PPB	05 Oct 2017 14:27:24	U	SMPL
I5616-04 - 1	STONE-SP-A-COMP	22	0.0308	-	7470A	PPB	05 Oct 2017 14:29:32	U	SMPL
I5616-08 - 1	STONE-SP-B-COMP	32	0.0336	-	7470A	PPB	05 Oct 2017 14:31:40	U	SMPL
I5644-04 - 1	TP-9-COMP	27	0.0322	-	7470A	PPB	05 Oct 2017 14:33:49	U	SMPL
I5644-08 - 1	TP-10-11-COMP	25	0.0316	-	7470A	PPB	05 Oct 2017 14:35:56	U	SMPL
CCV26 - 1	CCV26	17690	5.0004	-	7470A	PPB	05 Oct 2017 14:38:03	U	SMPL
CCB26 - 1	CCB26	-97	-0.0027	-	7470A	PPB	05 Oct 2017 14:40:11	U	SMPL
I5644-12 - 1	TP-12-13-COMP	-2	0.0240	-	7470A	PPB	05 Oct 2017 14:42:18	U	SMPL
I5644-16 - 1	TP-14-COMP	6	0.0263	-	7470A	PPB	05 Oct 2017 14:44:25	U	SMPL
I5647-03 - 1	OK-02-100317	26	0.0319	-	7470A	PPB	05 Oct 2017 14:46:32	U	SMPL
I5647-03DUP - 1	OK-02-100317DUP	5	0.0260	-	7470A	PPB	05 Oct 2017 14:48:39	U	SMPL
I5647-03MS - 1	OK-02-100317MS	13095	3.7079	-	7470A	PPB	05 Oct 2017 14:50:47	U	SMPL
I5647-03MSD - 1	OK-02-100317MSD	12573	3.5611	-	7470A	PPB	05 Oct 2017 14:52:55	U	SMPL
PB102924BL - 1	PBW	-83	0.0012	-	7470A	PPB	05 Oct 2017 14:55:02	U	SMPL
PB102924BS - 1	LCSW	13016	3.6857	-	7470A	PPB	05 Oct 2017 14:57:11	U	SMPL
PB102924TB - 1	PB102889TB	-53	0.0097	-	7470A	PPB	05 Oct 2017 14:59:20	U	SMPL
I5637-03 - 1	B-18(0-10)COMP	36	0.0347	-	7470A	PPB	05 Oct 2017 15:01:27	U	SMPL
CCV27 - 1	CCV27	17708	5.0055	-	7470A	PPB	05 Oct 2017 15:03:34	U	SMPL
CCB27 - 1	CCB27	-100	-0.0036	-	7470A	PPB	05 Oct 2017 15:05:42	U	SMPL
I5637-03DUP - 1	B-18(0-10)COMP DUP	40	0.0358	-	7470A	PPB	05 Oct 2017 15:07:49	U	SMPL
I5637-03MS - 1	B-18(0-10)COMPMS	13113	3.7130	-	7470A	PPB	05 Oct 2017 15:09:56	U	SMPL
I5637-03MSD - 1	B-18(0-10)COMPMSD	13204	3.7386	-	7470A	PPB	05 Oct 2017 15:12:04	U	SMPL
I5586-30LX5 - 1		-71	0.0046	-	7470A	PPB	05 Oct 2017 15:14:11	U	SMPL
I5586-30A - 1		14320	4.0525	-	7470A	PPB	05 Oct 2017 15:16:18	U	SMPL
I5647-03LX5 - 1		-68	0.0054	-	7470A	PPB	05 Oct 2017 15:18:26	U	SMPL
I5647-03A - 1		14051	3.9769	-	7470A	PPB	05 Oct 2017 15:20:34	U	SMPL
I5637-03LX5 - 1		-72	0.0043	-	7470A	PPB	05 Oct 2017 15:22:42	U	SMPL
I5637-03A - 1		14252	4.0334	-	7470A	PPB	05 Oct 2017 15:24:51	U	SMPL
CCV28 - 1	CCV28	17718	5.0083	-	7470A	PPB	05 Oct 2017 15:26:58	U	SMPL
CCB28 - 1	CCB28	-115	-0.0078	-	7470A	PPB	05 Oct 2017 15:29:06	U	SMPL