


7471A

LB 90597
INSTRUMENT ID: CV2

Linear

 μ Abs.:

41224



A= 0.0000e+000

B= 2.4363e-004

C= -4.3350e-002

Rho= 0.9997997

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	312	0.000	312				
0.2	0.200	0.231	0.031	1125	0.0 %	1125				
2.5	2.500	2.486	-0.014	10380	0.0 %	10380				
5.0	5.000	4.854	-0.146	20103	0.0 %	20103				
7.5	7.500	7.596	0.096	31358	0.0 %	31358	0			
10.0	10.000	10.000	0.000	41224	0.0 %	41224				

LB90597 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 05 Oct 2017 15:46:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	312	-	0.00007471A	PPB		05 Oct 2017 16:04:28	S	Std
0.2 - 1	50.2	1125	-	0.20007471A	PPB		05 Oct 2017 16:06:46	S	Std
2.5 - 1	52.5	10380	-	2.50007471A	PPB		05 Oct 2017 16:09:04	S	Std
5.0 - 1	55	20103	-	5.00007471A	PPB		05 Oct 2017 16:11:21	S	Std
7.5 - 1	57.5	31358	-	7.50007471A	PPB		05 Oct 2017 16:13:38	S	Std
10.0 - 1	510	41224	-	10.00007471A	PPB		05 Oct 2017 16:18:30	S	Std
ICV92 - 1	ICV92	16571	3.9939	-7471A	PPB		05 Oct 2017 16:21:23	U	SMPL
ICB92 - 1	ICB92	126	-0.0127	-7471A	PPB		05 Oct 2017 16:23:39	U	SMPL
CCV29 - 1	CCV29	20805	5.0254	-7471A	PPB		05 Oct 2017 16:25:55	U	SMPL
CCB29 - 1	CCB29	106	-0.0175	-7471A	PPB		05 Oct 2017 16:28:11	U	SMPL
CRA - 1	CRA	1084	0.2207	-7471A	PPB		05 Oct 2017 16:30:28	U	SMPL
HighStd - 1	HighStd	41632	10.0996	-7471A	PPB		05 Oct 2017 16:32:44	U	SMPL
ChkStd - 1	ChkStd	28520	6.9051	-7471A	PPB		05 Oct 2017 16:34:59	U	SMPL
PB102925BL - 1	PBS	155	-0.0056	-7471A	PPB		05 Oct 2017 16:39:51	U	SMPL
PB102925BS - 1	LCSS	15054	3.6243	-7471A	PPB		05 Oct 2017 16:42:08	U	SMPL
I5616-01 - 1	STONE-SP-A-COMP	453	0.0670	-7471A	PPB		05 Oct 2017 16:44:25	U	SMPL
I5616-05 - 1	STONE-SP-B-COMP	462	0.0692	-7471A	PPB		05 Oct 2017 16:46:42	U	SMPL
I5637-01 - 1	B-18(0-10)COMP	55457	13.4678	-7471A	PPB		05 Oct 2017 16:49:00	U	SMPL
I5640-01 - 1	DRUM-1	-299	-0.1162	-7471A	PPB		05 Oct 2017 16:51:18	U	SMPL
I5640-02 - 1	DRUM-2	7	-0.0416	-7471A	PPB		05 Oct 2017 16:53:36	U	SMPL
CCV30 - 1	CCV30	20741	5.0098	-7471A	PPB		05 Oct 2017 16:55:52	U	SMPL
CCB30 - 1	CCB30	124	-0.0131	-7471A	PPB		05 Oct 2017 16:58:08	U	SMPL
I5640-03 - 1	DRUM-3	115	-0.0153	-7471A	PPB		05 Oct 2017 17:00:24	U	SMPL
I5640-04 - 1	DRUM-4	66	-0.0273	-7471A	PPB		05 Oct 2017 17:02:39	U	SMPL
I5640-05 - 1	DRUM-5	79	-0.0241	-7471A	PPB		05 Oct 2017 17:04:56	U	SMPL
I5640-06 - 1	DRUM-6	98	-0.0195	-7471A	PPB		05 Oct 2017 17:07:12	U	SMPL
I5640-07 - 1	DRUM-7	70	-0.0263	-7471A	PPB		05 Oct 2017 17:09:29	U	SMPL
I5640-08 - 1	DRUM-8	168	-0.0024	-7471A	PPB		05 Oct 2017 17:11:45	U	SMPL
I5644-01 - 1	TP-9-COMP	490	0.0760	-7471A	PPB		05 Oct 2017 17:14:03	U	SMPL
I5644-05 - 1	TP-10-11-COMP	1395	0.2965	-7471A	PPB		05 Oct 2017 17:16:20	U	SMPL
I5644-09 - 1	TP-12-13-COMP	529	0.0855	-7471A	PPB		05 Oct 2017 17:18:38	U	SMPL
I5644-13 - 1	TP-14-COMP	809	0.1537	-7471A	PPB		05 Oct 2017 17:20:54	U	SMPL
CCV31 - 1	CCV31	20892	5.0466	-7471A	PPB		05 Oct 2017 17:23:10	U	SMPL
CCB31 - 1	CCB31	99	-0.0192	-7471A	PPB		05 Oct 2017 17:25:27	U	SMPL
I5644-13DUP - 1	TP-14-COMPDUP	834	0.1598	-7471A	PPB		05 Oct 2017 17:27:44	U	SMPL
I5644-13MS - 1	TP-14-COMPMS	16145	3.8901	-7471A	PPB		05 Oct 2017 17:30:01	U	SMPL
I5644-13MSD - 1	TP-14-COMPMSD	16039	3.8643	-7471A	PPB		05 Oct 2017 17:32:17	U	SMPL
I5647-01 - 1	OK-02-100317	1124	0.2305	-7471A	PPB		05 Oct 2017 17:34:33	U	SMPL
I5658-01 - 1	SU-02-100417	1625	0.3526	-7471A	PPB		05 Oct 2017 17:36:50	U	SMPL
I5676-01 - 1	352	21867	5.2842	-7471A	PPB		05 Oct 2017 17:39:07	U	SMPL
I5678-01 - 1	OR-03-100417	2935	0.6717	-7471A	PPB		05 Oct 2017 17:41:25	U	SMPL
PB102926BL - 1	PBS	1	-0.0431	-7471A	PPB		05 Oct 2017 17:43:43	U	SMPL
I5651-01 - 1	ENV-115-2(0-5)	5290	1.2455	-7471A	PPB		05 Oct 2017 17:46:01	U	SMPL
I5651-02 - 1	ENV-115-2(5-10)	330	0.0370	-7471A	PPB		05 Oct 2017 17:48:17	U	SMPL
CCV32 - 1	CCV32	20783	5.0201	-7471A	PPB		05 Oct 2017 17:50:33	U	SMPL
CCB32 - 1	CCB32	106	-0.0175	-7471A	PPB		05 Oct 2017 17:52:50	U	SMPL
I5651-03 - 1	ENV-115-2(10-15)	158	-0.0049	-7471A	PPB		05 Oct 2017 17:55:06	U	SMPL
I5651-04 - 1	ENV-115-2(15-20)	190	0.0029	-7471A	PPB		05 Oct 2017 17:57:22	U	SMPL
I5651-05 - 1	ENV-115-2(20-30)	205	0.0066	-7471A	PPB		05 Oct 2017 17:59:38	U	SMPL
I5651-06 - 1	ENV-115-2(30-40)	626	0.1092	-7471A	PPB		05 Oct 2017 18:01:55	U	SMPL
I5651-07 - 1	ENV-115-2(40-50)	421	0.0592	-7471A	PPB		05 Oct 2017 18:04:12	U	SMPL
I5651-08 - 1	ENV-116-1(0-5)	50333	12.2194	-7471A	PPB		05 Oct 2017 18:06:29	U	SMPL
I5651-09 - 1	ENV-116-1(5-10)	3055	0.7009	-7471A	PPB		05 Oct 2017 18:08:46	U	SMPL
I5651-09DUP - 1	ENV-116-1(5-10)DUP	3241	0.7463	-7471A	PPB		05 Oct 2017 18:11:06	U	SMPL
I5651-09MS - 1	ENV-116-1(5-10)MS	18854	4.5501	-7471A	PPB		05 Oct 2017 18:13:24	U	SMPL
I5651-09MSD - 1	ENV-116-1(5-10)MSD	19042	4.5959	-7471A	PPB		05 Oct 2017 18:15:40	U	SMPL
CCV33 - 1	CCV33	19886	4.8015	-7471A	PPB		05 Oct 2017 18:17:57	U	SMPL
CCB33 - 1	CCB33	71	-0.0261	-7471A	PPB		05 Oct 2017 18:20:14	U	SMPL
I5651-10 - 1	ENV-116-1(10-15)	336	0.0385	-7471A	PPB		05 Oct 2017 18:22:32	U	SMPL
I5651-11 - 1	ENV-116-7(0-5)	10759	2.5779	-7471A	PPB		05 Oct 2017 18:24:48	U	SMPL
I5651-12 - 1	ENV-116-7(5-10)	5411	1.2749	-7471A	PPB		05 Oct 2017 18:27:04	U	SMPL
I5651-13 - 1	ENV-116-7(10-15)	1044	0.2110	-7471A	PPB		05 Oct 2017 18:29:21	U	SMPL
PB102926BS - 1	LCSS	14865	3.5782	-7471A	PPB		05 Oct 2017 18:31:38	U	SMPL
I5644-13LX5 - 1		258	0.0195	-7471A	PPB		05 Oct 2017 18:33:55	U	SMPL
I5644-13A - 1		16976	4.0926	-7471A	PPB		05 Oct 2017 18:36:12	U	SMPL
I5651-09LX5 - 1		728	0.1340	-7471A	PPB		05 Oct 2017 18:38:30	U	SMPL
I5651-09A - 1		19574	4.7255	-7471A	PPB		05 Oct 2017 18:40:48	U	SMPL
I5637-01DLX2 - 1		28246	6.8383	-7471A	PPB		05 Oct 2017 18:43:04	U	SMPL
CCV34 - 1	CCV34	20742	5.0101	-7471A	PPB		05 Oct 2017 18:45:22	U	SMPL
CCB34 - 1	CCB34	107	-0.0173	-7471A	PPB		05 Oct 2017 18:47:40	U	SMPL
I5651-08DLX2 - 1		25819	6.2470	-7471A	PPB		05 Oct 2017 18:49:57	U	SMPL
CCV35 - 1	CCV35	19883	4.8008	-7471A	PPB		05 Oct 2017 18:52:14	U	SMPL
CCB35 - 1	CCB35	59	-0.0290	-7471A	PPB		05 Oct 2017 18:54:32	U	SMPL