



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: DHVANIDate: 10/06/2017DATE & TIME IN: 10/05/2017 17:30OVEN TEMP: 103 °CRun Number: LB90608DATE & TIME OUT: 10/06/2017 11:00OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB90608BL	LB90608BL	75.3369	75.3369	100	75.3369	75.3369	75.3369	0	0
2	I5706-01	MH-23	93.0551	93.0551	100	95.723	95.723	95.723	2.6679	26679
3	I5706-03	MH-A	79.5205	79.5205	100	79.6584	79.6584	79.6584	0.1379	1379
4	I5706-03DUP	MH-ADUP	81.5218	81.5218	100	81.6618	81.6618	81.6618	0.14	1400
5	I5707-01	2AVE-06	149.252	149.252	100	149.399	149.399	149.399	0.147	1470

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB90608

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS100517

WorkList ID : 103774

Date : 10/5/2017 5:42:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2017	Water	I5706-01 - <u>CT</u>	TS	Cool 4 deg C	DIST01	H11	MH-23	10/05/2017	SM2540 B
10/10/2017	Water	I5706-03 <u>F</u>	TS	Cool 4 deg C	DIST01	H11	MH-A	10/05/2017	SM2540 B
10/12/2017	Water	I5707-01 <u>F</u>	TS	Cool 4 deg C	DIST01	H11	2AVE-06	10/05/2017	SM2540 B

Date/Time 10/05/17 13:15
 Received by: DP
 Relinquished by: CP

Date/Time 10/05/17 17:45
 Received by: CP
 Relinquished by: DP