

7470A


LB90617

INSTRUMENT ID: CX1

Linear

 μ Abs.:

34269



A= 0.0000e+000

B= 2.9213e-004

C= 5.6340e-002

Rho= 0.9997548

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.050	0.050	-21	0.000	-21				
0.2	0.200	0.241	0.041	631	0.0 %	631				
2.5	2.500	2.495	-0.005	8347	0.0 %	8347				
5.0	5.000	4.824	-0.176	16322	0.0 %	16322				
7.5	7.500	7.523	0.023	25559	0.0 %	25559				
10.0	10.000	10.067	0.067	34269	0.0 %	34269				

LB90617 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Oct 2017 10:33:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.0000	7470A	PPB	06 Oct 2017 10:37:31	S	Std
0.2 - 1	50.2	631	-	0.2000	7470A	PPB	06 Oct 2017 10:39:39	S	Std
2.5 - 1	52.5	8347	-	2.5000	7470A	PPB	06 Oct 2017 10:41:47	S	Std
5.0 - 1	55	16322	-	5.0000	7470A	PPB	06 Oct 2017 10:43:56	S	Std
7.5 - 1	57.5	25559	-	7.5000	7470A	PPB	06 Oct 2017 10:46:04	S	Std
10.0 - 1	510	34269	-	10.0000	7470A	PPB	06 Oct 2017 10:48:11	S	Std
ICV93 - 1	ICV93	13291	3.9390	-7470A	PPB	PPB	06 Oct 2017 10:51:23	U	SMPL
ICB93 - 1	ICB93	-164	0.0084	-7470A	PPB	PPB	06 Oct 2017 10:53:29	U	SMPL
CCV36 - 1	CCV36	16775	4.9568	-7470A	PPB	PPB	06 Oct 2017 10:55:35	U	SMPL
CCB36 - 1	CCB36	-158	0.0102	-7470A	PPB	PPB	06 Oct 2017 10:57:42	U	SMPL
CRA - 1	CRA	620	0.2375	-7470A	PPB	PPB	06 Oct 2017 10:59:49	U	SMPL
HighStd - 1	HighStd	33865	9.9492	-7470A	PPB	PPB	06 Oct 2017 11:01:56	U	SMPL
ChkStd - 1	ChkStd	23505	6.9228	-7470A	PPB	PPB	06 Oct 2017 11:04:03	U	SMPL
PB102995BL - 1	PBW	-138	0.0160	-7470A	PPB	PPB	06 Oct 2017 11:06:12	U	SMPL
PB102995BS - 1	LCSW	13740	4.0702	-7470A	PPB	PPB	06 Oct 2017 11:08:20	U	SMPL
PB102995TB - 1	PB102914TB	-46	0.0429	-7470A	PPB	PPB	06 Oct 2017 11:10:28	U	SMPL
I5658-03 - 1	SU-02-100417	97	0.0847	-7470A	PPB	PPB	06 Oct 2017 11:12:36	U	SMPL
I5676-02 - 1	352	121	0.0917	-7470A	PPB	PPB	06 Oct 2017 11:14:44	U	SMPL
I5678-03 - 1	OR-03-100417	96	0.0844	-7470A	PPB	PPB	06 Oct 2017 11:16:53	U	SMPL
I5678-03DUP - 1	OR-03-100417DUP	126	0.0931	-7470A	PPB	PPB	06 Oct 2017 11:19:00	U	SMPL
CCV37 - 1	CCV37	16948	5.0073	-7470A	PPB	PPB	06 Oct 2017 11:21:06	U	SMPL
CCB37 - 1	CCB37	-64	0.0376	-7470A	PPB	PPB	06 Oct 2017 11:23:12	U	SMPL
I5678-03MS - 1	OR-03-100417MS	13682	4.0532	-7470A	PPB	PPB	06 Oct 2017 11:25:19	U	SMPL
I5678-03MSD - 1	OR-03-100417MSD	13636	4.0398	-7470A	PPB	PPB	06 Oct 2017 11:27:26	U	SMPL
I5678-03LX5 - 1		61	0.0742	-7470A	PPB	PPB	06 Oct 2017 11:29:33	U	SMPL
I5678-03A - 1		11942	3.5449	-7470A	PPB	PPB	06 Oct 2017 11:31:40	U	SMPL
CCV38 - 1	CCV38	17001	5.0228	-7470A	PPB	PPB	06 Oct 2017 11:33:47	U	SMPL
CCB38 - 1	CCB38	-85	0.0315	-7470A	PPB	PPB	06 Oct 2017 11:35:55	U	SMPL