


ISM02.4

LB90628
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

45087



A= 0.0000e+000

B= 2.2278e-004

C= -4.7787e-002

Rho= 0.9999765

Accept=Accepted

slope
 $y = 1.215 \times 10^{-4} x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.00	0.000	0.001	0.001	221	0.000	221					-
0.2	0.200	0.163	-0.037	948	0.0 %	948					19
2.5	2.500	2.547	0.047	11647	0.0 %	11647					2
5.0	5.000	5.007	0.007	22689	0.0 %	22689					0
7.5	7.500	7.485	-0.015	33811	0.0 %	33811					0
10.0	10.000	9.997	-0.003	45087	0.0 %	45087					0

NB
10/6/18

LB90628
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stdnd Conc	Method	Units	Date	Type
	0 S0	221	-		0 ISM02.4	PPB	10/6/2017 14:37	Std
	0.2 S0.2	948	-		0.2 ISM02.4	PPB	10/6/2017 14:39	Std
	2.5 S2.5	11647	-		2.5 ISM02.4	PPB	10/6/2017 14:42	Std
	5 S5	22689	-		5 ISM02.4	PPB	10/6/2017 14:44	Std
	7.5 S7.5	33811	-		7.5 ISM02.4	PPB	10/6/2017 14:46	Std
	10 S10	45087	-		10 ISM02.4	PPB	10/6/2017 14:48	Std
ICV	ICV	17744	3.9052 -		ISM02.4	PPB	10/6/2017 14:51	SMPL
ICB	ICB	-191	-0.0903 -		ISM02.4	PPB	10/6/2017 14:54	SMPL
CCV041	CCV041	22503	4.9654 -		ISM02.4	PPB	10/6/2017 14:56	SMPL
CCB041	CCB041	-206	-0.0937 -		ISM02.4	PPB	10/6/2017 14:58	SMPL
PB102976BL	PBS004	10	-0.0456 -		ISM02.4	PPB	10/6/2017 15:35	SMPL
I5693-01	B-55(0.5-1.5)-100317	8684	1.8868 -		ISM02.4	PPB	10/6/2017 15:37	SMPL
I5693-02	B-55(3.5-5)-100317	9903	2.1584 -		ISM02.4	PPB	10/6/2017 15:40	SMPL
I5693-03	B-32(1-2)-100417	24921	5.5041 -		ISM02.4	PPB	10/6/2017 15:42	SMPL
I5693-04	B-32(2-4)-100417	105875	23.5391 -		ISM02.4	PPB	10/6/2017 15:44	SMPL
I5693-05	B-33(0.5-1.5)-100417	19151	4.2187 -		ISM02.4	PPB	10/6/2017 15:47	SMPL
I5693-05D	B-33(0.5-1.5)-100417D	19199	4.2294 -		ISM02.4	PPB	10/6/2017 15:49	SMPL
I5693-05S	B-33(0.5-1.5)-100417S	30734	6.7991 -		ISM02.4	PPB	10/6/2017 15:51	SMPL
CCV042	CCV042	22078	4.8708 -		ISM02.4	PPB	10/6/2017 15:54	SMPL
CCB042	CCB042	-243	-0.1019 -		ISM02.4	PPB	10/6/2017 15:56	SMPL
I5693-06	B-33(3.5-4.5)-100417	29066	6.4275 -		ISM02.4	PPB	10/6/2017 15:58	SMPL
I5693-07	B-30(0-1)-100417	48904	10.8471 -		ISM02.4	PPB	10/6/2017 16:00	SMPL
I5693-08	B-30(3-3.8)-100417	49662	11.0159 -		ISM02.4	PPB	10/6/2017 16:03	SMPL
PB102977BL	PBS004	-34	-0.0554 -		ISM02.4	PPB	10/6/2017 16:08	SMPL
I5695-03	B-74(0-1)-100417	5537	1.1857 -		ISM02.4	PPB	10/6/2017 16:10	SMPL
I5695-03D	B-74(0-1)-100417D	5923	1.2717 -		ISM02.4	PPB	10/6/2017 16:12	SMPL
I5695-03S	B-74(0-1)-100417S	16318	3.5875 -		ISM02.4	PPB	10/6/2017 16:15	SMPL
I5695-04	B-74(3-4)-100417	6108	1.313 -		ISM02.4	PPB	10/6/2017 16:17	SMPL
I5693-04DLX5	B-32(2-4)-100417	21246	4.6854 -		ISM02.4	PPB	10/6/2017 16:28	SMPL
I5693-07DLX2	B-30(0-1)-100417	24517	5.4141 -		ISM02.4	PPB	10/6/2017 16:31	SMPL
I5693-08DLX2	B-30(3-3.8)-100417	25163	5.558 -		ISM02.4	PPB	10/6/2017 16:33	SMPL
CCV043	CCV043	22203	4.8986 -		ISM02.4	PPB	10/6/2017 16:35	SMPL
CCB043	CCB043	-109	-0.0721 -		ISM02.4	PPB	10/6/2017 16:37	SMPL