

7470A


LB 90633

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36723



A= 0.0000e+000

B= 2.7416e-004

C= 2.6813e-002

Rho= 0.9998653

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	17	0.000	17				
0.2	0.200	0.227	0.027	729	0.0 %	729				
2.5	2.500	2.492	-0.008	8992	0.0 %	8992				
5.0	5.000	4.947	-0.053	17947	0.0 %	17947				
7.5	7.500	7.408	-0.092	26923	0.0 %	26923				
10.0	10.000	10.095	0.095	36723	0.0 %	36723				

LB90633 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 06 Oct 2017 16:48:47				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	17	-	0.00007470A	PPB	06 Oct 2017 16:52:57	S	Std
0.2 - 1	50.2	729	-	0.20007470A	PPB	06 Oct 2017 16:55:05	S	Std
2.5 - 1	52.5	8992	-	2.50007470A	PPB	06 Oct 2017 16:57:14	S	Std
5.0 - 1	55	17947	-	5.00007470A	PPB	06 Oct 2017 16:59:22	S	Std
7.5 - 1	57.5	26923	-	7.50007470A	PPB	06 Oct 2017 17:01:29	S	Std
10.0 - 1	510	36723	-	10.00007470A	PPB	06 Oct 2017 17:03:37	S	Std
ICV94 - 1	ICV94	13680	3.7773	-7470A	PPB	06 Oct 2017 17:08:32	U	SMPL
ICB94 - 1	ICB94	-130	-0.0088	-7470A	PPB	06 Oct 2017 17:10:38	U	SMPL
CCV44 - 1	CCV44	17744	4.8915	-7470A	PPB	06 Oct 2017 17:12:44	U	SMPL
CCB44 - 1	CCB44	-164	-0.0181	-7470A	PPB	06 Oct 2017 17:14:51	U	SMPL
CRA - 1	CRA	684	0.2143	-7470A	PPB	06 Oct 2017 17:16:59	U	SMPL
HighStd - 1	HighStd	37040	10.1816	-7470A	PPB	06 Oct 2017 17:19:05	U	SMPL
ChkStd - 1	ChkStd	24805	6.8273	-7470A	PPB	06 Oct 2017 17:21:12	U	SMPL
PB103013BL - 1	PBW	-233	-0.0371	-7470A	PPB	06 Oct 2017 17:23:21	U	SMPL
PB103013BS - 1	LCSW	13795	3.8088	-7470A	PPB	06 Oct 2017 17:25:29	U	SMPL
PB103013TB - 1	PB102759TB	-121	-0.0064	-7470A	PPB	06 Oct 2017 17:27:37	U	SMPL
I5563-09 - 1	NYSEG-PH4-SB-ESMI-1	-52	0.0126	-7470A	PPB	06 Oct 2017 17:29:45	U	SMPL
I5563-10 - 1	NYSEG-PH4-SB-ESMI-2	-30	0.0186	-7470A	PPB	06 Oct 2017 17:31:53	U	SMPL
I5563-11 - 1	NYSEG-PH4-SB-ESMI-3	-42	0.0153	-7470A	PPB	06 Oct 2017 17:34:02	U	SMPL
I5563-12 - 1	NYSEG-PH4-SB-ESMI-4	-23	0.0205	-7470A	PPB	06 Oct 2017 17:36:08	U	SMPL
CCV45 - 1	CCV45	18005	4.9630	-7470A	PPB	06 Oct 2017 17:38:15	U	SMPL
CCB45 - 1	CCB45	-156	-0.0160	-7470A	PPB	06 Oct 2017 17:40:22	U	SMPL
I5563-13 - 1	NYSEG-PH4-SB-ESMI-5	-36	0.0169	-7470A	PPB	06 Oct 2017 17:42:30	U	SMPL
I5563-14 - 1	NYSEG-PH4-SB-ESMI-6	-22	0.0208	-7470A	PPB	06 Oct 2017 17:44:37	U	SMPL
I5563-15 - 1	NYSEG-PH4-SB-ESMI-7	-19	0.0216	-7470A	PPB	06 Oct 2017 17:46:43	U	SMPL
I5563-16 - 1	NYSEG-PH4-SB-ESMI-8	-32	0.0180	-7470A	PPB	06 Oct 2017 17:48:50	U	SMPL
I5563-16DUP - 1	NYSEG-PH4-SB-ESMI-8DUP	-36	0.0169	-7470A	PPB	06 Oct 2017 17:50:57	U	SMPL
I5563-16MS - 1	NYSEG-PH4-SB-ESMI-8MS	13355	3.6882	-7470A	PPB	06 Oct 2017 17:53:05	U	SMPL
I5563-16MSD - 1	NYSEG-PH4-SB-ESMI-8MSD	13407	3.7024	-7470A	PPB	06 Oct 2017 17:55:13	U	SMPL
I5563-16LX5 - 1		-133	-0.0097	-7470A	PPB	06 Oct 2017 17:57:21	U	SMPL
I5563-16A - 1		12105	3.3455	-7470A	PPB	06 Oct 2017 17:59:30	U	SMPL
CCV46 - 1	CCV46	17656	4.8673	-7470A	PPB	06 Oct 2017 18:01:37	U	SMPL
CCB46 - 1	CCB46	-172	-0.0203	-7470A	PPB	06 Oct 2017 18:03:44	U	SMPL