



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90634

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/06/2017	17:25
2	CAL2	1	Water			7.00	10/06/2017	17:26
3	CAL3	1	Water			10.03	10/06/2017	17:28
4	ICV	1	Water			6.99	10/06/2017	17:30
5	CCV1	1	Water			2.01	10/06/2017	17:31
6	I5745-03	1	Solid	20.02	20	8.76	10/06/2017	17:32
7	I5745-03DUP	1	Solid	20.04	20	8.77	10/06/2017	17:33
8	I5745-06	1	Solid	20.02	20	8.75	10/06/2017	17:35
9	I5745-09	1	Solid	20.04	20	8.92	10/06/2017	17:36
10	I5748-03	1	Solid	20.02	20	8.49	10/06/2017	17:37
11	I5748-06	1	Solid	20.03	20	8.52	10/06/2017	17:38
12	I5748-09	1	Solid	20.02	20	7.96	10/06/2017	17:40
13	I5749-02	1	Solid	20.01	20	6.74	10/06/2017	17:41
14	CCV2	1	Water			12.01	10/06/2017	17:43

10590634

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4749

WorkList ID : 103837

Date : 10/6/2017 4:53:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2017	Solid	I5745-03	Corrosivity	Cool 4 deg C	PSEG05	H52	HR-06-100617-A	10/06/2017	9045C
10/11/2017	Solid	I5745-06	Corrosivity	Cool 4 deg C	PSEG05	H52	HR-06-100617-B	10/06/2017	9045C
10/11/2017	Solid	I5745-09	Corrosivity	Cool 4 deg C	PSEG05	H52	HR-06-100617-C	10/06/2017	9045C
10/11/2017	Solid	I5748-03	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-01-100617-A	10/06/2017	9045C
10/11/2017	Solid	I5748-06	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-01-100617-B	10/06/2017	9045C
10/11/2017	Solid	I5748-09	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-01-100617-C	10/06/2017	9045C
10/20/2017	Solid	I5749-02	Corrosivity	Cool 4 deg C	PSEG01	H52	SP-COMP-B	10/06/2017	9045C

Date/Time 10/06/17-5:00 PM
 Received by: FA
 Relinquished by: [Signature]

Date/Time 10/06/17-7:00 PM
 Received by: CD
 Relinquished by: JD