


7471A

Linear

 μ Abs.:

42785



A= 0.0000e+000

B= 2.3448e-004

C= -2.2238e-004

Rho= 0.9999729

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	-44	0.000	-44				
0.2	0.200	0.209	0.009	892	0.0 %	892				
2.5	2.500	2.516	0.016	10732	0.0 %	10732				
5.0	5.000	5.007	0.007	21354	0.0 %	21354				
7.5	7.500	7.446	-0.054	31757	0.0 %	31757	0			
10.0	10.000	10.032	0.032	42785	0.0 %	42785				

LB90637 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 06 Oct 2017 18:25:16				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-44	-	0.0000	7471A	PPB	06 Oct 2017 18:32:13	S Std
0.2 - 1	50.2	892	-	0.2000	7471A	PPB	06 Oct 2017 18:34:31	S Std
2.5 - 1	52.5	10732	-	2.5000	7471A	PPB	06 Oct 2017 18:36:48	S Std
5.0 - 1	55	21354	-	5.0000	7471A	PPB	06 Oct 2017 18:39:06	S Std
7.5 - 1	57.5	31757	-	7.5000	7471A	PPB	06 Oct 2017 18:41:23	S Std
10.0 - 1	50	42785	-	10.0000	7471A	PPB	06 Oct 2017 18:46:14	S Std
ICV95 - 1	ICV95	16536	3.8772	-7471A	PPB	PPB	06 Oct 2017 18:48:49	U SMPL
ICB95 - 1	ICB95	-243	-0.0572	-7471A	PPB	PPB	06 Oct 2017 18:51:06	U SMPL
CCV47 - 1	CCV47	21050	4.9357	-7471A	PPB	PPB	06 Oct 2017 18:53:23	U SMPL
CCB47 - 1	CCB47	-250	-0.0588	-7471A	PPB	PPB	06 Oct 2017 18:55:39	U SMPL
CRA - 1	CRA	845	0.1979	-7471A	PPB	PPB	06 Oct 2017 18:57:55	U SMPL
HighStd - 1	HighStd	42680	10.0075	-7471A	PPB	PPB	06 Oct 2017 19:00:11	U SMPL
ChkStd - 1	ChkStd	28169	6.6049	-7471A	PPB	PPB	06 Oct 2017 19:02:27	U SMPL
PB102978BL - 1	PBS	-278	-0.0654	-7471A	PPB	PPB	06 Oct 2017 19:04:46	U SMPL
PB102978BS - 1	LCSS	14047	3.2936	-7471A	PPB	PPB	06 Oct 2017 19:07:04	U SMPL
I5691-01 - 1	ENV-115-8(0-5)	1163	0.2725	-7471A	PPB	PPB	06 Oct 2017 19:09:22	U SMPL
I5691-02 - 1	ENV-115-8(5-10)	1693	0.3968	-7471A	PPB	PPB	06 Oct 2017 19:11:39	U SMPL
I5691-03 - 1	ENV-115-8(10-15)	10260	2.4056	-7471A	PPB	PPB	06 Oct 2017 19:13:57	U SMPL
I5691-04 - 1	ENV-115-8(15-20)	7201	1.6883	-7471A	PPB	PPB	06 Oct 2017 19:16:14	U SMPL
I5691-05 - 1	ENV-117-1(0-5)	2970	0.6962	-7471A	PPB	PPB	06 Oct 2017 19:18:29	U SMPL
CCV48 - 1	CCV48	21048	4.9352	-7471A	PPB	PPB	06 Oct 2017 19:20:46	U SMPL
CCB48 - 1	CCB48	-273	-0.0642	-7471A	PPB	PPB	06 Oct 2017 19:23:01	U SMPL
I5691-06 - 1	ENV-117-1(5-10)	5555	1.3023	-7471A	PPB	PPB	06 Oct 2017 19:25:17	U SMPL
I5691-07 - 1	ENV-117-1(10-15)	3428	0.8036	-7471A	PPB	PPB	06 Oct 2017 19:27:33	U SMPL
I5702-01 - 1	PHOENIX-RAW	395	0.0924	-7471A	PPB	PPB	06 Oct 2017 19:29:49	U SMPL
I5702-01DUP - 1	PHOENIX-RAWDUP	372	0.0870	-7471A	PPB	PPB	06 Oct 2017 19:32:06	U SMPL
I5702-01MS - 1	PHOENIX-RAWMS	14895	3.4924	-7471A	PPB	PPB	06 Oct 2017 19:34:22	U SMPL
I5702-01MSD - 1	PHOENIX-RAWMSD	14514	3.4031	-7471A	PPB	PPB	06 Oct 2017 19:36:39	U SMPL
I5708-01 - 1	NB-08-100517	1164	0.2727	-7471A	PPB	PPB	06 Oct 2017 19:38:57	U SMPL
I5714-01 - 1	TR-06-100517	965	0.2261	-7471A	PPB	PPB	06 Oct 2017 19:41:14	U SMPL
I5715-01 - 1	PL-02-100517-A	1119	0.2622	-7471A	PPB	PPB	06 Oct 2017 19:43:32	U SMPL
I5715-04 - 1	PL-02-100517-B	1181	0.2767	-7471A	PPB	PPB	06 Oct 2017 19:45:48	U SMPL
CCV49 - 1	CCV49	21175	4.9650	-7471A	PPB	PPB	06 Oct 2017 19:48:05	U SMPL
CCB49 - 1	CCB49	-237	-0.0558	-7471A	PPB	PPB	06 Oct 2017 19:50:21	U SMPL
I5715-07 - 1	PL-02-100517-C	831	0.1946	-7471A	PPB	PPB	06 Oct 2017 19:52:38	U SMPL
I5715-10 - 1	PL-02-100517-D	584	0.1367	-7471A	PPB	PPB	06 Oct 2017 19:54:55	U SMPL
I5715-13 - 1	PL-02-100517-E	557	0.1304	-7471A	PPB	PPB	06 Oct 2017 19:57:11	U SMPL
I5716-01 - 1	KEMSCO-COMP	571	0.1337	-7471A	PPB	PPB	06 Oct 2017 19:59:28	U SMPL
I5702-01LX5 - 1		86	0.0199	-7471A	PPB	PPB	06 Oct 2017 20:01:44	U SMPL
I5702-01A - 1		14610	3.4256	-7471A	PPB	PPB	06 Oct 2017 20:04:02	U SMPL
CCV50 - 1	CCV50	21281	4.9898	-7471A	PPB	PPB	06 Oct 2017 20:06:19	U SMPL
CCB50 - 1	CCB50	-144	-0.0340	-7471A	PPB	PPB	06 Oct 2017 20:08:36	U SMPL