

L390655


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

31116



A= 0.0000e+000

B= 3.2028e-004

C= 6.5734e-004

Rho= 0.9999796

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	-18	0.000	-18				
0.2	0.200	0.184	-0.016	571	0.0 %	571				
2.5	2.500	2.496	-0.004	7792	0.0 %	7792				
5.0	5.000	5.039	0.039	15730	0.0 %	15730				
7.5	7.500	7.520	0.020	23478	0.0 %	23478				
10.0	10.000	9.966	-0.034	31116	0.0 %	31116				

LB90655 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Oct 2017 10:28:26

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	SO	-18	-	0.0000	245.1	PPB	09 Oct 2017 10:32:26	S	Std
0.2 - 1	SO.2	571	-	0.2000	245.1	PPB	09 Oct 2017 10:34:34	S	Std
2.5 - 1	SO.5	7792	-	2.5000	245.1	PPB	09 Oct 2017 10:36:42	S	Std
5.0 - 1	SO.5	15730	-	5.0000	245.1	PPB	09 Oct 2017 10:38:51	S	Std
7.5 - 1	SO.5	23478	-	7.5000	245.1	PPB	09 Oct 2017 10:40:59	S	Std
10.0 - 1	SO	31116	-	10.0000	245.1	PPB	09 Oct 2017 10:43:07	S	Std
ICV96 - 1	ICV96	12209	3.9109	-245.1	PPB	PPB	09 Oct 2017 10:45:50	U	SMPL
ICB96 - 1	ICB96	-169	-0.0535	-245.1	PPB	PPB	09 Oct 2017 10:47:56	U	SMPL
CCV51 - 1	CCV51	15629	5.0063	-245.1	PPB	PPB	09 Oct 2017 10:50:03	U	SMPL
CCB51 - 1	CCB51	-165	-0.0522	-245.1	PPB	PPB	09 Oct 2017 10:52:10	U	SMPL
CRA - 1	CRA	574	0.1845	-245.1	PPB	PPB	09 Oct 2017 10:54:17	U	SMPL
HighStd - 1	HighStd	31519	10.0955	-245.1	PPB	PPB	09 Oct 2017 10:56:24	U	SMPL
ChkStd - 1	ChkStd	22334	7.1538	-245.1	PPB	PPB	09 Oct 2017 10:58:31	U	SMPL
PB103012BL - 1	PBW	-248	-0.0788	-245.1	PPB	PPB	09 Oct 2017 11:00:39	U	SMPL
PB103012BS - 1	LCSW	12155	3.8937	-245.1	PPB	PPB	09 Oct 2017 11:02:49	U	SMPL
I5706-01 - 1	MH-23	68	0.0224	-245.1	PPB	PPB	09 Oct 2017 11:04:56	U	SMPL
I5706-03 - 1	MH-A	3013	0.9657	-245.1	PPB	PPB	09 Oct 2017 11:07:04	U	SMPL
I5707-01 - 1	2AVE-06	-90	-0.0282	-245.1	PPB	PPB	09 Oct 2017 11:09:12	U	SMPL
I5717-01 - 1	COMP	79	0.0260	-245.1	PPB	PPB	09 Oct 2017 11:11:20	U	SMPL
I5717-01DUP - 1	COMPDUP	100	0.0327	-245.1	PPB	PPB	09 Oct 2017 11:13:27	U	SMPL
CCV52 - 1	CCV52	15753	5.0460	-245.1	PPB	PPB	09 Oct 2017 11:15:34	U	SMPL
CCB52 - 1	CCB52	-181	-0.0573	-245.1	PPB	PPB	09 Oct 2017 11:17:41	U	SMPL
I5717-01MS - 1	COMPMS	12572	4.0272	-245.1	PPB	PPB	09 Oct 2017 11:19:48	U	SMPL
I5717-01MSD - 1	COMPMSD	12177	3.9007	-245.1	PPB	PPB	09 Oct 2017 11:21:55	U	SMPL
I5731-01 - 1	100417-TIDO-F-D-B	-144	-0.0455	-245.1	PPB	PPB	09 Oct 2017 11:24:02	U	SMPL
I5731-02 - 1	100417-TIDO-F-D-S	-57	-0.0176	-245.1	PPB	PPB	09 Oct 2017 11:26:09	U	SMPL
I5731-03 - 1	100417-TIDO-F-U-B	-47	-0.0144	-245.1	PPB	PPB	09 Oct 2017 11:28:16	U	SMPL
I5731-04 - 1	100417-TIDO-F-U-S	-68	-0.0211	-245.1	PPB	PPB	09 Oct 2017 11:30:24	U	SMPL
I5717-01LX5 - 1		-11	-0.0029	-245.1	PPB	PPB	09 Oct 2017 11:32:32	U	SMPL
I5717-01A - 1		11921	3.8187	-245.1	PPB	PPB	09 Oct 2017 11:34:40	U	SMPL
CCV53 - 1	CCV53	15693	5.0268	-245.1	PPB	PPB	09 Oct 2017 11:36:49	U	SMPL
CCB53 - 1	CCB53	-148	-0.0467	-245.1	PPB	PPB	09 Oct 2017 11:38:56	U	SMPL