


7470A

LB90659
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35237



A= 0.0000e+000

B= 2.8518e-004

C= 3.2641e-002

Rho= 0.9998903

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	-31	0.000	-31				
0.2	0.200	0.209	0.009	619	0.0 %	619				
2.5	2.500	2.503	0.003	8661	0.0 %	8661				
5.0	5.000	4.985	-0.015	17367	0.0 %	17367				
7.5	7.500	7.398	-0.102	25827	0.0 %	25827				
10.0	10.000	10.081	0.081	35237	0.0 %	35237				

LB90659 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Oct 2017 11:57:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-31	-	0.0000	7470A	PPB	09 Oct 2017 12:05:08	S	Std
0.2 - 1	50.2	619	-	0.2000	7470A	PPB	09 Oct 2017 12:07:17	S	Std
2.5 - 1	52.5	8661	-	2.5000	7470A	PPB	09 Oct 2017 12:09:25	S	Std
5.0 - 1	55	17367	-	5.0000	7470A	PPB	09 Oct 2017 12:11:33	S	Std
7.5 - 1	57.5	25827	-	7.5000	7470A	PPB	09 Oct 2017 12:13:41	S	Std
10.0 - 1	510	35237	-	10.0000	7470A	PPB	09 Oct 2017 12:18:14	S	Std
ICV97 - 1	ICV97	13716	3.9441	-	7470A	PPB	09 Oct 2017 12:22:25	U	SMPL
ICB97 - 1	ICB97	-171	-0.0161	-	7470A	PPB	09 Oct 2017 12:24:32	U	SMPL
CCV54 - 1	CCV54	17398	4.9941	-	7470A	PPB	09 Oct 2017 12:26:38	U	SMPL
CCB54 - 1	CCB54	-164	-0.0141	-	7470A	PPB	09 Oct 2017 12:28:45	U	SMPL
CRA - 1	CRA	599	0.2035	-	7470A	PPB	09 Oct 2017 12:30:52	U	SMPL
HighStd - 1	HighStd	35251	10.0853	-	7470A	PPB	09 Oct 2017 12:32:59	U	SMPL
ChkStd - 1	ChkStd	23787	6.8161	-	7470A	PPB	09 Oct 2017 12:35:06	U	SMPL
PB103025BL - 1	PBW	-251	-0.0389	-	7470A	PPB	09 Oct 2017 12:37:15	U	SMPL
PB103025BS - 1	LCSW	13131	3.7773	-	7470A	PPB	09 Oct 2017 12:39:23	U	SMPL
PB103025TB - 1	PB102962TB	-119	-0.0013	-	7470A	PPB	09 Oct 2017 12:41:31	U	SMPL
I5706-02 - 1	MH-21	-27	0.0249	-	7470A	PPB	09 Oct 2017 12:43:39	U	SMPL
I5708-03 - 1	NB-08-100517	-43	0.0204	-	7470A	PPB	09 Oct 2017 12:45:47	U	SMPL
I5714-03 - 1	TR-06-100517	-14	0.0286	-	7470A	PPB	09 Oct 2017 12:47:55	U	SMPL
I5715-03 - 1	PL-02-100517-A	-46	0.0195	-	7470A	PPB	09 Oct 2017 12:50:01	U	SMPL
CCV55 - 1	CCV55	17138	4.9200	-	7470A	PPB	09 Oct 2017 12:52:08	U	SMPL
CCB55 - 1	CCB55	-145	-0.0087	-	7470A	PPB	09 Oct 2017 12:54:15	U	SMPL
I5715-06 - 1	PL-02-100517-B	-39	0.0215	-	7470A	PPB	09 Oct 2017 12:56:22	U	SMPL
I5715-09 - 1	PL-02-100517-C	9	0.0352	-	7470A	PPB	09 Oct 2017 12:58:29	U	SMPL
I5715-12 - 1	PL-02-100517-D	-39	0.0215	-	7470A	PPB	09 Oct 2017 13:00:36	U	SMPL
I5715-15 - 1	PL-02-100517-E	-25	0.0255	-	7470A	PPB	09 Oct 2017 13:02:43	U	SMPL
I5716-04 - 1	KEMSCO-COMP	-48	0.0190	-	7470A	PPB	09 Oct 2017 13:04:50	U	SMPL
I5716-04DUP - 1	KEMSCO-COMPDUP	-32	0.0235	-	7470A	PPB	09 Oct 2017 13:06:58	U	SMPL
I5716-04MS - 1	KEMSCO-COMPMS	12833	3.6923	-	7470A	PPB	09 Oct 2017 13:09:06	U	SMPL
I5716-04MSD - 1	KEMSCO-COMPMSD	12579	3.6199	-	7470A	PPB	09 Oct 2017 13:11:15	U	SMPL
I5716-04LX5 - 1		-28	0.0247	-	7470A	PPB	09 Oct 2017 13:16:11	U	SMPL
I5716-04A - 1		12352	3.5551	-	7470A	PPB	09 Oct 2017 13:18:17	U	SMPL
CCV56 - 1	CCV56	17261	4.9550	-	7470A	PPB	09 Oct 2017 13:20:24	U	SMPL
CCB56 - 1	CCB56	-153	-0.0110	-	7470A	PPB	09 Oct 2017 13:22:31	U	SMPL