


7471A

LB90673
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41846



A= 0.0000e+000

B= 2.3846e-004

C= -1.2693e-002

Rho= 0.9999807

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	-5	0.000	-5				
0.2	0.200	0.196	-0.004	876	0.0 %	876				
2.5	2.500	2.509	0.009	10576	0.0 %	10576				
5.0	5.000	5.001	0.001	21027	0.0 %	21027				
7.5	7.500	7.541	0.041	31678	0.0 %	31678	0			
10.0	10.000	9.966	-0.034	41846	0.0 %	41846				

LB90673 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 09 Oct 2017 16:04:23				
Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-5	-	0.0000	7471A	PPB	09 Oct 2017 16:17:03	S Std
0.2 - 1	50.2	876	-	0.2000	7471A	PPB	09 Oct 2017 16:19:21	S Std
2.5 - 1	52.5	10576	-	2.5000	7471A	PPB	09 Oct 2017 16:21:38	S Std
5.0 - 1	55	21027	-	5.0000	7471A	PPB	09 Oct 2017 16:23:56	S Std
7.5 - 1	57.5	31678	-	7.5000	7471A	PPB	09 Oct 2017 16:26:13	S Std
10.0 - 1	510	41846	-	10.0000	7471A	PPB	09 Oct 2017 16:31:15	S Std
ICV98 - 1	ICV98	15394	3.6581	-	7471A	PPB	09 Oct 2017 16:33:57	U SMPL
ICB98 - 1	ICB98	-202	-0.0609	-	7471A	PPB	09 Oct 2017 16:36:13	U SMPL
CCV57 - 1	CCV57	22046	5.2444	-	7471A	PPB	09 Oct 2017 16:38:29	U SMPL
CCB57 - 1	CCB57	-202	-0.0609	-	7471A	PPB	09 Oct 2017 16:40:45	U SMPL
CRA - 1	CRA	828	0.1848	-	7471A	PPB	09 Oct 2017 16:43:01	U SMPL
HighStd - 1	HighStd	42582	10.1414	-	7471A	PPB	09 Oct 2017 16:45:16	U SMPL
ChkStd - 1	ChkStd	29732	7.0772	-	7471A	PPB	09 Oct 2017 16:47:33	U SMPL
PB103041BL - 1	PBS	-279	-0.0792	-	7471A	PPB	09 Oct 2017 16:49:51	U SMPL
PB103041BS - 1	LCSS	14828	3.5232	-	7471A	PPB	09 Oct 2017 16:52:09	U SMPL
I5729-01 - 1	1W-(1-2)	7816	1.8511	-	7471A	PPB	09 Oct 2017 16:54:26	U SMPL
I5729-02 - 1	1E-(3-4)	9208	2.1830	-	7471A	PPB	09 Oct 2017 16:56:44	U SMPL
I5729-03 - 1	2N-(1-2)	1762	0.4075	-	7471A	PPB	09 Oct 2017 16:59:01	U SMPL
I5729-04 - 1	2S-(1-2)	1656	0.3822	-	7471A	PPB	09 Oct 2017 17:01:19	U SMPL
I5729-05 - 1	3E-(1-2)	1170	0.2663	-	7471A	PPB	09 Oct 2017 17:03:34	U SMPL
CCV58 - 1	CCV58	22301	5.3052	-	7471A	PPB	09 Oct 2017 17:05:50	U SMPL
CCB58 - 1	CCB58	-208	-0.0623	-	7471A	PPB	09 Oct 2017 17:08:07	U SMPL
I5729-06 - 1	3W-(1-2)	2536	0.5920	-	7471A	PPB	09 Oct 2017 17:10:23	U SMPL
I5729-07 - 1	4W-(2-3)	24496	5.8286	-	7471A	PPB	09 Oct 2017 17:12:39	U SMPL
I5729-08 - 1	4E-(2-3)	16062	3.8174	-	7471A	PPB	09 Oct 2017 17:14:55	U SMPL
I5729-09 - 1	5N-(2-3)	7389	1.7493	-	7471A	PPB	09 Oct 2017 17:17:12	U SMPL
I5729-10 - 1	5S-(2-3)	3126	0.7327	-	7471A	PPB	09 Oct 2017 17:19:29	U SMPL
I5729-11 - 1	6W-(2-3)	6839	1.6181	-	7471A	PPB	09 Oct 2017 17:21:45	U SMPL
I5729-11DUP - 1	6W-(2-3)DUP	7032	1.6641	-	7471A	PPB	09 Oct 2017 17:24:02	U SMPL
I5729-11MS - 1	6W-(2-3)MS	20585	4.8960	-	7471A	PPB	09 Oct 2017 17:26:20	U SMPL
I5729-11MSD - 1	6W-(2-3)MSD	19669	4.6776	-	7471A	PPB	09 Oct 2017 17:28:38	U SMPL
I5729-12 - 1	6E-(1-2)	11475	2.7236	-	7471A	PPB	09 Oct 2017 17:30:54	U SMPL
CCV59 - 1	CCV59	21232	5.0503	-	7471A	PPB	09 Oct 2017 17:33:10	U SMPL
CCB59 - 1	CCB59	-216	-0.0642	-	7471A	PPB	09 Oct 2017 17:35:27	U SMPL
I5729-13 - 1	7E-(1-2)	5875	1.3883	-	7471A	PPB	09 Oct 2017 17:37:44	U SMPL
I5729-14 - 1	7W-(1-2)	5932	1.4018	-	7471A	PPB	09 Oct 2017 17:40:00	U SMPL
PB103042BL - 1	PBS	-89	-0.0339	-	7471A	PPB	09 Oct 2017 17:42:16	U SMPL
PB103042BS - 1	LCSS	15743	3.7414	-	7471A	PPB	09 Oct 2017 17:44:32	U SMPL
I5745-01 - 1	HR-06-100617-A	32529	7.7441	-	7471A	PPB	09 Oct 2017 17:46:48	U SMPL
I5745-04 - 1	HR-06-100617-B	6492	1.5354	-	7471A	PPB	09 Oct 2017 17:49:06	U SMPL
I5745-07 - 1	HR-06-100617-C	5809	1.3725	-	7471A	PPB	09 Oct 2017 17:51:24	U SMPL
I5748-01 - 1	CL-01-100617-A	1968	0.4566	-	7471A	PPB	09 Oct 2017 17:53:41	U SMPL
I5748-04 - 1	CL-01-100617-B	3840	0.9030	-	7471A	PPB	09 Oct 2017 17:56:00	U SMPL
I5748-07 - 1	CL-01-100617-C	1702	0.3932	-	7471A	PPB	09 Oct 2017 17:58:16	U SMPL
CCV60 - 1	CCV60	21415	5.0939	-	7471A	PPB	09 Oct 2017 18:00:32	U SMPL
CCB60 - 1	CCB60	-194	-0.0590	-	7471A	PPB	09 Oct 2017 18:02:49	U SMPL
I5748-07DUP - 1	CL-01-100617-CDUP	1672	0.3860	-	7471A	PPB	09 Oct 2017 18:10:37	U SMPL
I5748-07MS - 1	CL-01-100617-CMS	17429	4.1434	-	7471A	PPB	09 Oct 2017 18:12:53	U SMPL
I5748-07MSD - 1	CL-01-100617-CMSD	17423	4.1420	-	7471A	PPB	09 Oct 2017 18:15:10	U SMPL
I5749-01 - 1	SP-COMP	1260	0.2878	-	7471A	PPB	09 Oct 2017 18:17:26	U SMPL
I5729-11LX5 - 1		1319	0.3018	-	7471A	PPB	09 Oct 2017 18:19:43	U SMPL
I5729-11A - 1		21499	5.1139	-	7471A	PPB	09 Oct 2017 18:22:00	U SMPL
I5748-07LX5 - 1		147	0.0224	-	7471A	PPB	09 Oct 2017 18:24:17	U SMPL
I5748-07A - 1		16295	3.8730	-	7471A	PPB	09 Oct 2017 18:26:35	U SMPL
CCV61 - 1	CCV61	21354	5.0794	-	7471A	PPB	09 Oct 2017 18:28:52	U SMPL
CCB61 - 1	CCB61	-206	-0.0618	-	7471A	PPB	09 Oct 2017 18:31:08	U SMPL