



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90674

Temp.: 21.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/09/2017	16:30
2	CAL2	1	Water			7.01	10/09/2017	16:31
3	CAL3	1	Water			10.03	10/09/2017	16:33
4	ICV	1	Water			7.00	10/09/2017	16:35
5	CCV1	1	Water			2.01	10/09/2017	16:36
6	I5762-01	1	Water			4.14	10/09/2017	16:37
7	I5762-01DUP	1	Water			4.15	10/09/2017	16:38
8	CCV2	1	Water			12.01	10/09/2017	16:39

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i5762

WorkList ID : 103936

Date : 10/9/2017 4:16:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2017	Water	I5762-01-A	pH	Cool 4 deg C	STER04	H11	POOL-WATER	10/02/2017	SM 4500-PH B

Date/Time

Received by:

Relinquished by:

10/09/17 4:16 PM

JP

Date/Time

Received by:

Relinquished by:

10/09/17 5:00 PM

CP

JP