



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB90676

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	10/09/2017	09:05
2	I5706-01		1	>100.0	>212.0	10/09/2017	09:40
3	I5706-01DUP		1	>100.0	>212.0	10/09/2017	10:02
4	I5706-03		1	>100.0	>212.0	10/09/2017	10:35
5	I5707-01		1	>100.0	>212.0	10/09/2017	11:05

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 90676

WorkList Name : FP10517

WorkList ID : 103775

Date : 10/5/2017 5:45:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2017	Water	I5706-01	Flash Point	Cool 4 deg C	DIST01	H11	MH-23	10/05/2017	1010A
10/10/2017	Water	I5706-03	Flash Point	Cool 4 deg C	DIST01	H11	MH-A	10/05/2017	1010A
10/19/2017	Water	I5707-01	Flash Point	Cool 4 deg C	DIST01	H11	2AVE-06	10/05/2017	1010A

Date/Time 10/9/17 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 10/9/17 11:10 AM
 Received by: CP
 Relinquished by: KS