



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB90677

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	10/09/2017	11:32
2	I5640-01		1	>100.0	>212.0	10/09/2017	12:05
3	I5640-01DUP		1	>100.0	>212.0	10/09/2017	12:40
4	I5640-02		1	>100.0	>212.0	10/09/2017	13:06
5	I5640-03		1	>100.0	>212.0	10/09/2017	13:35
6	I5640-06		1	>100.0	>212.0	10/09/2017	13:56
7	I5640-07		1	>100.0	>212.0	10/09/2017	14:25
8	I5640-08		1	>100.0	>212.0	10/09/2017	14:50

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 90677

WorkList Name : FP10517-1

WorkList ID : 103776

Date : 10/5/2017 5:50:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/13/2017	Solid	I5640-01	Flash Point	Cool 4 deg C	ATLA08	J31	DRUM-1	09/29/2017	1010A
10/13/2017	Solid	I5640-02	Flash Point	Cool 4 deg C	ATLA08	J31	DRUM-2	09/29/2017	1010A
10/13/2017	Solid	I5640-03	Flash Point	Cool 4 deg C	ATLA08	J31	DRUM-3	09/29/2017	1010A
10/13/2017	Solid	I5640-06	Flash Point	Cool 4 deg C	ATLA08	J31	DRUM-6	09/29/2017	1010A
10/13/2017	Solid	I5640-07	Flash Point	Cool 4 deg C	ATLA08	J31	DRUM-7	09/29/2017	1010A
10/13/2017	Solid	I5640-08	Flash Point	Cool 4 deg C	ATLA08	J31	DRUM-8	09/29/2017	1010A

Date/Time 10/9/17 8:30 AM

Received by: KS

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

10/9/17 14:55
CP
KS