



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90678

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/09/2017	18:00
2	CAL2	1	Water			7.01	10/09/2017	18:01
3	CAL3	1	Water			10.03	10/09/2017	18:03
4	ICV	1	Water			6.99	10/09/2017	18:05
5	CCV1	1	Water			2.02	10/09/2017	18:06
6	I5750-01	1	Solid	20.02	20	6.15	10/09/2017	18:09
7	I5763-04	1	Solid	20.04	20	8.65	10/09/2017	18:10
8	I5765-03	1	Solid	20.02	20	6.60	10/09/2017	18:11
9	I5766-07	1	Solid	20.05	20	9.77	10/09/2017	18:13
10	I5766-07DUP	1	Solid	20.02	20	9.78	10/09/2017	18:15
11	CCV2	1	Water			12.01	10/09/2017	18:17

LB906F8

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5763

WorkList ID : 103941

Date : 10/9/2017 5:28:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2017	Solid	I5750-01	Corrosivity	Cool 4 deg C	PSEG02	H52	DRUM-COMP(1-10)	10/06/2017	9045C
10/15/2017	Solid	I5763-04	Corrosivity	Cool 4 deg C	PSEG01	H52	220-2(2TRH)-COMP	10/09/2017	9045C
	Solid	I5765-03	Corrosivity	Cool 4 deg C	PSEG05	H52	BU-03-100917	10/09/2017	9045C
10/12/2017	Solid	I5766-07	Corrosivity	Cool 4 deg C	PSEG01	H53	RS-CONCRETE	10/09/2017	9040C

Date/Time 10/09/17 5:30 PM
 Received by: JP
 Relinquished by: CJ

Date/Time 10/09/17 7:20 PM
 Received by: CJ
 Relinquished by: JP