



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90683

Temp.: 22.2 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/09/2017	19:00
2	CAL2	1	Water			7.01	10/09/2017	19:01
3	CAL3	1	Water			10.04	10/09/2017	19:03
4	ICV	1	Water			7.00	10/09/2017	19:05
5	CCV1	1	Water			2.01	10/09/2017	19:06
6	I5766-01	1	Water			11.80	10/09/2017	19:08
7	I5766-01DUP	1	Water			11.81	10/09/2017	19:10
8	CCV2	1	Water			12.01	10/09/2017	19:12

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5766

WorkList ID : 103946

Date : 10/9/2017 6:50:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2017	Water	I5766-01-13	pH	Cool 4 deg C	PSEG01	H53	RS-FRACTANK	10/09/2017	SM 4500-PH B

Date/Time

Received by:

Relinquished by:

10/09/17 6:50 PM

71

CS

Date/Time

Received by:

Relinquished by:

10/09/17 7:30 PM

CS

SP